

Data Path : Z:\HPCHEM1\BNA M\Data\BM111915\
 Data File : BM003188.D
 Acq On : 19 Nov 2015 11:47
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.453

Quant Time: Nov 20 03:35:55 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 03:16:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	14936	0.40	ng/ul	0.00
4) Naphthalene-d8	10.53	136	55491	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.38	164	27300	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.12	188	58023	0.40	ng/ul	0.00
18) Chrysene-d12	21.33	240	62460	0.40	ng/ul	0.00
22) Perylene-d12	23.60	264	58650	0.40	ng/ul	0.01
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.20	96	15738	0.99	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.12	152	31141	0.41	ng/ul	0.00
16) Fluoranthene-d10	19.17	212	62364	0.44	ng/ul	0.00
Target Compounds				Ovalue		
3) 1,4-Dioxane	3.23	88	18028	0.99	ng/ul#	100
5) Naphthalene	10.58	128	57702	0.41	ng/ul	99
7) 2-Methylnaphthalene	12.19	142	38538	0.42	ng/ul#	100
9) Acenaphthylene	14.09	152	50001	0.44	ng/ul	100
10) Acenaphthene	14.45	153	39390	0.41	ng/ul	97
11) Fluorene	15.44	166	43387	0.42	ng/ul	97
13) Pentachlorophenol	16.78	266	7717	0.56	ng/ul#	18
14) Phenanthrene	17.17	178	70248	0.42	ng/ul	98
15) Anthracene	17.26	178	63032	0.45	ng/ul	99
17) Fluoranthene	19.19	202	81170	0.43	ng/ul	93
19) Pyrene	19.57	202	83248	0.39	ng/ul	93
20) Benzo(a)anthracene	21.31	228	80971	0.45	ng/ul#	92
21) Chrysene	21.36	228	79532	0.39	ng/ul	92
23) Benzo(b)fluoranthene	22.91	252	84574	0.38	ng/ul	99
24) Benzo(k)fluoranthene	22.95	252	80305	0.38	ng/ul	95
25) Benzo(a)pyrene	23.49	252	80959	0.41	ng/ul	98
26) Indeno(1,2,3-cd)pyrene	25.87	276	92281	0.39	ng/ul#	100
27) Dibenzo(a,h)anthracene	25.89	278	74271	0.39	ng/ul	99
28) Benzo(g,h,i)perylene	26.57	276	79157	0.38	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\Data\BM111915\
Data File : BM003188.D
Acq On : 19 Nov 2015 11:47
Operator : UM/IZ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.453

Quant Time: Nov 20 03:35:55 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 20 03:16:26 2015
Response via : Initial Calibration

