

Data Path : Z:\HPCHEM1\BNA M\Data\BM111915\
 Data File : BM003188.D
 Acq On : 19 Nov 2015 11:47
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.453

Quant Time: Nov 20 03:35:55 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 03:16:26 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.421	1.4	101	0.00
3	1,4-Dioxane	0.486	0.483	0.6	103	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
5	Naphthalene	1.027	1.040	-1.3	107	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.561	-3.1	108	0.00
7	2-Methylnaphthalene	0.669	0.694	-3.7	109	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
9	Acenaphthylene	1.665	1.832	-10.0	120	0.00
10 C	Acenaphthene	1.406	1.443	-2.6	110	0.00
11	Fluorene	1.521	1.589	-4.5	111	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
13	Pentachlorophenol	0.094	0.133	-41.5#	161#	0.00
14	Phenanthrene	1.161	1.211	-4.3	114	0.00
15	Anthracene	0.972	1.086	-11.7	124	0.00
16 SURR	Fluoranthene-d10	0.980	1.075	-9.7	122	0.00
17 C	Fluoranthene	1.287	1.399	-8.7	121	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
19	Pyrene	1.371	1.333	2.8	125	0.01
20	Benzo(a)anthracene	1.160	1.296	-11.7	150	0.00
21	Chrysene	1.306	1.273	2.5	124	0.00
22 I	Perylene-d12	1.000	1.000	0.0	148	0.01
23	Benzo(b)fluoranthene	1.521	1.442	5.2	142	0.00
24	Benzo(k)fluoranthene	1.438	1.369	4.8	135	0.00
25 C	Benzo(a)pyrene	1.353	1.380	-2.0	149	0.01
26	Indeno(1,2,3-cd)pyrene	1.612	1.573	2.4	142	0.00
27	Dibenzo(a,h)anthracene	1.303	1.266	2.8	142	0.01
28	Benzo(q,h,i)perylene	1.416	1.350	4.7	138	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0