

Data Path : Z:\HPCHEM1\BNA M\Data\BM111915\
 Data File : BM003193.D
 Acq On : 19 Nov 2015 17:04
 Operator : UM/IZ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.454

Quant Time: Nov 20 02:43:01 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 02:40:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.418	2.1	95	0.00
3	1,4-Dioxane	0.486	0.466	4.1	94	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
5	Naphthalene	1.027	1.033	-0.6	100	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.559	-2.8	102	0.00
7	2-Methylnaphthalene	0.669	0.691	-3.3	103	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
9	Acenaphthylene	1.665	1.810	-8.7	110	0.00
10 C	Acenaphthene	1.406	1.425	-1.4	101	0.00
11	Fluorene	1.521	1.544	-1.5	101	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
13	Pentachlorophenol	0.094	0.092	2.1	99	0.00
14	Phenanthrene	1.161	1.188	-2.3	99	0.00
15	Anthracene	0.972	1.025	-5.5	104	0.00
16 SURR	Fluoranthene-d10	0.980	0.983	-0.3	100	0.00
17 C	Fluoranthene	1.287	1.265	1.7	97	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
19	Pyrene	1.371	1.396	-1.8	99	0.00
20	Benzo(a)anthracene	1.160	1.255	-8.2	110	0.00
21	Chrysene	1.306	1.311	-0.4	97	0.00
22 I	Perylene-d12	1.000	1.000	0.0	110	0.00
23	Benzo(b)fluoranthene	1.521	1.547	-1.7	114	0.00
24	Benzo(k)fluoranthene	1.438	1.294	10.0	95	0.00
25 C	Benzo(a)pyrene	1.353	1.379	-1.9	111	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.594	1.1	108	0.00
27	Dibenzo(a,h)anthracene	1.303	1.294	0.7	108	0.00
28	Benzo(q,h,i)perylene	1.416	1.372	3.1	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0