

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112015\
 Data File : BM003220.D
 Acq On : 20 Nov 2015 13:50
 Operator : UM/IZ
 Sample : G4323-18
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4J02

Quant Time: Nov 21 03:50:47 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 03:26:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	13562	0.40	ng/ul	0.00
4) Naphthalene-d8	10.53	136	53259	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.38	164	29491	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.12	188	66180	0.40	ng/ul	0.00
18) Chrysene-d12	21.33	240	69576	0.40	ng/ul	0.00
22) Perylene-d12	23.60	264	65134	0.40	ng/ul	0.01
System Monitoring Compounds						
2) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
6) 2-Methylnaphthalene-d10	12.12	152	23385	0.32	ng/ul	0.00
16) Fluoranthene-d10	19.17	212	66553	0.41	ng/ul	0.00
Target Compounds				Ovalue		
5) Naphthalene	10.58	128	114273	0.84	ng/ul	100
7) 2-Methylnaphthalene	12.19	142	52960	0.59	ng/ul#	100
11) Fluorene	15.44	166	71293	0.64	ng/ul	98
13) Pentachlorophenol	16.78	266	26793	1.72	ng/ul	96
14) Phenanthrene	17.17	178	158993	0.83	ng/ul	99
17) Fluoranthene	19.20	202	155837	0.73	ng/ul	91
19) Pyrene	19.57	202	2194	0.01	ng/ul#	74
20) Benzo(a)anthracene	21.31	228	5860	0.03	ng/ul#	88
24) Benzo(k)fluoranthene	22.96	252	135663	0.58	ng/ul	98
26) Indeno(1,2,3-cd)pyrene	25.87	276	4570	0.02	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112015\
Data File : BM003220.D
Acq On : 20 Nov 2015 13:50
Operator : UM/IZ
Sample : G4323-18
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4J02

Quant Time: Nov 21 03:50:47 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 21 03:26:56 2015
Response via : Initial Calibration

