

Data Path : Z:\HPCHEM1\BNA M\Data\BM112015\
 Data File : BM003215.D
 Acq On : 20 Nov 2015 09:25
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.456

Quant Time: Nov 21 03:20:40 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 13:20:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.416	2.6	105	0.00
3	1,4-Dioxane	0.486	0.473	2.7	106	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
5	Naphthalene	1.027	1.040	-1.3	113	0.01
6 SURR	2-Methylnaphthalene-d10	0.544	0.560	-2.9	114	0.00
7	2-Methylnaphthalene	0.669	0.697	-4.2	116	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
9	Acenaphthylene	1.665	1.791	-7.6	121	0.00
10 C	Acenaphthene	1.406	1.422	-1.1	111	0.02
11	Fluorene	1.521	1.522	-0.1	110	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
13	Pentachlorophenol	0.094	0.099	-5.3	115	0.00
14	Phenanthrene	1.161	1.184	-2.0	107	0.02
15	Anthracene	0.972	1.033	-6.3	113	0.00
16 SURR	Fluoranthene-d10	0.980	0.968	1.2	106	0.00
17 C	Fluoranthene	1.287	1.259	2.2	105	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	111	0.01
19	Pyrene	1.371	1.364	0.5	108	0.00
20	Benzo(a)anthracene	1.160	1.275	-9.9	124	0.01
21	Chrysene	1.306	1.327	-1.6	109	0.01
22 I	Perylene-d12	1.000	1.000	0.0	128	0.00
23	Benzo(b)fluoranthene	1.521	1.502	1.2	128	0.01
24	Benzo(k)fluoranthene	1.438	1.278	11.1	109	0.01
25 C	Benzo(a)pyrene	1.353	1.371	-1.3	129	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.608	0.2	126	0.01
27	Dibenzo(a,h)anthracene	1.303	1.277	2.0	124	0.00
28	Benzo(q,h,i)perylene	1.416	1.383	2.3	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0