

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112015\
 Data File : BM003224.D
 Acq On : 20 Nov 2015 16:56
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.457

Quant Time: Nov 21 04:23:05 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 03:26:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.411	3.7	95	0.00
3	1,4-Dioxane	0.486	0.471	3.1	96	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
5	Naphthalene	1.027	1.043	-1.6	104	0.01
6 SURR	2-Methylnaphthalene-d10	0.544	0.561	-3.1	105	0.00
7	2-Methylnaphthalene	0.669	0.697	-4.2	106	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
9	Acenaphthylene	1.665	1.844	-10.8	119	0.00
10 C	Acenaphthene	1.406	1.442	-2.6	108	0.02
11	Fluorene	1.521	1.578	-3.7	109	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
13	Pentachlorophenol	0.094	0.104	-10.6	124	0.00
14	Phenanthrene	1.161	1.174	-1.1	108	0.00
15	Anthracene	0.972	1.085	-11.6	121	0.00
16 SURR	Fluoranthene-d10	0.980	1.049	-7.0	117	0.00
17 C	Fluoranthene	1.287	1.366	-6.1	116	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	122	0.01
19	Pyrene	1.371	1.455	-6.1	126	0.00
20	Benzo(a)anthracene	1.160	1.331	-14.7	142	0.01
21	Chrysene	1.306	1.314	-0.6	119	0.01
22 I	Perylene-d12	1.000	1.000	0.0	140	0.00
23	Benzo(b)fluoranthene	1.521	1.489	2.1	139	0.01
24	Benzo(k)fluoranthene	1.438	1.313	8.7	122	0.01
25 C	Benzo(a)pyrene	1.353	1.370	-1.3	140	0.01
26	Indeno(1,2,3-cd)pyrene	1.612	1.566	2.9	134	0.01
27	Dibenzo(a,h)anthracene	1.303	1.253	3.8	133	0.01
28	Benzo(q,h,i)perylene	1.416	1.332	5.9	129	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0