

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112015\
 Data File : BM003239.D
 Acq On : 21 Nov 2015 01:55
 Operator : UM/IZ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.458

Quant Time: Nov 21 05:38:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 04:24:04 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.402	5.9	97	0.00
3	1,4-Dioxane	0.486	0.459	5.6	98	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
5	Naphthalene	1.027	1.052	-2.4	113	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.569	-4.6	115	0.00
7	2-Methylnaphthalene	0.669	0.718	-7.3	118	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
9	Acenaphthylene	1.665	1.980	-18.9	137	0.00
10 C	Acenaphthene	1.406	1.448	-3.0	116	0.00
11	Fluorene	1.521	1.553	-2.1	115	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
13	Pentachlorophenol	0.094	0.103	-9.6	120	0.00
14	Phenanthrene	1.161	1.198	-3.2	108	0.00
15	Anthracene	0.972	1.086	-11.7	119	0.00
16 SURR	Fluoranthene-d10	0.980	0.992	-1.2	109	0.00
17 C	Fluoranthene	1.287	1.307	-1.6	109	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
19	Pyrene	1.371	1.467	-7.0	114	0.00
20	Benzo(a)anthracene	1.160	1.332	-14.8	127	0.00
21	Chrysene	1.306	1.340	-2.6	109	0.00
22 I	Perylene-d12	1.000	1.000	0.0	128	0.00
23	Benzo(b)fluoranthene	1.521	1.493	1.8	127	0.00
24	Benzo(k)fluoranthene	1.438	1.304	9.3	111	0.00
25 C	Benzo(a)pyrene	1.353	1.396	-3.2	131	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.596	1.0	125	0.00
27	Dibenzo(a,h)anthracene	1.303	1.268	2.7	123	0.00
28	Benzo(q,h,i)perylene	1.416	1.354	4.4	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0