


Data Path : Z:\HPCHEM1\BNA M\Data\BM112015\
 Data File : BM003214.D
 Acq On : 20 Nov 2015 08:38
 Operator : UM/IZ
 Sample : DFTPP47
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP47

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 23 11:49:00 2015



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.0	30685	PASS
68	69	0.00	2	1.7	645	PASS
69	198	0.00	100	30.4	38856	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.4	58024	PASS
197	198	0.00	1	0.3	353	PASS
198	198	100	100	100.0	127853	PASS
199	198	5	9	6.8	8747	PASS
275	198	10	60	24.7	31522	PASS
365	198	1	100	3.0	3814	PASS
441	443	0.01	100	79.6	16291	PASS
442	198	50	100	83.1	106298	PASS
443	442	15	24	19.3	20469	PASS

Scan 2456 (17.533 min): BM003214.D
DFTPP47

m/z	Abundance
44.00	416.0
73.00	377.0
207.00	899.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
38.00	449.0
39.10	2773.0
50.10	9881.0
51.10	39352.0
52.00	1957.0
55.00	316.0
56.00	1249.0
57.00	2774.0
61.00	649.0
62.10	549.0
63.00	1710.0
65.00	873.0
68.00	779.0
69.00	49080.0
73.00	521.0
74.00	4898.0
75.00	8194.0
76.10	2751.0
77.10	60104.0
78.10	4446.0
79.00	3932.0
80.00	2990.0
81.00	4271.0
82.00	967.0
83.00	1028.0
85.00	1068.0
86.00	1103.0
87.00	572.0
91.00	1039.0
92.00	1182.0
93.00	6650.0
94.00	492.0
96.00	378.0
98.00	5143.0
99.00	4475.0
100.00	392.0
101.00	2546.0
103.00	754.0
104.00	1484.0
105.00	1312.0
106.00	535.0
107.00	18832.0
108.00	3039.0
109.10	645.0
110.00	35656.0
111.00	5416.0
112.00	692.0
116.00	1104.0
117.00	15830.0
118.00	1214.0
122.00	1425.0
123.00	1914.0
124.00	837.0
125.00	892.0
127.00	70576.0
128.00	5292.0
129.00	27024.0
130.00	2381.0
131.10	449.0
132.00	300.0
134.00	836.0
135.00	2218.0
136.00	917.0
137.00	1271.0
139.90	327.0
141.00	3361.0
142.00	1126.0
143.00	741.0
145.90	685.0
147.00	1744.0
148.00	4287.0
149.00	1106.0
151.00	393.0
151.90	332.0
153.00	1157.0
154.00	807.0
155.00	1998.0
156.00	2993.0
157.00	556.0
158.00	570.0
159.00	547.0
160.00	1072.0
161.00	1607.0
162.00	469.0
165.00	1230.0
166.00	953.0
167.00	7465.0
168.00	4881.0
169.00	692.0
171.00	305.0
171.90	610.0
173.10	847.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

174.00	1489.0
175.10	2995.0
176.00	870.0
177.00	1312.0
177.90	414.0
179.00	5108.0
180.00	3993.0
181.00	1574.0
184.00	516.0
185.00	2563.0
186.10	19952.0
187.00	6130.0
188.00	734.0
189.00	1230.0
191.00	599.0
192.00	1615.0
193.00	1997.0
194.00	465.0
196.10	3985.0
196.80	850.0
198.00	145408.0
199.00	10616.0
200.00	764.0
201.60	650.0
203.00	1023.0
204.00	5171.0
205.10	8984.0
206.10	38192.0
207.10	5870.0
208.00	1459.0
209.00	496.0
210.30	568.0
211.00	1493.0
215.00	478.0
216.00	820.0
217.00	9722.0
218.00	1272.0
221.10	5381.0
223.00	2177.0
224.10	21192.0
225.00	5183.0
226.00	619.0
227.00	8350.0
228.00	1154.0
229.00	1836.0
231.00	713.0
234.00	451.0
235.00	548.0
235.90	390.0
237.00	913.0
239.00	327.0
241.10	491.0
242.10	1097.0
243.10	1246.0
244.10	15315.0
245.10	2294.0
246.10	3271.0
247.00	576.0
249.00	477.0
253.00	322.0
254.10	529.0
255.10	79288.0
256.10	11253.0
257.00	1106.0
258.00	4741.0
259.00	831.0
265.00	1890.0
273.00	2379.0
274.10	6167.0
275.10	33424.0
276.10	4828.0
277.10	3082.0
278.00	474.0
283.00	304.0
285.00	578.0
293.00	727.0
296.10	9166.0
297.00	1328.0
303.10	1071.0
304.00	307.0
314.00	445.0
315.00	1000.0
316.00	568.0
321.10	352.0
323.10	3001.0
324.00	650.0
327.00	576.0
333.10	346.0
334.10	1636.0
335.10	452.0
341.10	337.0
346.00	613.0
352.00	757.0
353.00	471.0
354.10	966.0
365.00	3880.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

366.10	558.0
372.10	1352.0
373.00	343.0
383.10	357.0
402.00	504.0
403.10	900.0
404.10	385.0
421.10	839.0
422.10	653.0
423.10	4731.0
424.10	1578.0
441.10	14381.0
442.10	96936.0
443.10	19392.0
444.10	1690.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
38.00	421.0
39.10	2543.0
44.00	383.0
50.10	8673.0
51.10	34552.0
52.00	1780.0
56.00	1064.0
57.00	2449.0
61.00	541.0
62.00	606.0
63.00	1732.0
65.00	803.0
68.00	696.0
69.00	44152.0
73.00	576.0
74.00	4826.0
75.00	7536.0
76.10	2566.0
77.10	54320.0
78.10	3700.0
79.00	3443.0
80.00	2753.0
81.00	4194.0
82.00	951.0
83.10	980.0
85.00	1007.0
86.00	1271.0
87.00	542.0
91.00	812.0
92.00	978.0
93.00	6232.0
94.00	419.0
96.00	336.0
98.00	4733.0
99.00	4310.0
100.00	347.0
101.00	2163.0
103.00	758.0
104.00	1382.0
105.00	1318.0
106.00	471.0
107.00	17864.0
108.00	2710.0
109.10	562.0
110.00	32704.0
111.00	5107.0
112.00	570.0
116.00	927.0
117.00	15157.0
118.00	1269.0
122.00	1177.0
123.00	1747.0
124.00	851.0
124.90	807.0
127.00	65600.0
128.00	4651.0
129.00	25104.0
130.00	2153.0
131.00	427.0
134.00	818.0
135.00	2031.0
136.00	870.0
137.00	1335.0
141.00	3088.0
141.90	1025.0
143.00	789.0
146.00	592.0
147.00	1819.0
148.00	3963.0
149.00	795.0
151.10	440.0
152.90	1049.0
154.00	795.0
155.00	1917.0
156.00	2909.0
157.10	598.0
157.90	622.0
159.00	499.0
160.00	1322.0
161.00	1694.0
161.90	447.0
165.00	1267.0
166.00	991.0
167.00	7217.0
168.00	4170.0
169.00	782.0
172.00	580.0
173.00	757.0
174.00	1670.0
175.10	2659.0
176.10	738.0
177.00	1405.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

178.10	426.0
179.00	5299.0
180.00	3879.0
181.00	1898.0
184.00	441.0
185.10	2606.0
186.00	20008.0
187.00	5787.0
188.10	673.0
189.00	1381.0
191.10	697.0
192.10	1554.0
193.00	2186.0
194.00	417.0
196.00	4083.0
197.00	658.0
198.00	148288.0
199.00	9868.0
200.10	759.0
201.60	723.0
203.00	958.0
204.10	5542.0
205.10	9009.0
206.10	38416.0
207.10	5848.0
208.00	1591.0
209.10	490.0
210.10	589.0
211.00	1530.0
211.80	300.0
215.00	476.0
216.00	726.0
217.00	10375.0
218.00	1322.0
221.10	5964.0
221.80	1541.0
223.00	2234.0
224.10	22080.0
225.10	5519.0
226.10	612.0
227.00	9105.0
228.00	1279.0
229.00	1848.0
230.00	303.0
231.10	779.0
234.00	627.0
235.00	619.0
236.10	431.0
237.00	758.0
239.00	383.0
241.00	504.0
242.00	1206.0
243.10	1235.0
244.10	17592.0
245.10	1921.0
246.00	3423.0
247.00	661.0
249.00	641.0
252.90	305.0
254.10	699.0
255.10	86064.0
256.10	12545.0
257.10	1058.0
258.00	5022.0
259.00	794.0
265.00	2115.0
266.00	339.0
273.10	2560.0
274.10	6941.0
275.10	37152.0
276.10	5087.0
277.00	3067.0
278.10	529.0
283.10	376.0
285.00	535.0
293.00	760.0
296.10	10676.0
297.10	1564.0
303.10	1290.0
304.10	314.0
314.00	528.0
315.00	1095.0
316.10	687.0
321.10	305.0
323.10	3172.0
324.10	671.0
327.00	630.0
328.00	406.0
333.00	350.0
334.10	2273.0
335.10	447.0
341.10	341.0
346.00	779.0
352.10	1059.0
353.10	735.0
354.10	1134.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

365.10	4422.0
366.00	564.0
371.00	309.0
372.10	1811.0
373.10	395.0
383.00	516.0
402.10	659.0
403.10	889.0
404.10	376.0
421.00	841.0
422.10	941.0
423.10	6307.0
424.10	1790.0
441.10	19648.0
442.10	125624.0
443.10	24232.0
444.10	2295.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
39.00	1266.0
44.00	311.0
50.00	4715.0
51.10	18152.0
52.10	886.0
56.00	565.0
57.00	1442.0
63.00	899.0
65.00	464.0
68.00	460.0
69.00	23336.0
74.00	2443.0
75.00	4001.0
76.10	1319.0
77.10	29320.0
78.10	2087.0
79.00	1826.0
80.00	1380.0
81.00	2002.0
82.00	539.0
83.00	582.0
85.00	660.0
86.00	579.0
87.00	340.0
91.00	632.0
92.00	649.0
93.00	3235.0
98.00	2737.0
99.00	2473.0
101.00	1205.0
103.00	450.0
104.00	758.0
105.00	916.0
107.00	9747.0
108.00	1513.0
109.00	345.0
110.00	18528.0
111.00	2910.0
112.00	462.0
116.00	490.0
117.00	8897.0
118.00	700.0
122.00	631.0
123.00	1106.0
123.90	470.0
125.00	406.0
127.00	37896.0
128.00	2883.0
129.00	14294.0
130.00	1147.0
134.00	359.0
135.00	1134.0
136.00	438.0
137.00	758.0
141.00	2065.0
142.00	677.0
143.00	476.0
145.90	330.0
147.00	994.0
148.00	2450.0
149.00	457.0
153.00	641.0
154.00	533.0
155.00	1150.0
156.10	1769.0
157.00	339.0
158.00	359.0
159.90	680.0
161.00	855.0
161.90	303.0
165.00	674.0
166.00	608.0
167.00	4416.0
168.00	2510.0
169.00	392.0
172.00	369.0
173.00	493.0
174.10	798.0
175.00	1658.0
176.10	445.0
177.00	684.0
179.00	3151.0
180.00	2396.0
181.00	986.0
185.10	1725.0
186.00	12132.0
187.00	3477.0
188.10	341.0
189.00	773.0
191.00	435.0
192.00	1157.0
193.00	1344.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

196.00	2367.0
197.00	401.0
198.00	89864.0
199.00	5758.0
200.00	506.0
201.50	338.0
203.00	626.0
204.00	2983.0
205.10	5699.0
206.10	23752.0
207.10	3650.0
208.00	916.0
209.00	446.0
210.00	376.0
211.00	895.0
216.00	547.0
217.00	6003.0
218.00	717.0
221.10	4094.0
221.90	811.0
223.00	1492.0
224.10	12867.0
225.10	3289.0
226.10	416.0
227.00	5744.0
228.10	667.0
229.00	1158.0
231.00	506.0
234.00	429.0
235.10	412.0
236.00	317.0
237.00	606.0
241.00	397.0
242.00	724.0
243.00	788.0
244.10	10896.0
245.10	1384.0
246.00	2102.0
247.00	389.0
249.00	416.0
254.10	348.0
255.10	55536.0
256.10	7326.0
257.00	647.0
258.00	3347.0
259.00	524.0
265.00	1302.0
273.10	1579.0
274.10	4211.0
275.10	23992.0
276.10	3226.0
277.10	1934.0
278.00	317.0
285.00	348.0
293.00	489.0
296.10	6921.0
297.10	891.0
303.10	756.0
314.00	394.0
315.10	660.0
316.00	453.0
323.10	2107.0
324.10	353.0
327.00	481.0
334.10	1689.0
335.10	364.0
346.00	597.0
352.10	646.0
353.10	492.0
354.10	750.0
365.10	3142.0
366.10	511.0
372.10	1228.0
373.10	336.0
383.10	358.0
402.10	540.0
403.10	680.0
421.10	702.0
422.00	643.0
423.10	5072.0
424.10	1362.0
441.10	14845.0
442.10	96336.0
443.10	17784.0
444.10	1679.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47