

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112017\
 Data File : BM013074.D
 Acq On : 21 Nov 2017 09:27
 Operator : SJ/JU
 Sample : I6562-05
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Nov 21 14:01:53 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM111617.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 21 07:39:32 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	179494	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	788990	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.35	164	508971	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.10	188	1187194	20.00	ng/ul	0.00
75) Chrysene-d12	21.28	240	1007295	20.00	ng/ul	0.00
83) Perylene-d12	23.51	264	809680	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	10.52	131	161	0.01	ng/ul	-0.12
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
46) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
51) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
57) Fluorene-d10	15.34	176	1061	0.03	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
70) Anthracene-d10	17.10	188	1187194	19.53	ng/ul	-0.10
76) Pyrene-d10	0.00	212	0	0.00	ng/ul	
87) Benzo(a)pyrene-d12	23.51	264	807060	19.66	ng/ul	0.14

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112017\
Data File : BM013074.D
Acq On : 21 Nov 2017 09:27
Operator : SJ/JU
Sample : I6562-05
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Nov 21 14:01:53 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM111617.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 21 07:39:32 2017
Response via : Initial Calibration

