

Data Path : V:\HPCHEM1\BNA M\DATA\BM112115\
 Data File : BM003241.D
 Acq On : 21 Nov 2015 10:06
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.459

Quant Time: Nov 22 01:13:11 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 04:24:04 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.422	1.2	95	0.00
3	1,4-Dioxane	0.486	0.479	1.4	96	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
5	Naphthalene	1.027	1.048	-2.0	102	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.554	-1.8	102	0.00
7	2-Methylnaphthalene	0.669	0.700	-4.6	104	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
9	Acenaphthylene	1.665	1.882	-13.0	117	0.00
10 C	Acenaphthene	1.406	1.456	-3.6	105	0.00
11	Fluorene	1.521	1.581	-3.9	105	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
13	Pentachlorophenol	0.094	0.098	-4.3	112	0.00
14	Phenanthrene	1.161	1.165	-0.3	103	0.00
15	Anthracene	0.972	1.063	-9.4	114	0.00
16 SURR	Fluoranthene-d10	0.980	1.001	-2.1	108	0.00
17 C	Fluoranthene	1.287	1.323	-2.8	108	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
19	Pyrene	1.371	1.388	-1.2	112	0.00
20	Benzo(a)anthracene	1.160	1.291	-11.3	128	0.00
21	Chrysene	1.306	1.346	-3.1	113	0.00
22 I	Perylene-d12	1.000	1.000	0.0	130	0.00
23	Benzo(b)fluoranthene	1.521	1.444	5.1	125	0.00
24	Benzo(k)fluoranthene	1.438	1.409	2.0	123	0.00
25 C	Benzo(a)pyrene	1.353	1.411	-4.3	135	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.617	-0.3	129	0.00
27	Dibenzo(a,h)anthracene	1.303	1.283	1.5	127	0.00
28	Benzo(q,h,i)perylene	1.416	1.388	2.0	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0