

Data Path : V:\HPCHEM1\BNA M\DATA\BM112115\
 Data File : BM003256.D
 Acq On : 21 Nov 2015 19:18
 Operator : UM/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.460

Quant Time: Nov 22 04:36:14 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 01:13:23 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.418	2.1	107	0.00
3	1,4-Dioxane	0.486	0.473	2.7	108	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
5	Naphthalene	1.027	1.047	-1.9	117	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.555	-2.0	117	0.00
7	2-Methylnaphthalene	0.669	0.701	-4.8	120	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
9	Acenaphthylene	1.665	1.841	-10.6	130	0.00
10 C	Acenaphthene	1.406	1.440	-2.4	118	0.00
11	Fluorene	1.521	1.562	-2.7	117	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
13	Pentachlorophenol	0.094	0.089	5.3	110	0.00
14	Phenanthrene	1.161	1.185	-2.1	114	0.00
15	Anthracene	0.972	1.042	-7.2	122	0.00
16 SURR	Fluoranthene-d10	0.980	0.943	3.8	110	0.00
17 C	Fluoranthene	1.287	1.252	2.7	111	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
19	Pyrene	1.371	1.447	-5.5	116	0.00
20	Benzo(a)anthracene	1.160	1.280	-10.3	127	0.01
21	Chrysene	1.306	1.347	-3.1	113	0.01
22 I	Perylene-d12	1.000	1.000	0.0	128	0.00
23	Benzo(b)fluoranthene	1.521	1.556	-2.3	133	0.00
24	Benzo(k)fluoranthene	1.438	1.295	9.9	111	0.01
25 C	Benzo(a)pyrene	1.353	1.399	-3.4	131	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.627	-0.9	128	0.01
27	Dibenzo(a,h)anthracene	1.303	1.298	0.4	126	0.00
28	Benzo(q,h,i)perylene	1.416	1.391	1.8	124	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0