

Data Path : V:\HPCHEM1\BNA M\DATA\BM112115\
 Data File : BM003277.D
 Acq On : 22 Nov 2015 12:07
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.462

Quant Time: Nov 23 01:20:18 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 01:07:06 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.419	1.9	111	0.00
3	1,4-Dioxane	0.486	0.473	2.7	111	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
5	Naphthalene	1.027	1.044	-1.7	120	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.547	-0.6	118	0.00
7	2-Methylnaphthalene	0.669	0.693	-3.6	122	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
9	Acenaphthylene	1.665	1.863	-11.9	132	0.00
10 C	Acenaphthene	1.406	1.441	-2.5	118	0.00
11	Fluorene	1.521	1.578	-3.7	119	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
13	Pentachlorophenol	0.094	0.093	1.1	120	0.00
14	Phenanthrene	1.161	1.205	-3.8	121	0.00
15	Anthracene	0.972	1.049	-7.9	127	0.00
16 SURR	Fluoranthene-d10	0.980	1.012	-3.3	123	0.00
17 C	Fluoranthene	1.287	1.332	-3.5	123	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
19	Pyrene	1.371	1.372	-0.1	128	0.00
20	Benzo(a)anthracene	1.160	1.301	-12.2	150	0.00
21	Chrysene	1.306	1.319	-1.0	129	0.00
22 I	Perylene-d12	1.000	1.000	0.0	151#	0.00
23	Benzo(b)fluoranthene	1.521	1.422	6.5	143	0.00
24	Benzo(k)fluoranthene	1.438	1.398	2.8	141	0.00
25 C	Benzo(a)pyrene	1.353	1.399	-3.4	155#	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.591	1.3	148	0.00
27	Dibenzo(a,h)anthracene	1.303	1.266	2.8	146	0.00
28	Benzo(q,h,i)perylene	1.416	1.349	4.7	142	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0