

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112116\
 Data File : BM007961.D
 Acq On : 21 Nov 2016 16:56
 Operator : UM/SJ
 Sample : PB94803BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK03

Quant Time: Nov 21 17:52:32 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 15:11:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	562	0.40	ng/ul	0.00
2) Naphthalene-d8	10.50	136	1873	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.35	164	1195	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.19	188	188662	0.40	ng/ul	0.09
16) Chrysene-d12	21.31	240	2415	0.40	ng/ul	0.01
20) Perylene-d12	0.00	264	0	0.00	ng/ul	-23.56

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.12	152	950	0.36	ng/ul	0.00
14) Fluoranthene-d10	19.14	212	2253	0.00	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.55	128	8	0.00	ng/ul#	1
5) 2-Methylnaphthalene	12.19	142	2	0.00	ng/ul#	4
7) Acenaphthylene	14.04	152	14	0.00	ng/ul#	1
8) Acenaphthene	14.42	153	6	0.00	ng/ul#	68
9) Fluorene	15.36	166	383	0.08	ng/ul#	27
11) Pentachlorophenol	16.85	266	5	0.00	ng/ul#	20
12) Phenanthrene	17.19	178	115	0.00	ng/ul#	1
13) Anthracene	17.19	178	103	0.00	ng/ul#	1
15) Fluoranthene	19.22	202	1	0.00	ng/ul#	1
17) Pyrene	19.50	202	379	0.05	ng/ul#	1
18) Benzo(a)anthracene	21.36	228	268	0.04	ng/ul#	49
19) Chrysene	21.36	228	280	0.03	ng/ul#	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112116\
Data File : BM007961.D
Acq On : 21 Nov 2016 16:56
Operator : UM/SJ
Sample : PB94803BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK03

Quant Time: Nov 21 17:52:32 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM112116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 21 15:11:37 2016
Response via : Initial Calibration

