

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112116\
 Data File : BM007958.D
 Acq On : 21 Nov 2016 15:10
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.444

Quant Time: Nov 21 15:49:27 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 15:11:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
3	Naphthalene	1.119	1.093	2.3	91	0.00
4 SURR	2-Methylnaphthalene-d10	0.566	0.547	3.4	90	0.00
5	2-Methylnaphthalene	0.721	0.699	3.1	91	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
7	Acenaphthylene	1.810	1.822	-0.7	92	0.00
8 C	Acenaphthene	1.377	1.374	0.2	93	0.00
9	Fluorene	1.583	1.537	2.9	90	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.140	0.126	10.0	85	0.00
12	Phenanthrene	1.246	1.242	0.3	92	0.00
13	Anthracene	1.170	1.144	2.2	89	0.00
14 SURR	Fluoranthene-d10	1.132	1.101	2.7	90	0.00
15 C	Fluoranthene	1.608	1.522	5.3	89	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
17	Pyrene	1.389	1.375	1.0	88	0.00
18	Benzo(a)anthracene	1.228	1.227	0.1	92	0.00
19	Chrysene	1.527	1.425	6.7	86	0.00
20 I	Perylene-d12	1.000	1.000	0.0	86	0.00
21	Benzo(b)fluoranthene	1.596	1.496	6.3	82	0.00
22	Benzo(k)fluoranthene	1.647	1.693	-2.8	92	0.00
23 C	Benzo(a)pyrene	1.495	1.473	1.5	87	0.00
24	Indeno(1,2,3-cd)pyrene	1.848	1.798	2.7	84	0.00
25	Dibenzo(a,h)anthracene	1.481	1.432	3.3	84	0.00
26	Benzo(g,h,i)perylene	1.640	1.610	1.8	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0