

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
 Data File : BM037609.D
 Acq On : 21 Nov 2022 13:45
 Operator : CG/JU
 Sample : N5690-17
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGCE3

Quant Time: Nov 21 14:51:16 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 18 06:51:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.133	152	7020	0.400	ng/ul	0.00
4) Naphthalene-d8	10.953	136	20547	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.752	164	12634	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.483	188	26298	0.400	ng/ul	0.00
17) Chrysene-d12	21.639	240	17961	0.400	ng/ul	# 0.00
23) Perylene-d12	24.130	264	19555	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.450	96	39639	4.857	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.526	152	10264	0.301	ng/ul	0.00
18) Fluoranthene-d10	19.495	212	24418	0.411	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.488	88	2056	0.242	ng/ul#	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
Data File : BM037609.D
Acq On : 21 Nov 2022 13:45
Operator : CG/JU
Sample : N5690-17
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGCE3

Quant Time: Nov 21 14:51:16 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 18 06:51:41 2022
Response via : Initial Calibration

