

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
 Data File : BM037625.D
 Acq On : 22 Nov 2022 00:11
 Operator : CG/JU
 Sample : PB149135BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS135

Quant Time: Nov 22 01:37:13 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 22:36:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.129	152	4185	0.400	ng/ul	0.00
4) Naphthalene-d8	10.948	136	11648	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.747	164	6856	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.479	188	15518	0.400	ng/ul	0.00
17) Chrysene-d12	21.636	240	11402	0.400	ng/ul	# 0.00
23) Perylene-d12	24.124	264	10381	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.450	96	4489	0.923	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.526	152	8363	0.433	ng/ul	0.00
18) Fluoranthene-d10	19.490	212	18368	0.487	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.488	88	12070	2.383	ng/ul#	50
5) Naphthalene	10.997	128	16620	0.470	ng/ul	98
7) 2-Methylnaphthalene	12.597	142	10659	0.443	ng/ul	100
8) 1-Methylnaphthalene	12.812	142	11409	0.463	ng/ul	99
10) Acenaphthylene	14.470	152	15978	0.420	ng/ul	100
11) Acenaphthene	14.807	153	12835	0.483	ng/ul	98
12) Fluorene	15.788	166	15075	0.471	ng/ul	100
14) Pentachlorophenol	17.124	266	3490	1.002	ng/ul	100
15) Phenanthrene	17.521	178	25794	0.486	ng/ul	99
16) Anthracene	17.614	178	22498	0.440	ng/ul	99
19) Fluoranthene	19.523	202	30030	0.546	ng/ul	97
20) Pyrene	19.881	202	30752	0.538	ng/ul	97
21) Benzo(a)anthracene	21.619	228	26550	0.521	ng/ul	99
22) Chrysene	21.674	228	28484	0.563	ng/ul	99
24) Benzo(b)fluoranthene	23.358	252	29752	0.628	ng/ul	92
25) Benzo(k)fluoranthene	23.411	252	28227	0.591	ng/ul#	93
26) Benzo(a)pyrene	24.010	252	24605	0.598	ng/ul#	86
27) Indeno(1,2,3-cd)pyrene	26.728	276	30625	0.562	ng/ul#	92
28) Dibenzo(a,h)anthracene	26.748	278	23834	0.556	ng/ul	95
29) Benzo(g,h,i)perylene	27.533	276	25594	0.553	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
Data File : BM037625.D
Acq On : 22 Nov 2022 00:11
Operator : CG/JU
Sample : PB149135BS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS135

Quant Time: Nov 22 01:37:13 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 21 22:36:32 2022
Response via : Initial Calibration

