

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
 Data File : BM037658.D
 Acq On : 23 Nov 2022 01:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4178

Quant Time: Nov 23 03:18:31 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 22:36:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.125	152	4988	0.400	ng/ul	0.00
4) Naphthalene-d8	10.948	136	14411	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.747	164	8607	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.479	188	19311	0.400	ng/ul	0.00
17) Chrysene-d12	21.636	240	14741	0.400	ng/ul	# 0.00
23) Perylene-d12	24.121	264	13768	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.450	96	2664	0.459	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.520	152	9461	0.396	ng/ul	0.00
18) Fluoranthene-d10	19.490	212	20619	0.423	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.488	88	2592	0.429	ng/ul#	55
5) Naphthalene	10.997	128	18710	0.427	ng/ul	99
7) 2-Methylnaphthalene	12.592	142	12119	0.407	ng/ul	100
8) 1-Methylnaphthalene	12.806	142	12341	0.405	ng/ul	100
10) Acenaphthylene	14.465	152	18185	0.381	ng/ul	100
11) Acenaphthene	14.807	153	14354	0.430	ng/ul	98
12) Fluorene	15.784	166	16720	0.416	ng/ul	98
14) Pentachlorophenol	17.120	266	2085	0.481	ng/ul	98
15) Phenanthrene	17.517	178	28234	0.427	ng/ul	99
16) Anthracene	17.610	178	24830	0.390	ng/ul	99
19) Fluoranthene	19.518	202	32078	0.451	ng/ul#	95
20) Pyrene	19.880	202	32995	0.446	ng/ul	98
21) Benzo(a)anthracene	21.619	228	28485	0.432	ng/ul	99
22) Chrysene	21.674	228	30468	0.466	ng/ul	99
24) Benzo(b)fluoranthene	23.358	252	32444	0.517	ng/ul	91
25) Benzo(k)fluoranthene	23.408	252	29970	0.473	ng/ul#	94
26) Benzo(a)pyrene	24.010	252	27647	0.506	ng/ul#	85
27) Indeno(1,2,3-cd)pyrene	26.728	276	35800	0.495	ng/ul#	90
28) Dibenzo(a,h)anthracene	26.745	278	27469	0.483	ng/ul	95
29) Benzo(g,h,i)perylene	27.530	276	28917	0.471	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
Data File : BM037658.D
Acq On : 23 Nov 2022 01:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4178

Quant Time: Nov 23 03:18:31 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 21 22:36:32 2022
Response via : Initial Calibration

