

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
 Data File : BM037674.D
 Acq On : 23 Nov 2022 11:19
 Operator : CG/JU
 Sample : PB149045BS
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS045

Quant Time: Nov 24 01:48:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 22:36:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.125	152	6871	0.400	ng/ul	0.00
4) Naphthalene-d8	10.942	136	19732	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.743	164	11927	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.475	188	26507	0.400	ng/ul	0.00
17) Chrysene-d12	21.636	240	17813	0.400	ng/ul	# 0.00
23) Perylene-d12	24.124	264	14944	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.450	96	5563	0.696	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.520	152	11413	0.349	ng/ul	0.00
18) Fluoranthene-d10	19.490	212	24572	0.417	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.484	88	14116	1.698	ng/ul#	52
5) Naphthalene	10.997	128	21255	0.355	ng/ul	98
7) 2-Methylnaphthalene	12.592	142	13817	0.339	ng/ul	100
8) 1-Methylnaphthalene	12.806	142	14834	0.355	ng/ul	100
10) Acenaphthylene	14.465	152	20998	0.317	ng/ul	99
11) Acenaphthene	14.807	153	16658	0.360	ng/ul	98
12) Fluorene	15.783	166	19593	0.352	ng/ul	100
14) Pentachlorophenol	17.120	266	4320	0.726	ng/ul	100
15) Phenanthrene	17.517	178	32515	0.359	ng/ul	99
16) Anthracene	17.610	178	28466	0.326	ng/ul	100
19) Fluoranthene	19.518	202	36066	0.420	ng/ul#	95
20) Pyrene	19.880	202	36506	0.409	ng/ul	97
21) Benzo(a)anthracene	21.619	228	29746	0.374	ng/ul	99
22) Chrysene	21.674	228	31678	0.401	ng/ul	100
24) Benzo(b)fluoranthene	23.358	252	30895	0.453	ng/ul	92
25) Benzo(k)fluoranthene	23.408	252	28550	0.415	ng/ul#	93
26) Benzo(a)pyrene	24.010	252	25197	0.425	ng/ul#	86
27) Indeno(1,2,3-cd)pyrene	26.724	276	30385	0.387	ng/ul#	90
28) Dibenzo(a,h)anthracene	26.741	278	23525	0.381	ng/ul	95
29) Benzo(g,h,i)perylene	27.526	276	25408	0.381	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
Data File : BM037674.D
Acq On : 23 Nov 2022 11:19
Operator : CG/JU
Sample : PB149045BS
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS045

Quant Time: Nov 24 01:48:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 21 22:36:32 2022
Response via : Initial Calibration

