

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
 Data File : BM037622.D
 Acq On : 21 Nov 2022 22:21
 Operator : CG/JU
 Sample : PB149198BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK198

Quant Time: Nov 22 00:10:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 21 22:36:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.129	152	6823	0.400	ng/u1	0.00
4) Naphthalene-d8	10.948	136	20249	0.400	ng/u1	# 0.00
9) Acenaphthene-d10	14.747	164	12196	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.479	188	25258	0.400	ng/u1	0.00
17) Chrysene-d12	21.637	240	17419	0.400	ng/u1	# 0.00
23) Perylene-d12	24.124	264	18712	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.450	96	47765	6.021	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.526	152	8979	0.267	ng/u1	0.00
18) Fluoranthene-d10	19.495	212	18479	0.321	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112122\
Data File : BM037622.D
Acq On : 21 Nov 2022 22:21
Operator : CG/JU
Sample : PB149198BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK198

Quant Time: Nov 22 00:10:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 21 22:36:32 2022
Response via : Initial Calibration

