




Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112216\
Data File : BM008001.D
Acq On : 22 Nov 2016 19:11
Operator : UM/SJ
Sample : H5695-11
Misc :
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Nov 23 01:00:28 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM112116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 22 16:05:09 2016
Response via : Initial Calibration



TIC: BM008001.D

(12) Phenanthrene

17.228min (+0.068) 0.43ng/ul

response 4305

Ion	Exp%	Act%
178.00	100	100
176.00	23.00	20.43
179.00	17.00	21.83#
0.00	0.00	0.00