




Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112216\
Data File : BM008001.D
Acq On : 22 Nov 2016 19:11
Operator : UM/SJ
Sample : H5695-11
Misc :
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Nov 23 01:00:28 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM112116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 22 16:05:09 2016
Response via : Initial Calibration



TIC: BM008001.D

(23) Benzo(a)pyrene (C)

23.576min (+0.125) 0.37ng/ul

response 4601

Ion	Exp%	Act%
252.00	100	100
253.00	35.60	33.39
125.00	22.60	19.62
0.00	0.00	0.00