


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM112216\
Data File : BM007985.D
Acq On : 22 Nov 2016 08:46
Operator : UM/SJ
Sample : DFTPP53
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
Title : SVOA CALIBRATION
Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2541, 2542, 2543; Background Corrected with Scan 2535

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.3	43376	PASS
68	69	0.00	2	1.8	896	PASS
69	198	0.00	100	35.3	50562	PASS
70	69	0.00	2	0.2	105	PASS
127	198	10	80	40.9	58578	PASS
197	198	0.00	2	0.4	523	PASS
198	198	100	100	100.0	143258	PASS
199	198	5	9	6.5	9333	PASS
275	198	10	60	30.4	43557	PASS
365	198	1	100	4.4	6281	PASS
441	443	0.01	100	78.3	19213	PASS
442	198	50	100	88.4	126642	PASS
443	442	15	24	19.4	24552	PASS

m/z	Abundance
38.00	812.0
39.10	3604.0
40.00	622.0
44.00	1008.0
49.20	414.0
50.10	13341.0
51.10	54080.0
52.10	2861.0
54.90	324.0
56.00	1520.0
57.00	3413.0
58.10	400.0
61.00	886.0
62.00	872.0
63.10	2607.0
64.10	354.0
65.10	952.0
68.10	1371.0
69.00	63552.0
70.00	315.0
73.10	1066.0
74.10	7015.0
75.00	11092.0
76.10	3607.0
77.10	73496.0
78.10	4918.0
79.00	4956.0
80.00	4040.0
81.00	5906.0
82.00	1748.0
83.10	1122.0
85.00	695.0
86.10	1922.0
87.00	681.0
91.00	1451.0
92.00	1859.0
93.00	8542.0
94.00	654.0
98.00	7833.0
99.00	6318.0
99.90	740.0
101.00	2131.0
102.90	873.0
104.10	1936.0
105.00	1881.0
106.00	668.0
107.00	23168.0
108.00	3787.0
109.00	673.0
110.00	38704.0
111.10	5735.0
112.00	873.0
116.00	1367.0
117.00	21600.0
118.00	1622.0
122.00	1197.0
123.10	2286.0
124.10	1028.0
125.10	788.0
126.00	334.0
127.00	70184.0
128.00	5659.0
129.00	30408.0
130.00	2865.0
131.00	486.0
134.10	732.0
135.10	2446.0
136.00	949.0
137.00	1423.0
138.00	399.0
141.00	4063.0
142.00	1367.0
143.00	816.0
144.00	350.0
145.90	564.0
147.10	2034.0
148.00	5198.0
148.90	1499.0
149.80	332.0
151.00	496.0
153.00	1068.0
153.90	1037.0
155.00	2168.0
156.10	3125.0
157.00	652.0
157.90	641.0
159.00	358.0
160.00	1248.0
161.00	2054.0
162.00	512.0
163.90	336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

165.00	1446.0
166.00	1169.0
167.10	8142.0
168.00	5655.0
169.00	687.0
172.00	618.0
173.00	1122.0
174.10	1783.0
175.00	3351.0
176.00	565.0
177.00	1732.0
178.10	702.0
178.90	6109.0
180.00	4953.0
181.00	1937.0
182.20	439.0
184.00	766.0
185.00	3028.0
186.00	22736.0
187.00	6041.0
188.00	926.0
189.00	2191.0
190.00	481.0
191.00	885.0
192.00	2065.0
193.00	2589.0
194.10	537.0
196.00	4990.0
196.90	935.0
198.00	161280.0
199.00	10624.0
200.10	911.0
201.50	531.0
203.00	1448.0
204.00	6119.0
205.00	11107.0
206.00	45288.0
207.10	7455.0
208.00	1803.0
209.00	777.0
210.00	886.0
211.10	1695.0
212.10	305.0
215.00	701.0
216.00	847.0
217.00	12496.0
218.00	1402.0
221.00	5823.0
222.00	1451.0
223.00	2253.0
224.00	26400.0
225.00	5919.0
226.00	1189.0
227.00	10525.0
227.90	1760.0
229.00	2210.0
229.90	343.0
231.00	1073.0
234.00	659.0
235.10	784.0
236.10	581.0
236.90	1356.0
240.00	493.0
241.00	825.0
242.10	1335.0
243.10	1694.0
244.10	18576.0
245.00	2915.0
246.00	4465.0
247.00	845.0
248.90	811.0
253.00	436.0
254.00	583.0
255.00	103304.0
256.00	14825.0
257.10	1530.0
258.00	6886.0
258.90	929.0
265.00	2692.0
270.90	333.0
273.00	3188.0
274.00	7626.0
275.00	46328.0
276.00	6122.0
277.00	4071.0
278.00	1050.0
281.00	573.0
283.00	514.0
285.00	615.0
292.90	778.0
295.00	361.0
296.00	13384.0
297.00	2272.0
303.10	1626.0
304.00	375.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

314.00	736.0
314.90	1569.0
316.10	1089.0
321.00	349.0
323.10	4014.0
324.00	702.0
327.00	1011.0
328.00	538.0
332.00	392.0
332.90	377.0
334.00	2386.0
335.00	855.0
340.90	402.0
342.00	319.0
346.00	889.0
352.10	1562.0
353.10	787.0
354.10	1439.0
355.20	367.0
365.00	5796.0
366.00	1110.0
372.00	2194.0
372.80	471.0
383.00	375.0
402.10	807.0
403.10	879.0
404.10	398.0
421.00	937.0
422.00	755.0
423.00	5566.0
424.10	2614.0
441.10	16920.0
442.10	110080.0
443.10	20864.0
444.10	1828.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
38.00	584.0
39.10	3297.0
40.00	537.0
43.90	884.0
49.10	441.0
50.10	13433.0
51.10	51256.0
52.10	2191.0
55.00	484.0
56.00	1515.0
57.00	3420.0
61.00	640.0
62.00	741.0
63.10	2125.0
65.10	1029.0
68.00	857.0
69.00	59712.0
73.00	1183.0
74.10	6480.0
75.00	8771.0
76.10	3489.0
77.10	65040.0
78.10	4676.0
79.10	5595.0
80.00	3407.0
81.00	5440.0
81.90	1142.0
83.00	1071.0
85.00	726.0
86.00	1777.0
87.00	686.0
88.10	303.0
91.00	1272.0
92.00	1184.0
93.00	8905.0
93.90	588.0
95.90	561.0
98.00	6376.0
99.00	6284.0
100.00	463.0
101.00	3096.0
103.00	1114.0
104.10	1574.0
105.00	2198.0
106.10	625.0
107.00	21728.0
108.00	3468.0
109.00	643.0
110.00	38088.0
111.00	5749.0
112.00	944.0
116.00	1105.0
117.00	20600.0
118.00	1494.0
118.80	328.0
120.10	301.0
122.00	1459.0
123.00	2886.0
123.90	1068.0
125.00	871.0
126.20	351.0
127.00	67176.0
128.00	5458.0
129.00	28680.0
130.00	2655.0
131.10	648.0
134.00	956.0
135.00	2507.0
136.00	1151.0
137.00	1737.0
138.10	401.0
140.00	374.0
141.00	3780.0
142.00	1325.0
143.00	749.0
144.00	309.0
146.00	529.0
147.00	2733.0
148.00	5484.0
149.00	1365.0
150.00	399.0
151.30	415.0
153.00	1611.0
154.00	1345.0
155.00	2355.0
156.00	3506.0
157.00	635.0
158.00	796.0
159.00	330.0
160.00	1295.0
161.00	2464.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

161.90	509.0
164.90	1498.0
166.00	1487.0
167.00	9038.0
168.10	5508.0
169.10	1013.0
170.00	466.0
170.90	381.0
172.00	966.0
172.90	1040.0
174.00	1922.0
175.00	2662.0
176.00	838.0
177.00	1854.0
178.00	912.0
179.00	6683.0
180.00	5762.0
181.00	2041.0
182.00	596.0
184.00	529.0
185.00	3176.0
186.00	22816.0
187.10	8232.0
188.10	551.0
189.00	1973.0
189.90	332.0
190.90	1113.0
192.00	2336.0
193.00	2459.0
194.00	386.0
196.10	4342.0
197.10	1003.0
198.00	168448.0
199.00	10827.0
200.10	957.0
201.70	833.0
203.00	1208.0
204.00	7039.0
205.00	12067.0
206.10	46648.0
207.10	8071.0
207.90	1845.0
209.00	1024.0
210.00	710.0
211.10	2218.0
211.90	523.0
215.00	621.0
216.00	1080.0
217.00	13522.0
218.00	2009.0
220.30	499.0
221.00	6784.0
221.90	1497.0
223.00	3255.0
224.00	26952.0
225.10	6769.0
226.00	823.0
227.00	12071.0
228.10	1970.0
229.00	2969.0
230.10	373.0
231.00	788.0
234.00	805.0
235.00	1099.0
236.10	556.0
237.00	1855.0
239.00	666.0
240.00	394.0
240.90	994.0
242.00	1665.0
243.00	1210.0
244.10	19512.0
245.00	2729.0
246.00	5098.0
246.90	879.0
248.00	314.0
248.90	691.0
251.00	329.0
253.00	638.0
254.00	747.0
255.00	119416.0
256.00	16293.0
257.10	1521.0
258.00	7542.0
259.10	1374.0
265.00	3328.0
266.00	459.0
272.00	400.0
273.00	3963.0
274.00	7742.0
275.00	51256.0
276.00	6494.0
277.00	4868.0
277.90	956.0
281.00	682.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

283.00	547.0
285.10	1130.0
289.00	324.0
292.90	953.0
293.90	464.0
294.90	491.0
296.00	16504.0
297.00	2433.0
302.00	330.0
303.10	1653.0
303.90	385.0
314.00	860.0
315.00	1938.0
316.00	897.0
321.90	310.0
323.00	5149.0
324.10	1033.0
326.90	1021.0
328.10	342.0
331.90	534.0
333.10	567.0
334.10	3028.0
335.00	893.0
340.90	509.0
345.90	978.0
352.00	1651.0
353.00	1394.0
354.00	1642.0
355.00	643.0
365.00	7871.0
366.00	1007.0
371.10	429.0
372.10	2245.0
373.10	766.0
383.00	836.0
402.00	1314.0
403.10	1423.0
403.90	500.0
421.00	1114.0
422.10	997.0
423.10	7663.0
424.10	3207.0
425.00	532.0
441.10	22488.0
442.10	153600.0
443.10	31416.0
444.00	2362.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
39.10	1881.0
40.00	564.0
44.00	927.0
49.00	342.0
50.10	6347.0
51.10	24792.0
52.00	1420.0
56.00	799.0
57.10	1665.0
60.90	463.0
62.10	400.0
63.10	1257.0
64.90	435.0
68.10	461.0
69.00	28424.0
73.00	1139.0
74.10	3676.0
75.10	5009.0
76.10	1731.0
77.10	34520.0
78.00	2060.0
79.00	2636.0
79.90	1972.0
81.00	2774.0
82.00	546.0
83.10	946.0
84.90	545.0
86.00	1000.0
87.00	394.0
91.00	611.0
92.10	669.0
93.00	4357.0
96.00	384.0
98.00	3204.0
99.00	2778.0
101.00	1334.0
102.80	447.0
104.00	840.0
105.10	1047.0
107.00	11723.0
108.00	1785.0
109.10	335.0
110.00	19712.0
111.00	2766.0
111.90	387.0
115.90	554.0
117.00	11551.0
118.00	1010.0
121.90	619.0
123.00	1129.0
123.90	575.0
124.90	372.0
127.00	38376.0
128.00	2694.0
129.00	16070.0
130.00	1373.0
134.00	408.0
135.10	1550.0
136.10	681.0
137.10	850.0
140.00	424.0
141.00	2242.0
142.00	704.0
143.00	410.0
146.00	541.0
147.00	1680.0
148.10	2748.0
149.10	521.0
150.90	463.0
153.00	800.0
153.90	451.0
155.00	1417.0
156.10	1957.0
157.20	382.0
157.90	391.0
160.00	1156.0
161.00	1067.0
162.00	375.0
165.00	1255.0
166.00	465.0
167.00	5536.0
168.00	4025.0
169.10	595.0
171.80	326.0
173.00	522.0
173.90	724.0
175.10	2094.0
176.00	737.0
177.00	864.0
178.00	333.0
179.00	4263.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

180.00	3139.0
181.00	1023.0
183.90	374.0
185.00	2595.0
186.10	14375.0
187.00	3977.0
188.00	516.0
189.00	1134.0
191.10	411.0
191.90	1251.0
193.00	1903.0
194.00	634.0
196.00	3094.0
197.10	567.0
198.00	100048.0
199.00	6549.0
200.00	498.0
201.40	624.0
203.00	1109.0
204.10	3554.0
205.00	6410.0
206.10	27592.0
207.10	4744.0
208.00	1823.0
209.00	591.0
210.40	364.0
211.00	1150.0
215.00	396.0
215.90	692.0
217.00	8438.0
218.00	781.0
221.00	3836.0
222.00	960.0
223.00	2012.0
224.10	17024.0
225.00	4660.0
226.00	604.0
227.00	7430.0
228.00	1446.0
228.90	1541.0
231.10	564.0
234.00	479.0
235.00	560.0
236.00	471.0
236.90	1041.0
239.90	317.0
241.00	343.0
242.00	889.0
243.10	1036.0
244.10	13448.0
245.00	1970.0
246.00	3104.0
247.00	614.0
249.00	338.0
251.00	326.0
252.80	358.0
254.10	539.0
255.00	71728.0
256.00	9486.0
257.00	899.0
258.00	4803.0
258.90	906.0
265.00	1642.0
271.00	334.0
273.00	2326.0
274.00	6050.0
275.00	33088.0
276.10	4177.0
277.00	3298.0
278.00	759.0
281.10	728.0
283.00	541.0
284.00	305.0
285.10	687.0
293.00	936.0
295.00	451.0
296.00	10963.0
297.00	2069.0
303.00	1315.0
314.10	625.0
314.90	1195.0
316.00	675.0
321.00	463.0
323.00	3490.0
324.00	804.0
326.90	881.0
328.00	300.0
332.90	321.0
334.00	2161.0
335.00	548.0
341.10	465.0
346.10	863.0
352.00	1243.0
353.00	782.0
354.00	986.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

365.00	5176.0
366.00	531.0
371.10	352.0
371.90	2001.0
373.00	429.0
383.00	546.0
402.10	693.0
403.10	1180.0
404.10	490.0
421.00	816.0
422.10	746.0
423.10	6324.0
424.00	1928.0
441.10	18232.0
442.10	116248.0
443.00	21376.0
444.00	1901.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

Scan 2535 (17.498 min): BM007985.D
DFTPP53

m/z	Abundance
40.10	499.0
44.00	877.0
73.00	428.0
132.90	378.0
190.90	309.0
207.00	1676.0
208.10	306.0
281.00	414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53