

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112316\
 Data File : BM008080.D
 Acq On : 24 Nov 2016 18:38
 Operator : UM/SJ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.455

Quant Time: Nov 25 01:30:42 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM112116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 25 01:27:01 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	136	0.00
3	Naphthalene	1.119	1.106	1.2	133	0.00
4 SURR	2-Methylnaphthalene-d10	0.566	0.580	-2.5	139	0.00
5	2-Methylnaphthalene	0.721	0.738	-2.4	139	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
7	Acenaphthylene	1.810	1.889	-4.4	147	0.00
8 C	Acenaphthene	1.377	1.349	2.0	141	0.00
9	Fluorene	1.583	1.552	2.0	140	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
11	Pentachlorophenol	0.140	0.135	3.6	139	0.00
12	Phenanthrene	1.246	1.232	1.1	139	0.00
13	Anthracene	1.170	1.186	-1.4	141	-0.02
14 SURR	Fluoranthene-d10	1.132	1.080	4.6	134	0.00
15 C	Fluoranthene	1.608	1.501	6.7	134	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
17	Pyrene	1.389	1.661	-19.6	136	0.00
18	Benzo(a)anthracene	1.228	1.453	-18.3	138	0.00
19	Chrysene	1.527	1.421	6.9	108	0.00
20 I	Perylene-d12	1.000	1.000	0.0	119	0.00
21	Benzo(b)fluoranthene	1.596	1.524	4.5	116	0.00
22	Benzo(k)fluoranthene	1.647	1.535	6.8	115	0.00
23 C	Benzo(a)pyrene	1.495	1.473	1.5	120	-0.01
24	Indeno(1,2,3-cd)pyrene	1.848	1.755	5.0	113	0.00
25	Dibenzo(a,h)anthracene	1.481	1.375	7.2	111	0.00
26	Benzo(a,h,i)perylene	1.640	1.507	8.1	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0