

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112415\
 Data File : BM003288.D
 Acq On : 24 Nov 2015 15:43
 Operator : UM/IZ
 Sample : SSTDCCC002
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD00264

Quant Time: Nov 25 04:08:10 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-1,4-DIO SIM-BM110915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 25 04:04:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 SURR	1,4-Dioxane-d8	0.391	0.393	-0.5	108	0.00
3	1,4-Dioxane	0.466	0.458	1.7	106	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
5 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
6 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
7 I	Chrysene-d12	1.000	1.000	0.0	135	0.00
8 I	Perylene-d12	1.000	1.000	0.0	150#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0