

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112415\
Data File : BM003295.D
Acq On : 24 Nov 2015 19:56
Operator : UM/IZ
Sample : SSTDCCC002EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP51

Quant Time: Nov 25 04:18:56 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-1,4-DIO SIM-BM110915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 25 04:16:44 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 SURR	1,4-Dioxane-d8	0.391	0.391	0.0	95	0.00
3	1,4-Dioxane	0.466	0.458	1.7	94	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
5 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
6 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
7 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
8 I	Perylene-d12	1.000	1.000	0.0	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0