

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
 Data File : BM033242.D
 Acq On : 24 Nov 2021 11:01
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Nov 24 14:39:53 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 14:37:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	115285	20.000	ng	0.00
21) Naphthalene-d8	10.736	136	465912	20.000	ng	# 0.00
39) Acenaphthene-d10	14.560	164	294636	20.000	ng	0.00
64) Phenanthrene-d10	17.295	188	608192	20.000	ng	# 0.00
76) Chrysene-d12	21.453	240	582869	20.000	ng	0.00
86) Perylene-d12	23.789	264	564369	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.507	112	159534	24.125	ng	0.00
7) Phenol-d6	7.090	99	213038	21.634	ng	0.00
23) Nitrobenzene-d5	9.101	82	218898	25.715	ng	0.00
42) 2,4,6-Tribromophenol	16.042	330	68051	20.623	ng	0.00
45) 2-Fluorobiphenyl	13.183	172	468936	24.685	ng	0.00
79) Terphenyl-d14	19.901	244	671655	25.158	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.425	88	34041	12.064	ng	# 76
3) Pyridine	3.837	79	81255	10.488	ng	# 69
4) n-Nitrosodimethylamine	3.743	42	42024	12.782	ng	# 63
6) Aniline	7.266	93	119256	10.846	ng	99
8) 2-Chlorophenol	7.501	128	84647	10.986	ng	100
9) Benzaldehyde	7.084	77	75522	13.588	ng	99
10) Phenol	7.119	94	107327	11.355	ng	95
11) bis(2-Chloroethyl)ether	7.360	93	84332	11.301	ng	99
12) 1,3-Dichlorobenzene	7.825	146	99239	11.657	ng	97
13) 1,4-Dichlorobenzene	7.972	146	101725	11.664	ng	98
14) 1,2-Dichlorobenzene	8.290	146	96529	11.247	ng	98
15) Benzyl Alcohol	8.178	79	81565	11.636	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.466	45	152126	14.182	ng	94
17) 2-Methylphenol	8.372	107	72871	10.975	ng	92
18) Hexachloroethane	9.019	117	35296	11.445	ng	92
19) n-Nitroso-di-n-propyla...	8.742	70	70601	11.710	ng	94
20) 3+4-Methylphenols	8.695	107	98025	10.787	ng	98
22) Acetophenone	8.760	105	132576	11.649	ng	# 98
24) Nitrobenzene	9.142	77	103975	12.748	ng	97
25) Isophorone	9.666	82	168084	11.174	ng	99
26) 2-Nitrophenol	9.848	139	36136	9.490	ng	94
27) 2,4-Dimethylphenol	9.901	122	65879	10.515	ng	99
28) bis(2-Chloroethoxy)met...	10.142	93	111886	11.133	ng	98
29) 2,4-Dichlorophenol	10.378	162	74238	10.332	ng	99
30) 1,2,4-Trichlorobenzene	10.595	180	91451	11.452	ng	99
31) Naphthalene	10.784	128	271545	11.228	ng	99
32) Benzoic acid	9.972	122	25552	5.043	ng	90
33) 4-Chloroaniline	10.895	127	100828	9.984	ng	99
34) Hexachlorobutadiene	11.060	225	56169	11.571	ng	97
35) Caprolactam	11.672	113	12869	5.341	ng	96
36) 4-Chloro-3-methylphenol	12.001	107	83472	10.156	ng	99
37) 2-Methylnaphthalene	12.389	142	183453	10.793	ng	97
38) 1-Methylnaphthalene	12.389	142	183453	10.913	ng	95
40) 1,2,4,5-Tetrachloroben...	12.754	216	97991	12.030	ng	# 96
41) Hexachlorocyclopentadiene	12.730	237	44069	10.302	ng	96
43) 2,4,6-Trichlorophenol	12.989	196	59237	10.392	ng	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
 Data File : BM033242.D
 Acq On : 24 Nov 2021 11:01
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Nov 24 14:39:53 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 14:37:22 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.060	196	66489	10.555	ng	96
46) 1,1'-Biphenyl	13.395	154	250237	11.753	ng	99
47) 2-Chloronaphthalene	13.436	162	188510	11.532	ng	99
48) 2-Nitroaniline	13.642	65	55031	11.458	ng	95
49) Acenaphthylene	14.283	152	287138	10.948	ng	100
50) Dimethylphthalate	14.013	163	232441	10.843	ng	99
51) 2,6-Dinitrotoluene	14.136	165	44374	10.127	ng	# 77
52) Acenaphthene	14.624	154	177173	11.341	ng	99
53) 3-Nitroaniline	14.466	138	43932	9.212	ng	99
54) 2,4-Dinitrophenol	14.666	184	18753	7.092	ng	# 77
55) Dibenzofuran	14.954	168	299789	11.431	ng	94
56) 4-Nitrophenol	14.760	139	32654	8.403	ng	# 88
57) 2,4-Dinitrotoluene	14.919	165	54254	9.140	ng	# 75
58) Fluorene	15.601	166	234358	11.919	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.177	232	57131	10.117	ng	# 87
60) Diethylphthalate	15.371	149	224999	10.470	ng	97
61) 4-Chlorophenyl-phenyle...	15.595	204	119662	11.511	ng	92
62) 4-Nitroaniline	15.624	138	43201	8.618	ng	# 83
63) Azobenzene	15.889	77	249003	12.110	ng	92
65) 4,6-Dinitro-2-methylph...	15.677	198	29273	7.635	ng	92
66) n-Nitrosodiphenylamine	15.807	169	197280	11.065	ng	99
67) 4-Bromophenyl-phenylether	16.489	248	71864	11.498	ng	93
68) Hexachlorobenzene	16.601	284	78577	11.470	ng	# 83
69) Atrazine	16.754	200	39345	6.573	ng	94
70) Pentachlorophenol	16.942	266	39757	9.274	ng	98
71) Phenanthrene	17.336	178	370944	11.503	ng	99
72) Anthracene	17.430	178	351707	11.050	ng	99
73) Carbazole	17.695	167	346150	10.913	ng	97
74) Di-n-butylphthalate	18.254	149	339669	9.765	ng	98
75) Fluoranthene	19.342	202	414182	11.194	ng	98
77) Benzidine	19.530	184	72904	4.564	ng	97
78) Pyrene	19.706	202	433985	11.114	ng	99
80) Butylbenzylphthalate	20.595	149	140137	8.929	ng	# 98
81) Benzo(a)anthracene	21.436	228	409887	11.022	ng	99
82) 3,3'-Dichlorobenzidine	21.371	252	183282	15.699	ng	98
83) Chrysene	21.489	228	400420	10.960	ng	99
84) Bis(2-ethylhexyl)phtha...	21.359	149	197448	9.431	ng	94
85) Di-n-octyl phthalate	22.265	149	325577	8.720	ng	98
87) Indeno(1,2,3-cd)pyrene	26.171	276	417511	11.970	ng	98
88) Benzo(b)fluoranthene	23.077	252	373678	10.875	ng	98
89) Benzo(k)fluoranthene	23.124	252	383266	11.442	ng	99
90) Benzo(a)pyrene	23.683	252	309794	10.912	ng	99
91) Dibenzo(a,h)anthracene	26.188	278	350288	12.060	ng	98
92) Benzo(g,h,i)perylene	26.900	276	343503	11.890	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
 Data File : BM033242.D
 Acq On : 24 Nov 2021 11:01
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Nov 24 14:39:53 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 14:37:22 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.937 min Scan# 782
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010


Tgt Ion:152 Resp: 115285
 Ion Ratio Lower Upper
 152 100
 150 160.8 122.4 183.6
 115 63.2 53.8 80.8



#2
 1,4-Dioxane
 Concen: 12.064 ng
 RT: 3.425 min Scan# 15
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01

Tgt Ion: 88 Resp: 34041
 Ion Ratio Lower Upper
 88 100
 58 84.6 50.4 75.6#
 43 36.1 22.2 33.2#









#3
Pyridine
Concen: 10.488 ng
RT: 3.837 min Scan# 81
Delta R.T. 0.012 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 81255
Ion Ratio Lower Upper
79 100
52 87.6 48.6 73.0#
51 47.1 27.4 41.2#



#4
n-Nitrosodimethylamine
Concen: 12.782 ng
RT: 3.743 min Scan# 69
Delta R.T. 0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 42 Resp: 42024
Ion Ratio Lower Upper
42 100
74 104.1 99.5 149.3
44 13.4 57.3 85.9#









#5
2-Fluorophenol
Concen: 24.125 ng
RT: 5.507 min Scan# 369
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 159534
Ion Ratio Lower Upper
112 100
64 68.1 53.7 80.5
63 38.7 33.9 50.9



#6
Aniline
Concen: 10.846 ng
RT: 7.266 min Scan# 668
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 93 Resp: 119256
Ion Ratio Lower Upper
93 100
66 44.6 35.3 52.9
65 22.6 18.7 28.1









#7
Phenol-d6
Concen: 21.634 ng
RT: 7.090 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 99 Resp: 213038
Ion Ratio Lower Upper
99 100
42 23.8 23.0 34.6
71 41.8 37.8 56.8



#8
2-Chlorophenol
Concen: 10.986 ng
RT: 7.501 min Scan# 708
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 128 Resp: 84647
Ion Ratio Lower Upper
128 100
130 32.8 13.0 53.0
64 53.8 33.4 73.4









#9
Benzaldehyde
Concen: 13.588 ng
RT: 7.084 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 75522
Ion Ratio Lower Upper
77 100
105 87.2 68.9 108.9
106 80.5 61.2 101.2



#10
Phenol
Concen: 11.355 ng
RT: 7.119 min Scan# 643
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 94 Resp: 107327
Ion Ratio Lower Upper
94 100
65 34.5 16.9 56.9
66 46.4 30.0 70.0









#11
bis(2-Chloroethyl)ether
Concen: 11.301 ng
RT: 7.360 min Scan# 685
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 93 Resp: 84332
Ion Ratio Lower Upper
93 100
63 83.8 63.7 103.7
95 31.4 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 11.657 ng
RT: 7.825 min Scan# 763
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 146 Resp: 99239
Ion Ratio Lower Upper
146 100
148 65.9 52.3 78.5
75 38.1 34.2 51.2









#13
1,4-Dichlorobenzene
Concen: 11.664 ng
RT: 7.972 min Scan# 788
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:146 Resp: 101725
Ion Ratio Lower Upper
146 100
148 63.1 51.1 76.7
111 42.7 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 11.247 ng
RT: 8.290 min Scan# 842
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:146 Resp: 96529
Ion Ratio Lower Upper
146 100
148 66.6 51.7 77.5
111 46.4 38.2 57.2









#15
Benzyl Alcohol
Concen: 11.636 ng
RT: 8.178 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 81565
Ion Ratio Lower Upper
79 100
108 66.7 53.6 80.4
77 67.4 57.7 86.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 14.182 ng
RT: 8.466 min Scan# 872
Delta R.T. 0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 45 Resp: 152126
Ion Ratio Lower Upper
45 100
77 13.1 0.0 36.1
79 9.3 0.0 30.5









#17
2-Methylphenol
Concen: 10.975 ng
RT: 8.372 min Scan# 856
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:	107	Resp:	72871
Ion	Ratio	Lower	Upper
107	100		
108	104.4	87.9	131.9
77	51.7	48.2	72.2
79	49.2	44.7	67.1



#18
Hexachloroethane
Concen: 11.445 ng
RT: 9.019 min Scan# 966
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:	117	Resp:	35296
Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.7	116.5
201	102.8	70.2	105.2









#19
n-Nitroso-di-n-propylamine
Concen: 11.710 ng
RT: 8.742 min Scan# 919
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:	70	Resp:	70601
Ion	Ratio	Lower	Upper
70	100		
42	64.1	56.3	84.5
101	9.0	7.8	11.6
130	18.6	14.2	21.4



#20
3+4-Methylphenols
Concen: 10.787 ng
RT: 8.695 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:	107	Resp:	98025
Ion	Ratio	Lower	Upper
107	100		
108	88.0	67.6	107.6
77	37.3	21.9	61.9
79	34.7	15.3	55.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.736 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:136 Resp: 465912
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 10.1 7.4 11.0
68 6.5 3.8 5.8#



#22
Acetophenone
Concen: 11.649 ng
RT: 8.760 min Scan# 922
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:105 Resp: 132576
Ion Ratio Lower Upper
105 100
71 1.5 3.8 5.6#
51 35.3 29.5 44.3
120 20.3 16.3 24.5





#23

Nitrobenzene-d5

Concen: 25.715 ng

RT: 9.101 min Scan# 91

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion: 82 Resp: 218898

Ion Ratio Lower Upper

82 100

128 35.9 29.0 43.6

54 56.8 45.8 68.8



#24

Nitrobenzene

Concen: 12.748 ng

RT: 9.142 min Scan# 987

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion: 77 Resp: 103975


Ion Ratio Lower Upper

77 100

123 39.8 30.1 45.1

65 13.4 11.7 17.5









#25
Isophorone
Concen: 11.174 ng
RT: 9.666 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 168084
Ion Ratio Lower Upper
82 100
95 8.0 6.4 9.6
138 15.9 12.4 18.6



#26
2-Nitrophenol
Concen: 9.490 ng
RT: 9.848 min Scan# 1107
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 139 Resp: 36136
Ion Ratio Lower Upper
139 100
109 38.2 34.8 52.2
65 71.3 59.9 89.9









#27
2,4-Dimethylphenol
Concen: 10.515 ng
RT: 9.901 min Scan# 1116
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:122 Resp: 65879
Ion Ratio Lower Upper
122 100
107 126.6 100.6 150.8
121 61.1 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 11.133 ng
RT: 10.142 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 93 Resp: 111886
Ion Ratio Lower Upper
93 100
95 31.3 25.8 38.8
123 10.6 8.8 13.2





#29

2,4-Dichlorophenol

Concen: 10.332 ng

RT: 10.378 min Scan# 1197

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 74238

Ion Ratio Lower Upper

162 100

164 67.2 47.4 87.4

98 38.0 18.9 58.9



#30

1,2,4-Trichlorobenzene

Concen: 11.452 ng

RT: 10.595 min Scan# 1234

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:180 Resp: 91451


Ion Ratio Lower Upper

180 100

182 94.0 75.5 113.3

145 30.6 26.2 39.2









#31
Naphthalene
Concen: 11.228 ng
RT: 10.784 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:128 Resp: 271545
Ion Ratio Lower Upper
128 100
129 11.6 8.6 13.0
127 12.7 10.4 15.6



#32
Benzoic acid
Concen: 5.043 ng
RT: 9.972 min Scan# 1128
Delta R.T. -0.053 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:122 Resp: 25552
Ion Ratio Lower Upper
122 100
105 125.5 120.3 160.3
77 111.7 100.0 140.0









#33
4-Chloroaniline
Concen: 9.984 ng
RT: 10.895 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:	127	Resp:	100828
Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.6	38.4
65	37.6	30.4	45.6
92	21.8	16.2	24.4



#34
Hexachlorobutadiene
Concen: 11.571 ng
RT: 11.060 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:	225	Resp:	56169
Ion	Ratio	Lower	Upper
225	100		
223	63.1	52.6	78.8
227	66.8	51.2	76.8









#35
Caprolactam
Concen: 5.341 ng
RT: 11.672 min Scan# 1417
Delta R.T. -0.012 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:113 Resp: 12869
Ion Ratio Lower Upper
113 100
55 184.7 170.5 210.5
56 151.1 125.2 165.2



#36
4-Chloro-3-methylphenol
Concen: 10.156 ng
RT: 12.001 min Scan# 1473
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:107 Resp: 83472
Ion Ratio Lower Upper
107 100
144 22.9 19.3 28.9
142 74.6 60.5 90.7









#37
2-Methylnaphthalene
Concen: 10.793 ng
RT: 12.389 min Scan# 1539
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:142 Resp: 183453
Ion Ratio Lower Upper
142 100
141 89.8 71.8 107.6
115 39.0 36.3 54.5



#38
1-Methylnaphthalene
Concen: 10.913 ng
RT: 12.389 min Scan# 1539
Delta R.T. -0.218 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:142 Resp: 183453
Ion Ratio Lower Upper
142 100
141 89.8 75.6 113.4
116 4.2 4.0 6.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.560 min Scan# 1908
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:164 Resp: 294636
Ion Ratio Lower Upper
164 100
162 101.6 80.6 121.0
160 44.5 35.1 52.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 12.030 ng
RT: 12.754 min Scan# 1601
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:216 Resp: 97991
Ion Ratio Lower Upper
216 100
214 77.0 62.5 93.7
179 21.5 18.7 28.1
108 20.1 21.0 31.4#









#41
Hexachlorocyclopentadiene
Concen: 10.302 ng
RT: 12.730 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:	237	Resp:	44069
Ion Ratio	Lower	Upper	
237	100		
235	67.0	44.4	84.4
272	14.0	0.0	31.4



#42
2,4,6-Tribromophenol
Concen: 20.623 ng
RT: 16.042 min Scan# 2160
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:	330	Resp:	68051
Ion Ratio	Lower	Upper	
330	100		
332	98.3	75.7	113.5
141	42.1	50.7	76.1#





#43

2,4,6-Trichlorophenol

Concen: 10.392 ng

RT: 12.989 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 59237

Ion Ratio Lower Upper

196 100

198 98.8 73.6 110.4

200 32.7 22.1 33.1



#44

2,4,5-Trichlorophenol

Concen: 10.555 ng

RT: 13.060 min Scan# 1653

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Tgt Ion:196 Resp: 66489




Ion Ratio Lower Upper


196 100

198 95.2 77.0 115.4

97 59.9 53.9 80.9

132 28.6 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 24.685 ng

RT: 13.183 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:172 Resp: 468936

Ion Ratio Lower Upper

172 100

171 35.3 28.1 42.1

170 22.9 18.6 27.8



#46

1,1'-Biphenyl

Concen: 11.753 ng

RT: 13.395 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:154 Resp: 250237


Ion Ratio Lower Upper

154 100

153 39.7 20.3 60.3

76 13.9 0.0 33.5









#47
2-Chloronaphthalene
Concen: 11.532 ng
RT: 13.436 min Scan# 1717
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:162 Resp: 188510
Ion Ratio Lower Upper
162 100
127 40.7 33.2 49.8
164 32.0 25.7 38.5



#48
2-Nitroaniline
Concen: 11.458 ng
RT: 13.642 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 65 Resp: 55031
Ion Ratio Lower Upper
65 100
92 62.0 51.2 76.8
138 77.5 66.8 100.2









#49
Acenaphthylene
Concen: 10.948 ng
RT: 14.283 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:152 Resp: 287138
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.1 10.6 15.8



#50
Dimethylphthalate
Concen: 10.843 ng
RT: 14.013 min Scan# 1815
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:163 Resp: 232441
Ion Ratio Lower Upper
163 100
194 3.8 2.9 4.3
164 10.3 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 10.127 ng
RT: 14.136 min Scan# 1836
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 44374
Ion Ratio Lower Upper
165 100
63 67.4 72.4 108.6#
89 66.8 69.3 103.9#



#52
Acenaphthene
Concen: 11.341 ng
RT: 14.624 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:154 Resp: 177173
Ion Ratio Lower Upper
154 100
153 114.5 92.4 138.6
152 53.1 43.7 65.5









#53
3-Nitroaniline
Concen: 9.212 ng
RT: 14.466 min Scan# 1892
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:138 Resp: 43932
Ion Ratio Lower Upper
138 100
108 11.0 8.3 12.5
92 133.2 105.2 157.8



#54
2,4-Dinitrophenol
Concen: 7.092 ng
RT: 14.666 min Scan# 1926
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:184 Resp: 18753
Ion Ratio Lower Upper
184 100
63 85.5 93.8 140.6#
154 67.0 64.0 96.0









#55
Dibenzofuran
Concen: 11.431 ng
RT: 14.954 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:168 Resp: 299789
Ion Ratio Lower Upper
168 100
139 39.0 35.0 52.4
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 8.403 ng
RT: 14.760 min Scan# 1942
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:139 Resp: 32654
Ion Ratio Lower Upper
139 100
109 87.9 92.3 132.3#
65 128.3 110.5 150.5









#57
2,4-Dinitrotoluene
Concen: 9.140 ng
RT: 14.919 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 54254
Ion Ratio Lower Upper
165 100
63 56.8 70.3 105.5#
89 85.4 83.1 124.7
182 3.1 2.2 3.4



#58
Fluorene
Concen: 11.919 ng
RT: 15.601 min Scan# 2085
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:166 Resp: 234358
Ion Ratio Lower Upper
166 100
165 100.3 79.3 118.9
167 13.2 10.6 16.0





#59

2,3,4,6-Tetrachlorophenol

Concen: 10.117 ng

RT: 15.177 min Scan# 2013

Delta R.T. -0.000 min

Lab File: BM033242.D


Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDIC010



Tgt Ion:232 Resp: 57131




Ion Ratio Lower Upper

232 100

131 49.1 50.6 75.8#

130 2.6 3.0 4.4#

166 33.2 28.8 43.2



#60

Diethylphthalate

Concen: 10.470 ng

RT: 15.371 min Scan# 2046

Delta R.T. -0.006 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:149 Resp: 224999


Ion Ratio Lower Upper

149 100

177 21.9 15.7 23.5

150 12.3 9.6 14.4









#61
4-Chlorophenyl-phenylether
Concen: 11.511 ng
RT: 15.595 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 119662
Ion Ratio Lower Upper
204 100
206 31.9 26.7 40.1
141 60.9 55.8 83.8



#62
4-Nitroaniline
Concen: 8.618 ng
RT: 15.624 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:138 Resp: 43201
Ion Ratio Lower Upper
138 100
92 60.6 40.6 80.6
108 95.8 103.8 143.8#









#63
Azobenzene
Concen: 12.110 ng
RT: 15.889 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 249003
Ion Ratio Lower Upper
77 100
182 22.5 0.0 39.3
105 19.0 0.0 36.1
51 33.9 19.1 59.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.295 min Scan# 2373
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion: 188 Resp: 608192
Ion Ratio Lower Upper
188 100
94 9.5 6.3 9.5#
80 11.3 7.3 10.9#









#65
4,6-Dinitro-2-methylphenol
Concen: 7.635 ng
RT: 15.677 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:198 Resp: 29273
Ion Ratio Lower Upper
198 100
51 60.8 49.7 89.7
105 51.4 34.9 74.9



#66
n-Nitrosodiphenylamine
Concen: 11.065 ng
RT: 15.807 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:169 Resp: 197280
Ion Ratio Lower Upper
169 100
168 67.7 55.0 82.4
167 37.5 29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 11.498 ng
RT: 16.489 min Scan# 211
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:248 Resp: 71864
Ion Ratio Lower Upper
248 100
250 96.6 77.4 116.2
141 78.1 74.1 111.1



#68
Hexachlorobenzene
Concen: 11.470 ng
RT: 16.601 min Scan# 2255
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:284 Resp: 78577
Ion Ratio Lower Upper
284 100
142 35.9 43.0 64.6#
249 31.2 27.6 41.4









#69
 Atrazine
 Concen: 6.573 ng
 RT: 16.754 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010


Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.3	5.8	45.8
215	47.5	23.6	63.6



#70
 Pentachlorophenol
 Concen: 9.274 ng
 RT: 16.942 min Scan# 2313
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.8	50.3	75.5
264	65.4	52.1	78.1









#71
Phenanthrene
Concen: 11.503 ng
RT: 17.336 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 370944
Ion Ratio Lower Upper
178 100
176 19.5 15.6 23.4
179 14.9 12.4 18.6



#72
Anthracene
Concen: 11.050 ng
RT: 17.430 min Scan# 2396
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:178 Resp: 351707
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.6 12.3 18.5









#73
 Carbazole
 Concen: 10.913 ng
 RT: 17.695 min Scan# 2442
 Delta R.T. -0.006 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010


Tgt Ion:167 Resp: 346150
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.4 26.0
 139 12.9 11.9 17.9



#74
 Di-n-butylphthalate
 Concen: 9.765 ng
 RT: 18.254 min Scan# 2536
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01

Tgt Ion:149 Resp: 339669
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.3 10.9
 104 6.0 5.6 8.4









#75
Fluoranthene
Concen: 11.194 ng
RT: 19.342 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:202 Resp: 414182
Ion Ratio Lower Upper
202 100
101 11.3 0.0 29.9
203 17.5 0.0 36.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.453 min Scan# 3080
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:240 Resp: 582869
Ion Ratio Lower Upper
240 100
120 10.4 7.8 11.6
236 24.3 20.9 31.3





#77

Benzidine

Concen: 4.564 ng

RT: 19.530 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:184 Resp: 72904

Ion Ratio Lower Upper

184 100

185 14.3 10.6 16.0

183 12.5 9.0 13.4



#78

Pyrene

Concen: 11.114 ng

RT: 19.706 min Scan# 2783

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:202 Resp: 433985


Ion Ratio Lower Upper

202 100

200 20.2 16.6 25.0

203 17.2 13.7 20.5









#79
 Terphenyl-d14
 Concen: 25.158 ng
 RT: 19.901 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010


Tgt Ion:244 Resp: 671655
 Ion Ratio Lower Upper
 244 100
 212 7.0 6.5 9.7
 122 11.0 8.7 13.1



#80
 Butylbenzylphthalate
 Concen: 8.929 ng
 RT: 20.595 min Scan# 2934
 Delta R.T. -0.000 min
 Lab File: BM033242.D
 Acq: 24 Nov 2021 11:01

Tgt Ion:149 Resp: 140137
 Ion Ratio Lower Upper
 149 100
 91 79.3 64.0 96.0
 206 21.8 14.4 21.6#





#81

Benzo(a)anthracene

Concen: 11.022 ng

RT: 21.436 min Scan# 3078

Delta R.T. -0.006 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 409887

Ion Ratio Lower Upper

228 100

226 25.6 21.4 32.2

229 19.3 15.3 22.9



#82

3,3'-Dichlorobenzidine

Concen: 15.699 ng

RT: 21.371 min Scan# 3066

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:252 Resp: 183282


Ion Ratio Lower Upper

252 100

254 63.3 51.6 77.4

126 10.7 9.7 14.5









#83
Chrysene
Concen: 10.960 ng
RT: 21.489 min Scan# 3087
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 400420
Ion Ratio Lower Upper
228 100
226 29.3 24.0 36.0
229 19.9 16.0 24.0



#84
Bis(2-ethylhexyl)phthalate
Concen: 9.431 ng
RT: 21.359 min Scan# 3064
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:149 Resp: 197448
Ion Ratio Lower Upper
149 100
167 27.8 19.7 29.5
279 3.5 2.6 4.0









#85
Di-n-octyl phthalate
Concen: 8.720 ng
RT: 22.265 min Scan# 3218
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:149 Resp: 325577
Ion Ratio Lower Upper
149 100
167 1.7 1.2 1.8
43 12.3 10.4 15.6



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.789 min Scan# 3477
Delta R.T. -0.000 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:264 Resp: 564369
Ion Ratio Lower Upper
264 100
260 22.9 19.3 28.9
265 22.5 18.3 27.5





#87

Indeno(1,2,3-cd)pyrene

Concen: 11.970 ng

RT: 26.171 min Scan# 3

Delta R.T. -0.012 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:276 Resp: 417511

Ion Ratio Lower Upper

276 100

138 24.3 18.4 27.6

277 24.8 20.1 30.1



#88

Benzo(b)fluoranthene

Concen: 10.875 ng

RT: 23.077 min Scan# 3356

Delta R.T. -0.006 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01




Tgt Ion:252 Resp: 373678


Ion Ratio Lower Upper

252 100

253 21.8 18.0 27.0

125 10.3 7.5 11.3









#89
Benzo(k)fluoranthene
Concen: 11.442 ng
RT: 23.124 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:252 Resp: 383266
Ion Ratio Lower Upper
252 100
253 22.0 17.3 25.9
125 10.0 7.5 11.3



#90
Benzo(a)pyrene
Concen: 10.912 ng
RT: 23.683 min Scan# 3459
Delta R.T. -0.006 min
Lab File: BM033242.D
Acq: 24 Nov 2021 11:01

Tgt Ion:252 Resp: 309794
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 11.5 8.1 12.1





#91

Dibenzo(a,h)anthracene

Concen: 12.060 ng

RT: 26.188 min Scan# 3885

Delta R.T. -0.000 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:278 Resp: 350288

Ion Ratio Lower Upper

278 100

139 14.6 10.9 16.3

279 23.3 18.2 27.2



#92

Benzo(g,h,i)perylene

Concen: 11.890 ng

RT: 26.900 min Scan# 4006

Delta R.T. -0.012 min

Lab File: BM033242.D

Acq: 24 Nov 2021 11:01

Tgt Ion:276 Resp: 343503

Ion Ratio Lower Upper

276 100

277 23.0 19.4 29.0

138 21.6 14.6 22.0

