

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
 Data File : BM033257.D
 Acq On : 24 Nov 2021 21:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 05:47:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	126	0.00
2	1,4-Dioxane	40.000	34.430	13.9	115	0.00
3	Pyridine	40.000	38.389	4.0	121	0.00
4	n-Nitrosodimethylamine	40.000	36.150	9.6	117	0.00
5 S	2-Fluorophenol	80.000	78.205	2.2	129	0.00
6	Aniline	40.000	39.179	2.1	126	0.00
7 S	Phenol-d6	80.000	80.837	-1.0	132	0.00
8	2-Chlorophenol	40.000	40.026	-0.1	131	0.00
9	Benzaldehyde	40.000	42.670	-6.7	134	0.00
10 C	Phenol	40.000	39.879	0.3	131	0.01
11	bis(2-Chloroethyl)ether	40.000	39.058	2.4	126	0.00
12	1,3-Dichlorobenzene	40.000	37.641	5.9	125	0.00
13 C	1,4-Dichlorobenzene	40.000	37.714	5.7	126	0.00
14	1,2-Dichlorobenzene	40.000	37.941	5.1	125	0.00
15	Benzyl Alcohol	40.000	39.945	0.1	127	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.850	7.9	121	0.00
17	2-Methylphenol	40.000	40.281	-0.7	130	0.00
18	Hexachloroethane	40.000	39.000	2.5	127	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.299	1.8	127	0.00
20	3+4-Methylphenols	40.000	41.213	-3.0	132	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	123	0.00
22	Acetophenone	40.000	38.396	4.0	128	0.00
23 S	Nitrobenzene-d5	80.000	77.188	3.5	125	0.00
24	Nitrobenzene	40.000	37.910	5.2	125	0.00
25	Isophorone	40.000	39.794	0.5	128	0.00
26 C	2-Nitrophenol	40.000	44.509	-11.3	140	0.00
27	2,4-Dimethylphenol	40.000	39.397	1.5	128	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.810	3.0	126	0.00
29 C	2,4-Dichlorophenol	40.000	40.860	-2.1	130	0.00
30	1,2,4-Trichlorobenzene	40.000	37.642	5.9	124	0.00
31	Naphthalene	40.000	37.762	5.6	124	0.00
32	Benzoic acid	40.000	33.668	15.8	106	0.00
33	4-Chloroaniline	40.000	38.781	3.0	124	0.00
34 C	Hexachlorobutadiene	40.000	37.124	7.2	123	0.00
35	Caprolactam	40.000	42.731	-6.8	133	0.01
36 C	4-Chloro-3-methylphenol	40.000	39.124	2.2	124	0.00
37	2-Methylnaphthalene	40.000	37.869	5.3	124	0.00
38	1-Methylnaphthalene	40.000	37.863	5.3	123	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.099	4.8	125	0.00
41 P	Hexachlorocyclopentadiene	40.000	30.454	23.9	93	0.00
42 S	2,4,6-Tribromophenol	80.000	79.905	0.1	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.714	-4.3	130	0.00
44	2,4,5-Trichlorophenol	40.000	40.145	-0.4	127	0.00
45 S	2-Fluorobiphenyl	80.000	75.802	5.2	123	0.00
46	1,1'-Biphenyl	40.000	38.107	4.7	124	0.00
47	2-Chloronaphthalene	40.000	38.245	4.4	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
 Data File : BM033257.D
 Acq On : 24 Nov 2021 21:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 05:47:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.978	-2.4	125	0.00
49	Acenaphthylene	40.000	38.641	3.4	121	0.00
50	Dimethylphthalate	40.000	38.051	4.9	123	0.00
51	2,6-Dinitrotoluene	40.000	39.885	0.3	123	0.00
52 C	Acenaphthene	40.000	37.633	5.9	122	0.00
53	3-Nitroaniline	40.000	40.669	-1.7	122	0.00
54 P	2,4-Dinitrophenol	40.000	8.216	79.5#	25	0.00
55	Dibenzofuran	40.000	37.316	6.7	120	0.00
56 P	4-Nitrophenol	40.000	40.978	-2.4	120	0.02
57	2,4-Dinitrotoluene	40.000	40.260	-0.6	122	0.00
58	Fluorene	40.000	36.884	7.8	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.342	1.6	124	0.00
60	Diethylphthalate	40.000	38.549	3.6	124	0.00
61	4-Chlorophenyl-phenylether	40.000	36.948	7.6	119	0.00
62	4-Nitroaniline	40.000	41.730	-4.3	125	0.00
63	Azobenzene	40.000	38.771	3.1	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	40.000	17.387	56.5#	49	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.723	0.7	120	0.00
67	4-Bromophenyl-phenylether	40.000	40.131	-0.3	122	0.00
68	Hexachlorobenzene	40.000	39.082	2.3	120	0.00
69	Atrazine	40.000	43.018	-7.5	124	0.00
70 C	Pentachlorophenol	40.000	38.481	3.8	110	0.00
71	Phenanthrene	40.000	37.762	5.6	116	0.00
72	Anthracene	40.000	38.913	2.7	118	0.00
73	Carbazole	40.000	38.223	4.4	116	0.00
74	Di-n-butylphthalate	40.000	43.127	-7.8	129	0.00
75 C	Fluoranthene	40.000	36.936	7.7	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	55.469	-38.7#	137	0.00
78	Pyrene	40.000	39.829	0.4	112	0.00
79 S	Terphenyl-d14	80.000	78.949	1.3	112	0.00
80	Butylbenzylphthalate	40.000	47.638	-19.1	128	0.00
81	Benzo(a)anthracene	40.000	38.163	4.6	110	0.00
82	3,3'-Dichlorobenzidine	40.000	40.424	-1.1	112	0.00
83	Chrysene	40.000	37.457	6.4	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	48.053	-20.1	131	0.00
85 c	Di-n-octyl phthalate	40.000	46.801	-17.0	129	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.628	5.9	110	0.00
88	Benzo(b)fluoranthene	40.000	38.269	4.3	109	0.00
89	Benzo(k)fluoranthene	40.000	36.085	9.8	106	0.00
90 C	Benzo(a)pyrene	40.000	37.638	5.9	109	0.00
91	Dibenzo(a,h)anthracene	40.000	37.716	5.7	110	0.00
92	Benzo(g,h,i)perylene	40.000	37.464	6.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112421\
Data File : BM033257.D
Acq On : 24 Nov 2021 21:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 05:47:49 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 24 15:16:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0