

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112522\
 Data File : BM037750.D
 Acq On : 27 Nov 2022 01:26
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020056

Quant Time: Nov 27 21:31:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 01:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	242114	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	1117690	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.733	164	734094	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.468	188	1592099	20.000	ng/u1	0.00
79) Chrysene-d12	21.633	240	1585316	20.000	ng/u1	0.00
88) Perylene-d12	24.127	264	1456268	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	44526	7.627	ng/uL	0.00
4) Pyridine-d5	3.869	84	368176	19.375	ng/u1	0.00
7) Phenol-d5	7.257	99	453136	18.760	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.428	67	295567	18.343	ng/u1	0.00
11) 2-Chlorophenol-d4	7.634	132	343247	20.451	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	363359	18.904	ng/u1	0.00
21) Nitrobenzene-d5	9.281	128	189733	21.654	ng/u1	0.00
24) 2-Nitrophenol-d4	10.004	143	196693	22.782	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.545	165	355453	22.497	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	542092	20.676	ng/u1	-0.01
46) Dimethylphthalate-d6	14.139	166	1109435	21.336	ng/u1	0.00
49) Acenaphthylene-d8	14.427	160	1310646	21.355	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.916	143	230481	20.123	ng/u1	0.00
60) Fluorene-d10	15.722	176	985629	21.643	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.833	200	169361	19.775	ng/u1	0.00
73) Anthracene-d10	17.568	188	1591713	21.379	ng/u1	0.00
81) Pyrene-d10	19.851	212	1793832	22.267	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.962	264	1598824	21.330	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.469	88	50102	7.806	ng/uL#	89
5) Pyridine	3.893	79	370197	19.158	ng/u1	99
6) Benzaldehyde	7.240	77	238955	18.848	ng/u1	99
8) Phenol	7.287	94	477281	18.851	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.522	93	385916	18.801	ng/u1	97
12) 2-Chlorophenol	7.663	128	373901	20.393	ng/u1	99
13) 2-Methylphenol	8.551	108	364859	18.890	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.640	45	663968	17.908	ng/u1	97
16) Acetophenone	8.940	105	611424	19.323	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.922	70	338668	18.262	ng/u1	99
18) 4-Methylphenol	8.881	108	403178	18.893	ng/u1	98
19) Hexachloroethane	9.198	117	165290	20.843	ng/u1	90
22) Nitrobenzene	9.322	77	501492	20.990	ng/u1	100
23) Isophorone	9.851	82	940565	19.533	ng/u1	98
25) 2-Nitrophenol	10.040	139	216518	22.173	ng/u1	96
26) 2,4-Dimethylphenol	10.092	107	458758	21.392	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.328	93	533411	19.311	ng/u1	99
29) 2,4-Dichlorophenol	10.569	162	368169	22.777	ng/u1	98
30) Naphthalene	10.981	128	1304882	21.352	ng/u1	99
32) 4-Chloroaniline	11.086	127	561055	20.535	ng/u1	98
33) Hexachlorobutadiene	11.263	225	221269	26.077	ng/u1	97
34) Caprolactam	11.869	113	128236	19.049	ng/u1	94
35) 4-Chloro-3-methylphenol	12.198	107	431923	20.904	ng/u1	99
36) 2-Methylnaphthalene	12.575	142	896059	21.107	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112522\
 Data File : BM037750.D
 Acq On : 27 Nov 2022 01:26
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020056

Quant Time: Nov 27 21:31:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 01:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.792	142	897016	21.003	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.939	216	414201	22.979	ng/ul	97
40) Hexachlorocyclopentadiene	12.922	237	243455	21.689	ng/ul	99
41) 2,4,6-Trichlorophenol	13.175	196	286291	22.057	ng/ul	100
42) 2,4,5-Trichlorophenol	13.245	196	312538	22.505	ng/ul	99
43) 1,1'-Biphenyl	13.575	154	1174939	21.077	ng/ul	98
44) 2-Chloronaphthalene	13.616	162	921033	21.257	ng/ul	99
45) 2-Nitroaniline	13.816	65	331252	20.165	ng/ul	95
47) Dimethylphthalate	14.186	163	1167169	21.117	ng/ul	100
48) 2,6-Dinitrotoluene	14.310	165	233443	21.873	ng/ul	99
50) Acenaphthylene	14.457	152	1472884	21.231	ng/ul	99
51) 3-Nitroaniline	14.633	138	259103	20.007	ng/ul	93
52) Acenaphthene	14.798	153	1038717	20.935	ng/ul	100
53) 2,4-Dinitrophenol	14.833	184	118796	18.971	ng/ul	98
55) 4-Nitrophenol	14.933	109	220269	21.461	ng/ul	89
56) Dibenzofuran	15.127	168	1391582	21.489	ng/ul	99
57) 2,4-Dinitrotoluene	15.086	165	345671	22.072	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.351	232	265276	23.281	ng/ul	94
59) Diethylphthalate	15.539	149	1233365	20.675	ng/ul	98
61) Fluorene	15.774	166	1198576	21.367	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.763	204	549284	22.446	ng/ul	99
63) 4-Nitroaniline	15.792	138	260681	21.179	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.845	198	179633	19.762	ng/ul#	98
67) N-Nitrosodiphenylamine	15.974	169	1010327	20.424	ng/ul	98
68) 4-Bromophenyl-phenylether	16.657	248	306842	23.961	ng/ul	99
69) Hexachlorobenzene	16.774	284	349848	21.272	ng/ul	94
70) Atrazine	16.921	200	343652	20.466	ng/ul	98
71) Pentachlorophenol	17.116	266	225576	20.714	ng/ul	97
72) Phenanthrene	17.515	178	1945109	20.856	ng/ul	99
74) Anthracene	17.604	178	1972253	20.830	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.539	216	422115	22.163	ng/ul	99
76) Pentachlorobenzene	15.051	250	442964	21.867	ng/ul	99
77) Carbazole	17.874	167	1822488	20.246	ng/ul	100
78) Di-n-butylphthalate	18.421	149	2271935	17.436	ng/ul	99
80) Fluoranthene	19.515	202	2245108	21.816	ng/ul	96
82) Pyrene	19.874	202	2371822	21.820	ng/ul	97
83) Butylbenzylphthalate	20.756	149	1085755	21.442	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.545	252	650533	20.432	ng/ul	99
85) Benzo(a)anthracene	21.621	228	2244948	21.167	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.527	149	1602746	20.950	ng/ul#	98
87) Chrysene	21.674	228	2101127	21.053	ng/ul	99
89) Di-n-octyl phthalate	22.480	149	2750704	22.112	ng/ul	100
90) Benzo(b)fluoranthene	23.362	252	2119845	21.413	ng/ul	97
91) Benzo(k)fluoranthene	23.409	252	2051906	21.253	ng/ul	98
93) Benzo(a)pyrene	24.015	252	1802788	20.952	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.733	276	1834602	18.585	ng/ul	95
95) Dibenzo(a,h)anthracene	26.750	278	1635397	18.452	ng/ul	97
96) Benzo(g,h,i)perylene	27.538	276	1578266	18.361	ng/ul	94





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.806 ng/uL
RT: 3.469 min Scan# 14
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion: 88 Resp: 50102
Ion Ratio Lower Upper
88 100
43 42.3 42.8 64.2#
58 85.1 62.6 94.0



#5
Pyridine
Concen: 19.158 ng/uL
RT: 3.893 min Scan# 86
Delta R.T. 0.000 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion: 79 Resp: 370197
Ion Ratio Lower Upper
79 100
52 85.4 68.2 102.4
51 42.3 32.4 48.6









#6
Benzaldehyde
Concen: 18.848 ng/ul
RT: 7.240 min Scan# 611
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion: 77 Resp: 238955
Ion Ratio Lower Upper
77 100
105 90.7 73.4 110.0
106 88.8 70.2 105.2



#8
Phenol
Concen: 18.851 ng/ul
RT: 7.287 min Scan# 663
Delta R.T. 0.000 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion: 94 Resp: 477281
Ion Ratio Lower Upper
94 100
65 32.7 25.0 37.4
66 42.7 31.0 46.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.801 ng/ul
RT: 7.522 min Scan# 704
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion: 93 Resp: 385916
Ion Ratio Lower Upper
93 100
63 82.9 68.6 103.0
95 32.9 27.1 40.7



#12
2-Chlorophenol
Concen: 20.393 ng/ul
RT: 7.663 min Scan# 727
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:128 Resp: 373901
Ion Ratio Lower Upper
128 100
64 51.9 42.2 63.4
130 31.6 26.1 39.1





#13

2-Methylphenol

Concen: 18.890 ng/ul

RT: 8.551 min Scan# 81

Delta R.T. -0.006 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :





SSTD020056

Tgt Ion:108 Resp: 364859

Ion Ratio Lower Upper

108 100

107 89.3 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.908 ng/ul

RT: 8.640 min Scan# 893

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26




Tgt Ion: 45 Resp: 663968


Ion Ratio Lower Upper

45 100

77 12.8 9.4 14.0

79 10.3 7.3 10.9









#16
Acetophenone
Concen: 19.323 ng/ul
RT: 8.940 min Scan# 945
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:105 Resp: 611424
Ion Ratio Lower Upper
105 100
77 85.8 66.6 99.8
51 34.0 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 18.262 ng/ul
RT: 8.922 min Scan# 941
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion: 70 Resp: 338668
Ion Ratio Lower Upper
70 100
42 70.6 57.1 85.7
101 9.6 7.0 10.6
130 19.3 14.8 22.2









#18
4-Methylphenol
Concen: 18.893 ng/ul
RT: 8.881 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:108 Resp: 403178
Ion Ratio Lower Upper
108 100
107 110.8 90.4 135.6



#19
Hexachloroethane
Concen: 20.843 ng/ul
RT: 9.198 min Scan# 988
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:117 Resp: 165290
Ion Ratio Lower Upper
117 100
201 83.4 60.4 90.6
199 54.9 38.1 57.1









#22
Nitrobenzene
Concen: 20.990 ng/ul
RT: 9.322 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion: 77 Resp: 501492
Ion Ratio Lower Upper
77 100
123 40.6 32.4 48.6
65 13.9 11.4 17.0



#23
Isophorone
Concen: 19.533 ng/ul
RT: 9.851 min Scan# 1099
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion: 82 Resp: 940565
Ion Ratio Lower Upper
82 100
95 7.2 5.3 7.9
138 16.6 12.5 18.7









#25
2-Nitrophenol
Concen: 22.173 ng/ul
RT: 10.040 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:139 Resp: 216518
Ion Ratio Lower Upper
139 100
65 58.8 49.6 74.4
109 34.8 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 21.392 ng/ul
RT: 10.092 min Scan# 1140
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:107 Resp: 458758
Ion Ratio Lower Upper
107 100
121 48.2 39.6 59.4
122 85.4 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.311 ng/ul
 RT: 10.328 min Scan# 1180
 Delta R.T. -0.012 min
 Lab File: BM037750.D
 Acq: 27 Nov 2022 01:26




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020056


Tgt Ion: 93 Resp: 533411
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.4 39.6
 123 10.5 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 22.777 ng/ul
 RT: 10.569 min Scan# 1221
 Delta R.T. -0.012 min
 Lab File: BM037750.D
 Acq: 27 Nov 2022 01:26

Tgt Ion:162 Resp: 368169
 Ion Ratio Lower Upper
 162 100
 164 66.1 52.6 79.0
 98 39.2 33.6 50.4









#30
Naphthalene
Concen: 21.352 ng/ul
RT: 10.981 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:128 Resp: 1304882
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 20.535 ng/ul
RT: 11.086 min Scan# 1309
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:127 Resp: 561055
Ion Ratio Lower Upper
127 100
129 31.6 26.1 39.1









#33
Hexachlorobutadiene
Concen: 26.077 ng/ul
RT: 11.263 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:225 Resp: 221269
Ion Ratio Lower Upper
225 100
223 61.6 51.1 76.7
227 60.6 50.7 76.1



#34
Caprolactam
Concen: 19.049 ng/ul
RT: 11.869 min Scan# 1442
Delta R.T. 0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:113 Resp: 128236
Ion Ratio Lower Upper
113 100
55 194.9 163.3 244.9
56 142.1 120.3 180.5









#35
4-Chloro-3-methylphenol
Concen: 20.904 ng/ul
RT: 12.198 min Scan# 1498
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:107 Resp: 431923
Ion Ratio Lower Upper
107 100
144 25.3 20.2 30.4
142 79.0 62.6 93.8



#36
2-Methylnaphthalene
Concen: 21.107 ng/ul
RT: 12.575 min Scan# 1562
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:142 Resp: 896059
Ion Ratio Lower Upper
142 100
141 89.1 72.2 108.4





#37

1-Methylnaphthalene

Concen: 21.003 ng/ul

RT: 12.792 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:142 Resp: 897016

Ion Ratio Lower Upper

142 100

141 92.8 73.3 109.9

116 4.3 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.979 ng/ul

RT: 12.939 min Scan# 1624

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Tgt Ion:216 Resp: 414201




Ion Ratio Lower Upper


216 100

214 78.9 60.2 90.4

179 22.9 18.6 27.8

108 22.1 18.3 27.5









#40
Hexachlorocyclopentadiene
Concen: 21.689 ng/ul
RT: 12.922 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:237 Resp: 243455
Ion Ratio Lower Upper
237 100
235 62.2 49.1 73.7
272 11.9 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 22.057 ng/ul
RT: 13.175 min Scan# 1664
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:196 Resp: 286291
Ion Ratio Lower Upper
196 100
198 96.0 76.9 115.3
200 30.6 24.1 36.1









#42
2,4,5-Trichlorophenol
Concen: 22.505 ng/ul
RT: 13.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:196 Resp: 312538
Ion Ratio Lower Upper
196 100
198 95.0 75.4 113.2
200 30.8 23.7 35.5



#43
1,1'-Biphenyl
Concen: 21.077 ng/ul
RT: 13.575 min Scan# 1732
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:154 Resp: 1174939
Ion Ratio Lower Upper
154 100
153 40.4 32.1 48.1
76 13.1 12.4 18.6





#44

2-Chloronaphthalene

Concen: 21.257 ng/ul

RT: 13.616 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:162 Resp: 921033

Ion Ratio Lower Upper

162 100

164 33.2 26.5 39.7

127 41.6 32.4 48.6



#45

2-Nitroaniline

Concen: 20.165 ng/ul

RT: 13.816 min Scan# 1773

Delta R.T. -0.006 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26




Tgt Ion: 65 Resp: 331252


Ion Ratio Lower Upper

65 100

92 62.8 48.6 73.0

138 89.3 67.0 100.4









#47
Dimethylphthalate
Concen: 21.117 ng/ul
RT: 14.186 min Scan# 1836
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:163 Resp: 1167169
Ion Ratio Lower Upper
163 100
194 3.8 2.9 4.3
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 21.873 ng/ul
RT: 14.310 min Scan# 1857
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:165 Resp: 233443
Ion Ratio Lower Upper
165 100
89 69.2 56.4 84.6
121 23.2 18.9 28.3









#50
Acenaphthylene
Concen: 21.231 ng/ul
RT: 14.457 min Scan# 1882
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:152 Resp: 1472884
Ion Ratio Lower Upper
152 100
151 19.9 16.3 24.5
153 13.1 10.3 15.5



#51
3-Nitroaniline
Concen: 20.007 ng/ul
RT: 14.633 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:138 Resp: 259103
Ion Ratio Lower Upper
138 100
108 10.5 8.4 12.6
92 121.3 104.0 156.0









#52
Acenaphthene
Concen: 20.935 ng/ul
RT: 14.798 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:153 Resp: 1038717
Ion Ratio Lower Upper
153 100
152 48.1 38.4 57.6
154 88.7 70.7 106.1



#53
2,4-Dinitrophenol
Concen: 18.971 ng/ul
RT: 14.833 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:184 Resp: 118796
Ion Ratio Lower Upper
184 100
63 92.3 74.9 112.3
154 70.1 57.8 86.6









#55
4-Nitrophenol
Concen: 21.461 ng/ul
RT: 14.933 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:109 Resp: 220269
Ion Ratio Lower Upper
109 100
139 116.0 102.4 153.6
65 134.3 119.6 179.4



#56
Dibenzofuran
Concen: 21.489 ng/ul
RT: 15.127 min Scan# 1996
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:168 Resp: 1391582
Ion Ratio Lower Upper
168 100
139 40.6 32.2 48.2









#57
2,4-Dinitrotoluene
Concen: 22.072 ng/ul
RT: 15.086 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:165 Resp: 345671
Ion Ratio Lower Upper
165 100
63 58.0 46.6 70.0
182 2.6 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.281 ng/ul
RT: 15.351 min Scan# 2034
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:232 Resp: 265276
Ion Ratio Lower Upper
232 100
131 57.4 50.2 75.4
130 3.3 2.8 4.2
166 36.1 31.2 46.8









#59
Diethylphthalate
Concen: 20.675 ng/ul
RT: 15.539 min Scan# 2066
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:149 Resp: 1233365
Ion Ratio Lower Upper
149 100
177 20.8 16.1 24.1
150 12.5 9.4 14.0



#61
Fluorene
Concen: 21.367 ng/ul
RT: 15.774 min Scan# 2106
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:166 Resp: 1198576
Ion Ratio Lower Upper
166 100
165 98.4 79.3 118.9
167 13.5 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 22.446 ng/ul
RT: 15.763 min Scan# 2104
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:204 Resp: 549284
Ion Ratio Lower Upper
204 100
206 31.6 25.8 38.8
141 68.7 54.3 81.5



#63
4-Nitroaniline
Concen: 21.179 ng/ul
RT: 15.792 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:138 Resp: 260681
Ion Ratio Lower Upper
138 100
92 55.6 43.7 65.5
108 92.6 68.1 102.1









#66
4,6-Dinitro-2-methylphenol
Concen: 19.762 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:198 Resp: 179633
Ion Ratio Lower Upper
198 100
182 4.0 4.0 6.0#
77 33.6 27.9 41.9



#67
N-Nitrosodiphenylamine
Concen: 20.424 ng/ul
RT: 15.974 min Scan# 2140
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:169 Resp: 1010327
Ion Ratio Lower Upper
169 100
168 68.9 54.0 81.0
167 37.7 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 23.961 ng/ul
RT: 16.657 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:248 Resp: 306842
Ion Ratio Lower Upper
248 100
250 97.7 78.5 117.7
141 97.3 78.9 118.3



#69
Hexachlorobenzene
Concen: 21.272 ng/ul
RT: 16.774 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:284 Resp: 349848
Ion Ratio Lower Upper
284 100
142 45.7 40.4 60.6
249 32.7 24.5 36.7









#70
Atrazine
Concen: 20.466 ng/ul
RT: 16.921 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:200 Resp: 343652
Ion Ratio Lower Upper
200 100
173 26.9 22.6 34.0
215 45.3 36.6 55.0



#71
Pentachlorophenol
Concen: 20.714 ng/ul
RT: 17.116 min Scan# 2334
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:266 Resp: 225576
Ion Ratio Lower Upper
266 100
264 62.0 52.0 78.0
268 63.2 52.2 78.2









#72
Phenanthrene
Concen: 20.856 ng/ul
RT: 17.515 min Scan# 2402
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:178 Resp: 1945109
Ion Ratio Lower Upper
178 100
179 15.2 12.0 18.0
176 19.6 15.4 23.2



#74
Anthracene
Concen: 20.830 ng/ul
RT: 17.604 min Scan# 2417
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:178 Resp: 1972253
Ion Ratio Lower Upper
178 100
179 15.3 12.6 18.8
176 19.0 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.163 ng/uL
RT: 13.539 min Scan# 1726
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:216 Resp: 422115
Ion Ratio Lower Upper
216 100
214 78.9 63.3 94.9
218 47.2 38.6 58.0



#76
Pentachlorobenzene
Concen: 21.867 ng/uL
RT: 15.051 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:250 Resp: 442964
Ion Ratio Lower Upper
250 100
248 63.0 50.1 75.1
252 63.4 49.9 74.9





#77

Carbazole

Concen: 20.246 ng/ul

RT: 17.874 min Scan# 2463

Delta R.T. 0.000 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:167 Resp: 1822488

Ion Ratio Lower Upper

167 100

166 21.7 17.5 26.3

139 13.6 10.9 16.3



#78

Di-n-butylphthalate

Concen: 17.436 ng/ul

RT: 18.421 min Scan# 2556

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26




Tgt Ion:149 Resp: 2271935


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.0

104 6.0 4.4 6.6





#80

Fluoranthene

Concen: 21.816 ng/ul

RT: 19.515 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:202 Resp: 2245108

Ion Ratio Lower Upper

202 100

101 11.5 10.6 16.0

100 8.8 8.2 12.4



#82

Pyrene

Concen: 21.820 ng/ul

RT: 19.874 min Scan# 2803

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26




Tgt Ion:202 Resp: 2371822


Ion Ratio Lower Upper

202 100

101 14.1 12.2 18.4

100 11.5 10.2 15.2









#83
Butylbenzylphthalate
Concen: 21.442 ng/ul
RT: 20.756 min Scan# 2953
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:149 Resp: 1085755
Ion Ratio Lower Upper
149 100
91 80.4 65.9 98.9
206 18.0 13.4 20.2



#84
3,3'-Dichlorobenzidine
Concen: 20.432 ng/ul
RT: 21.545 min Scan# 3087
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:252 Resp: 650533
Ion Ratio Lower Upper
252 100
254 63.5 51.0 76.4
126 13.3 12.2 18.4









#85
Benzo(a)anthracene
Concen: 21.167 ng/ul
RT: 21.621 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:228 Resp: 2244948
Ion Ratio Lower Upper
228 100
229 19.6 15.6 23.4
226 27.7 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.950 ng/ul
RT: 21.527 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:149 Resp: 1602746
Ion Ratio Lower Upper
149 100
167 26.4 20.2 30.4
279 3.4 2.2 3.4#





#87

Chrysene

Concen: 21.053 ng/ul

RT: 21.674 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:228 Resp: 2101127

Ion Ratio Lower Upper

228 100

226 30.5 24.2 36.4

229 19.5 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 22.112 ng/ul




RT: 22.480 min Scan# 3246


Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Tgt Ion:149 Resp: 2750704









#90
Benzo(b)fluoranthene
Concen: 21.413 ng/ul
RT: 23.362 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26




Instrument :
BNA_M
ClientSampleId :
SSTD020056


Tgt Ion:252 Resp: 2119845
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 11.1 11.0 16.4



#91
Benzo(k)fluoranthene
Concen: 21.253 ng/ul
RT: 23.409 min Scan# 3404
Delta R.T. -0.012 min
Lab File: BM037750.D
Acq: 27 Nov 2022 01:26

Tgt Ion:252 Resp: 2051906
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 11.0 10.6 16.0





#93

Benzo(a)pyrene

Concen: 20.952 ng/ul

RT: 24.015 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:252 Resp: 1802788

Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 12.5 12.0 18.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.585 ng/ul

RT: 26.733 min Scan# 3969

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26




Tgt Ion:276 Resp: 1834602


Ion Ratio Lower Upper

276 100

138 28.8 26.9 40.3

277 25.6 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 18.452 ng/ul

RT: 26.750 min Scan# 3972

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Instrument :

BNA_M

ClientSampleId :

SSTD020056





Tgt Ion:278 Resp: 1635397

Ion Ratio Lower Upper

278 100

139 18.7 16.6 25.0

279 23.4 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 18.361 ng/ul

RT: 27.538 min Scan# 4106

Delta R.T. -0.012 min

Lab File: BM037750.D

Acq: 27 Nov 2022 01:26

Tgt Ion:276 Resp: 1578266

Ion Ratio Lower Upper

276 100

138 24.4 23.4 35.2

277 23.3 19.1 28.7

