

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112522\
 Data File : BM037696.D
 Acq On : 25 Nov 2022 15:28
 Operator : CG/JU
 Sample : N5591-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C9804

Quant Time: Nov 26 02:44:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 01:27:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	100007	20.000	ng/u1	0.00
20) Naphthalene-d8	10.939	136	489130	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	311768	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.474	188	642533	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	638382	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	653627	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	15331	6.358	ng/uL	0.00
4) Pyridine-d5	3.887	84	50786	6.470	ng/u1	0.02
7) Phenol-d5	7.257	99	100658	10.089	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	270520	40.646	ng/u1	0.00
11) 2-Chlorophenol-d4	7.639	132	220982	31.875	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	168648	21.242	ng/u1	0.00
21) Nitrobenzene-d5	9.286	128	146073	38.094	ng/u1	0.00
24) 2-Nitrophenol-d4	10.010	143	129012	34.145	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	221429	32.024	ng/u1	0.00
31) 4-Chloroaniline-d4	11.069	131	306236	26.690	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	890346	40.318	ng/u1	0.00
49) Acenaphthylene-d8	14.433	160	997092	38.253	ng/u1	0.00
54) 4-Nitrophenol-d4	14.915	143	39720	8.166	ng/u1	0.00
60) Fluorene-d10	15.727	176	765735	39.592	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.833	200	108976	31.528	ng/u1	0.00
73) Anthracene-d10	17.574	188	1333570	44.383	ng/u1	0.00
81) Pyrene-d10	19.856	212	1520440	46.868	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1477502	43.916	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112522\
Data File : BM037696.D
Acq On : 25 Nov 2022 15:28
Operator : CG/JU
Sample : N5591-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C9804

Quant Time: Nov 26 02:44:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 24 01:27:15 2022
Response via : Initial Calibration

