

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
 Data File : BM003309.D
 Acq On : 27 Nov 2015 17:21
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICBM112715

Quant Time: Nov 27 18:23:20 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.488	0.435	10.9	110	-0.02
3	n-Nitrosodimethylamine	0.462	0.455	1.5	117	0.00
4 S	2-Fluorophenol	0.986	0.964	2.2	120	0.00
5 S	Phenol-d6	1.131	1.141	-0.9	124	0.00
6	bis(2-Chloroethyl)ether	1.156	1.145	1.0	121	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.222	0.219	1.4	128	0.00
9	Nitrobenzene	0.243	0.240	1.2	130	0.00
10	Naphthalene	1.081	1.076	0.5	125	0.01
11	Hexachlorobutadiene	0.164	0.166	-1.2	129	0.00
12	2-Methylnaphthalene	0.678	0.703	-3.7	130	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.090	0.081	10.0	122	0.00
15 S	2-Fluorobiphenyl	1.614	1.494	7.4	133	0.00
16	Acenaphthylene	1.832	1.797	1.9	136	0.00
17	Acenaphthene	1.414	1.200	15.1	131	0.00
18	Fluorene	1.427	1.349	5.5	131	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
20	4-Bromophenyl-phenylether	0.170	0.134	21.2	127	-0.03
21	Hexachlorobenzene	0.176	0.175	0.6	130	0.00
22	Pentachlorophenol	0.066	0.041	37.9#	110	0.00
23	Phenanthrene	1.145	0.913	20.3	119	0.00
24	Anthracene	0.933	0.935	-0.2	120	0.00
25	Fluoranthene	1.090	0.871	20.1	107	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
27	Benzydine	0.229	0.240	-4.8	123	0.00
28	Pvrene	1.528	1.642	-7.5	107	0.00
29 S	Terphenyl-d14	0.697	0.777	-11.5	102	0.00
30	Benzo(a)anthracene	1.266	1.311	-3.6	97	0.00
31	3,3'-Dichlorobenzidine	0.270	0.219	18.9	100	0.02
32	Chrysene	1.515	1.434	5.3	95	0.00
33	Bis(2-ethylhexyl)phthalate	0.518	0.450	13.1	120	0.00
34	Indeno(1,2,3-cd)pyrene	1.160	1.213	-4.6	101	0.01
35 I	Perylene-d12	1.000	1.000	0.0	92	0.01
36	Benzo(b)fluoranthene	1.439	1.399	2.8	88	0.01
37	Benzo(k)fluoranthene	1.294	1.351	-4.4	101	0.00
38 C	Benzo(a)pyrene	1.146	1.141	0.4	96	0.01
39	Dibenzo(a,h)anthracene	0.992	1.117	-12.6	105	0.01
40	Benzo(a,h,i)perylene	1.190	1.215	-2.1	97	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
Data File : BM003309.D
Acq On : 27 Nov 2015 17:21
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM112715

Quant Time: Nov 27 18:23:20 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				