

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
 Data File : BM003311.D
 Acq On : 27 Nov 2015 19:09
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Nov 27 22:49:17 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.488	0.494	-1.2	108	-0.02
3	n-Nitrosodimethylamine	0.462	0.462	0.0	103	0.00
4 S	2-Fluorophenol	0.986	0.971	1.5	104	0.00
5 S	Phenol-d6	1.131	1.095	3.2	103	0.00
6	bis(2-Chloroethyl)ether	1.156	1.123	2.9	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.222	0.214	3.6	102	0.00
9	Nitrobenzene	0.243	0.231	4.9	102	0.00
10	Naphthalene	1.081	1.078	0.3	102	0.00
11	Hexachlorobutadiene	0.164	0.174	-6.1	110	0.00
12	2-Methylnaphthalene	0.678	0.672	0.9	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.090	0.088	2.2	102	0.00
15 S	2-Fluorobiphenyl	1.614	1.505	6.8	102	0.00
16	Acenaphthylene	1.832	1.763	3.8	101	0.00
17	Acenaphthene	1.414	1.223	13.5	101	0.00
18	Fluorene	1.427	1.356	5.0	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.170	0.123	27.6#	101	0.00
21	Hexachlorobenzene	0.176	0.157	10.8	100	0.00
22	Pentachlorophenol	0.066	0.057	13.6	130	0.00
23	Phenanthrene	1.145	0.859	25.0	96	0.03
24	Anthracene	0.933	0.906	2.9	100	0.00
25	Fluoranthene	1.090	0.961	11.8	101	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
27	Benzydine	0.229	0.270	-17.9	153#	0.00
28	Pvrene	1.528	1.417	7.3	102	0.00
29 S	Terphenyl-d14	0.697	0.669	4.0	97	0.00
30	Benzo(a)anthracene	1.266	1.284	-1.4	105	0.00
31	3,3'-Dichlorobenzidine	0.270	0.215	20.4	108	0.02
32	Chrysene	1.515	1.384	8.6	102	0.00
33	Bis(2-ethylhexyl)phthalate	0.518	0.413	20.3	122	0.00
34	Indeno(1,2,3-cd)pyrene	1.160	1.156	0.3	106	0.01
35 I	Perylene-d12	1.000	1.000	0.0	103	0.01
36	Benzo(b)fluoranthene	1.439	1.388	3.5	97	0.01
37	Benzo(k)fluoranthene	1.294	1.339	-3.5	112	0.00
38 C	Benzo(a)pyrene	1.146	1.107	3.4	104	0.01
39	Dibenzo(a,h)anthracene	0.992	1.042	-5.0	109	0.01
40	Benzo(a,h,i)perylene	1.190	1.177	1.1	105	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
Data File : BM003311.D
Acq On : 27 Nov 2015 19:09
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 27 22:49:17 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				