

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
 Data File : BM003319.D
 Acq On : 27 Nov 2015 23:58
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Nov 28 01:17:54 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.488	0.453	7.2	103	-0.02
3	n-Nitrosodimethylamine	0.462	0.473	-2.4	110	0.00
4 S	2-Fluorophenol	0.986	1.013	-2.7	113	0.00
5 S	Phenol-d6	1.131	1.173	-3.7	115	0.00
6	bis(2-Chloroethyl)ether	1.156	1.163	-0.6	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.222	0.228	-2.7	117	0.00
9	Nitrobenzene	0.243	0.248	-2.1	118	0.00
10	Naphthalene	1.081	1.066	1.4	109	0.01
11	Hexachlorobutadiene	0.164	0.167	-1.8	114	0.00
12	2-Methylnaphthalene	0.678	0.677	0.1	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.090	0.101	-12.2	122	0.00
15 S	2-Fluorobiphenyl	1.614	1.544	4.3	109	0.00
16	Acenaphthylene	1.832	1.874	-2.3	112	0.00
17	Acenaphthene	1.414	1.243	12.1	108	0.00
18	Fluorene	1.427	1.431	-0.3	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.170	0.135	20.6	105	-0.03
21	Hexachlorobenzene	0.176	0.178	-1.1	108	0.00
22	Pentachlorophenol	0.066	0.053	19.7	115	0.00
23	Phenanthrene	1.145	0.932	18.6	99	0.00
24	Anthracene	0.933	1.005	-7.7	105	0.00
25	Fluoranthene	1.090	0.989	9.3	99	-0.02
26 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
27	Benzydine	0.229	0.267	-16.6	152#	0.00
28	Pvrene	1.528	1.341	12.2	97	0.00
29 S	Terphenyl-d14	0.697	0.625	10.3	91	0.00
30	Benzo(a)anthracene	1.266	1.270	-0.3	105	0.00
31	3,3'-Dichlorobenzidine	0.270	0.277	-2.6	140	0.00
32	Chrysene	1.515	1.196	21.1	88	-0.02
33	Bis(2-ethylhexyl)phthalate	0.518	0.347	33.0#	103	-0.02
34	Indeno(1,2,3-cd)pyrene	1.160	1.135	2.2	105	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Benzo(b)fluoranthene	1.439	1.403	2.5	92	0.00
37	Benzo(k)fluoranthene	1.294	1.342	-3.7	105	-0.01
38 C	Benzo(a)pyrene	1.146	1.190	-3.8	105	0.00
39	Dibenzo(a,h)anthracene	0.992	1.102	-11.1	108	-0.01
40	Benzo(a,h,i)perylene	1.190	1.218	-2.4	101	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
Data File : BM003319.D
Acq On : 27 Nov 2015 23:58
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 28 01:17:54 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0