

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
 Data File : BM003319.D
 Acq On : 27 Nov 2015 23:58
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Nov 28 01:17:54 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	107	0.00
2	1,4-Dioxane	1.000	0.994	0.6	103	-0.02
3	n-Nitrosodimethylamine	1.000	1.024	-2.4	110	0.00
4 S	2-Fluorophenol	1.000	1.027	-2.7	113	0.00
5 S	Phenol-d6	1.000	1.037	-3.7	115	0.00
6	bis(2-Chloroethyl)ether	1.000	1.006	-0.6	111	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
8 S	Nitrobenzene-d5	1.000	1.025	-2.5	117	0.00
9	Nitrobenzene	1.000	1.019	-1.9	118	0.00
10	Naphthalene	1.000	0.986	1.4	109	0.01
11	Hexachlorobutadiene	1.000	1.018	-1.8	114	0.00
12	2-Methylnaphthalene	1.000	0.999	0.1	110	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	1.000	1.073	-7.3	122	0.00
15 S	2-Fluorobiphenyl	1.000	0.956	4.4	109	0.00
16	Acenaphthylene	1.000	1.023	-2.3	112	0.00
17	Acenaphthene	1.000	0.879	12.1	108	0.00
18	Fluorene	1.000	1.003	-0.3	111	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	1.000	0.898	10.2	105	-0.03
21	Hexachlorobenzene	1.000	1.013	-1.3	108	0.00
22	Pentachlorophenol	1.000	0.997	0.3	115	0.00
23	Phenanthrene	1.000	0.897	10.3	99	0.00
24	Anthracene	1.000	1.077	-7.7	105	0.00
25	Fluoranthene	1.000	0.907	9.3	99	-0.02
26 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
27	Benzydine	1.000	1.312	-31.2#	152	0.00
28	Pvrene	1.000	0.877	12.3	97	0.00
29 S	Terphenyl-d14	1.000	0.897	10.3	91	0.00
30	Benzo(a)anthracene	1.000	1.003	-0.3	105	0.00
31	3,3'-Dichlorobenzidine	1.000	1.137	-13.7	140	0.00
32	Chrysene	1.000	0.850	15.0	88	-0.02
33	Bis(2-ethylhexyl)phthalate	1.000	0.905	9.5	103	-0.02
34	Indeno(1,2,3-cd)pyrene	1.000	0.978	2.2	105	0.00
35 I	Perylene-d12	5.000	5.000	0.0	96	0.00
36	Benzo(b)fluoranthene	1.000	0.975	2.5	92	0.00
37	Benzo(k)fluoranthene	1.000	1.037	-3.7	105	-0.01
38 C	Benzo(a)pyrene	1.000	1.038	-3.8	105	0.00
39	Dibenzo(a,h)anthracene	1.000	1.112	-11.2	108	-0.01
40	Benzo(a,h,i)perylene	1.000	1.023	-2.3	101	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM112715\
Data File : BM003319.D
Acq On : 27 Nov 2015 23:58
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 28 01:17:54 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0