


Data Path : Z:\HPCHEM1\BNA M\DATA\BM112816\
 Data File : BM008081.D
 Acq On : 28 Nov 2016 09:40
 Operator : UM/SJ
 Sample : DFTPP55
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
 Title : SVOA CALIBRATION
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2534

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	46592	PASS
68	69	0.00	2	1.9	976	PASS
69	198	0.00	100	36.8	51973	PASS
70	69	0.00	2	0.3	163	PASS
127	198	10	80	42.0	59317	PASS
197	198	0.00	2	0.4	588	PASS
198	198	100	100	100.0	141368	PASS
199	198	5	9	6.7	9460	PASS
275	198	10	60	29.1	41176	PASS
365	198	1	100	4.0	5624	PASS
441	443	0.01	100	75.5	17008	PASS
442	198	50	100	79.4	112269	PASS
443	442	15	24	20.1	22532	PASS

Scan 2534 (17.492 min): BM008081.D
DFTPP55

m/z	Abundance
40.00	641.0
44.00	638.0
73.10	709.0
146.90	391.0
207.00	2313.0
208.00	447.0
281.00	799.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
38.10	408.0
39.00	3078.0
40.00	587.0
43.90	717.0
50.10	11639.0
51.10	42408.0
52.00	2104.0
55.00	458.0
56.00	1596.0
57.00	3431.0
61.00	647.0
62.00	641.0
63.10	2088.0
64.00	355.0
65.10	1047.0
67.90	875.0
69.00	47432.0
73.00	1052.0
74.10	5544.0
75.00	7387.0
76.10	2564.0
77.10	52272.0
78.00	3912.0
79.10	3788.0
80.00	3304.0
81.00	4757.0
82.00	934.0
82.90	1056.0
83.90	345.0
85.00	853.0
86.00	1023.0
87.00	439.0
90.90	929.0
92.10	1148.0
93.00	7070.0
94.00	511.0
96.10	483.0
98.00	5219.0
99.00	5014.0
99.90	529.0
101.00	1688.0
103.00	883.0
104.00	1317.0
105.00	1439.0
106.20	406.0
107.00	16063.0
108.00	2815.0
109.10	659.0
110.00	28968.0
111.00	4486.0
112.00	656.0
116.00	1093.0
117.00	16106.0
118.00	1472.0
122.00	1050.0
123.10	1632.0
123.80	574.0
125.00	770.0
126.20	369.0
127.00	52896.0
128.00	4254.0
129.00	20328.0
130.00	2039.0
131.00	633.0
133.10	399.0
134.00	757.0
135.00	1573.0
136.00	651.0
137.00	1341.0
137.90	320.0
141.00	3192.0
142.00	947.0
142.90	768.0
146.00	411.0
147.00	1573.0
148.00	3777.0
149.00	1177.0
150.10	316.0
151.00	450.0
151.90	313.0
153.00	926.0
154.10	802.0
155.10	1303.0
156.00	2172.0
157.00	472.0
158.00	780.0
158.90	392.0
159.90	796.0
161.10	1494.0
162.00	566.0
165.00	1188.0
166.00	926.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

167.00	6116.0
168.00	4530.0
169.00	798.0
170.90	300.0
171.90	344.0
172.90	727.0
174.00	1223.0
175.00	2375.0
176.00	425.0
177.00	1536.0
178.00	504.0
179.00	4352.0
180.00	3625.0
180.90	1493.0
185.00	2315.0
186.00	16392.0
187.00	5298.0
188.00	713.0
188.90	1514.0
191.00	945.0
192.00	1351.0
192.90	1745.0
193.90	469.0
196.10	2830.0
196.80	549.0
198.00	112744.0
199.00	8534.0
199.90	575.0
201.30	378.0
203.00	904.0
204.00	4813.0
205.10	7335.0
206.10	31312.0
207.10	5786.0
208.00	1667.0
208.90	426.0
209.90	406.0
210.90	1478.0
214.90	390.0
215.90	735.0
217.00	8458.0
218.00	1092.0
221.00	3643.0
221.90	1036.0
223.00	1730.0
224.00	18792.0
225.00	4843.0
226.10	609.0
227.00	7261.0
228.00	1080.0
229.00	1274.0
229.80	305.0
231.00	534.0
233.90	647.0
235.00	500.0
236.00	443.0
237.00	798.0
239.80	310.0
241.00	566.0
242.00	866.0
243.00	1018.0
244.00	12457.0
245.00	2168.0
246.00	3371.0
247.00	556.0
249.00	473.0
254.10	524.0
255.00	69128.0
256.00	9971.0
257.00	843.0
258.00	4729.0
259.00	625.0
265.00	1663.0
266.10	360.0
273.00	1818.0
274.00	5152.0
275.00	30712.0
276.00	4592.0
277.00	2818.0
278.00	649.0
281.00	562.0
282.90	549.0
285.10	828.0
293.00	559.0
294.00	304.0
296.00	9205.0
297.00	1534.0
303.00	1224.0
314.10	531.0
315.00	1140.0
316.10	576.0
323.00	2456.0
324.10	660.0
327.00	591.0
328.00	313.0
332.90	402.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

334.10	2146.0
334.90	500.0
346.10	419.0
352.10	730.0
353.00	777.0
354.00	722.0
355.10	537.0
365.00	3755.0
366.00	797.0
372.00	1364.0
373.10	371.0
401.90	470.0
403.00	767.0
421.00	573.0
422.00	628.0
423.10	4015.0
424.10	1370.0
441.00	9488.0
442.10	67184.0
443.00	13901.0
444.00	1528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
38.10	670.0
39.10	3919.0
40.00	683.0
44.00	809.0
49.00	523.0
50.10	14191.0
51.10	57336.0
52.10	2681.0
55.10	623.0
56.10	1904.0
57.10	3822.0
61.00	851.0
62.00	922.0
63.10	2621.0
64.10	315.0
65.00	1180.0
68.00	1340.0
69.00	63152.0
70.00	491.0
73.00	923.0
74.00	7376.0
75.00	10404.0
76.10	3458.0
77.10	68664.0
78.00	4775.0
79.00	5170.0
80.10	4079.0
81.00	5276.0
82.00	1366.0
83.00	1266.0
84.00	362.0
85.10	651.0
86.00	1749.0
87.00	633.0
88.00	317.0
89.10	437.0
91.10	1333.0
92.00	1477.0
93.00	9556.0
93.90	535.0
96.00	646.0
98.00	6905.0
99.10	6565.0
100.00	561.0
101.00	2622.0
103.00	947.0
104.00	1667.0
105.00	1962.0
106.20	417.0
107.00	24136.0
108.10	3284.0
109.00	776.0
110.00	39848.0
111.00	6246.0
112.00	753.0
113.00	355.0
116.00	1158.0
117.00	21472.0
118.00	1646.0
119.00	355.0
119.80	349.0
122.00	1507.0
123.00	2542.0
124.00	833.0
125.10	945.0
127.00	71288.0
128.00	5630.0
129.00	30504.0
130.00	2350.0
131.10	500.0
132.00	472.0
133.10	367.0
134.00	804.0
135.10	2756.0
136.00	1355.0
137.10	1884.0
138.00	348.0
140.00	458.0
141.00	4435.0
142.00	1426.0
142.90	1218.0
146.00	527.0
147.00	2091.0
148.00	5177.0
149.00	1210.0
150.90	544.0
152.00	395.0
153.00	1446.0
154.00	844.0
155.00	2288.0
156.00	3827.0
157.10	701.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

158.00	990.0
159.00	781.0
160.00	1201.0
161.10	2351.0
162.10	530.0
164.10	358.0
165.00	1625.0
165.90	1213.0
167.00	9945.0
168.00	5570.0
169.00	661.0
170.00	332.0
170.90	409.0
171.90	822.0
173.00	973.0
174.00	1938.0
175.00	3848.0
176.10	1063.0
176.90	1946.0
177.90	639.0
179.00	6481.0
180.10	5081.0
181.00	1772.0
181.90	307.0
183.90	648.0
185.00	3090.0
186.00	25720.0
187.00	7515.0
188.00	1443.0
189.00	1587.0
190.00	329.0
191.10	1063.0
192.00	2243.0
193.00	2488.0
193.90	598.0
195.20	331.0
196.00	4773.0
196.80	653.0
197.20	694.0
198.00	170688.0
199.00	11601.0
199.90	1404.0
201.40	918.0
203.00	1539.0
204.00	6589.0
205.10	12140.0
206.10	49688.0
207.10	7740.0
208.10	1966.0
208.90	737.0
210.20	949.0
211.00	1342.0
212.10	363.0
215.00	618.0
216.00	1159.0
217.00	13590.0
218.00	1723.0
221.10	5904.0
221.90	2148.0
223.00	3153.0
224.10	26280.0
225.00	6712.0
226.00	808.0
227.00	12233.0
228.00	1819.0
229.00	2121.0
229.90	402.0
231.00	987.0
233.00	352.0
234.00	613.0
235.00	971.0
236.00	519.0
237.00	1521.0
239.00	614.0
240.10	342.0
241.10	646.0
242.00	1479.0
243.00	1470.0
244.10	19976.0
245.00	3047.0
246.00	4027.0
247.00	1182.0
249.00	798.0
250.90	334.0
252.00	341.0
253.10	522.0
254.10	848.0
255.00	115720.0
256.00	15636.0
257.00	1438.0
258.00	7229.0
259.00	1289.0
265.00	3191.0
266.10	327.0
272.00	362.0
273.00	3408.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

274.00	9249.0
275.00	49040.0
276.00	6897.0
277.00	5198.0
278.00	882.0
281.10	701.0
282.90	790.0
284.10	462.0
285.00	1169.0
290.90	310.0
292.90	797.0
293.90	344.0
294.90	370.0
296.00	16230.0
297.00	2993.0
303.10	2062.0
303.90	569.0
313.90	869.0
315.00	1777.0
315.90	1025.0
321.00	598.0
323.10	4685.0
324.00	1067.0
327.00	894.0
328.00	501.0
331.90	317.0
333.00	424.0
334.00	2770.0
335.00	784.0
341.10	597.0
345.90	1029.0
352.00	1309.0
352.90	975.0
354.00	1630.0
355.00	640.0
365.00	6648.0
366.00	1294.0
370.70	305.0
372.00	2274.0
373.10	914.0
383.00	580.0
390.00	326.0
402.00	718.0
403.00	1124.0
404.00	480.0
421.00	791.0
422.00	831.0
423.10	6115.0
424.00	2256.0
424.90	384.0
441.00	18776.0
442.10	127864.0
443.10	26592.0
444.00	2329.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
38.00	458.0
39.00	2611.0
40.00	539.0
44.00	698.0
48.90	406.0
50.10	11112.0
51.10	40032.0
52.10	1998.0
54.90	455.0
56.00	1295.0
57.00	2966.0
61.10	561.0
62.00	725.0
63.00	1716.0
64.10	310.0
65.10	736.0
68.10	713.0
69.00	45336.0
73.10	1229.0
74.00	5139.0
75.00	6627.0
76.00	2383.0
77.10	48672.0
78.00	3467.0
79.00	3840.0
80.00	2405.0
81.00	3710.0
82.00	1070.0
82.90	582.0
85.00	775.0
86.00	1308.0
87.00	665.0
91.00	1087.0
92.10	1458.0
93.00	6907.0
93.90	512.0
96.10	329.0
98.00	5766.0
99.00	4177.0
99.90	327.0
101.00	2416.0
103.00	1084.0
104.00	1883.0
105.00	1583.0
106.10	628.0
107.00	16221.0
108.00	2839.0
109.00	405.0
110.00	28504.0
111.00	4647.0
111.90	532.0
116.00	1005.0
117.00	15833.0
117.90	1365.0
122.00	921.0
123.00	1676.0
124.00	963.0
125.10	633.0
127.00	53768.0
128.00	4614.0
129.00	21952.0
130.00	1724.0
130.90	333.0
133.00	459.0
134.00	463.0
135.00	1784.0
136.10	647.0
137.10	1322.0
139.90	336.0
141.10	3117.0
142.00	929.0
142.90	715.0
146.10	829.0
147.00	2109.0
148.00	4252.0
149.00	868.0
150.90	323.0
153.00	1171.0
154.00	846.0
155.00	1592.0
156.10	2864.0
156.80	330.0
157.20	417.0
158.10	733.0
159.00	607.0
160.00	1285.0
161.00	1357.0
162.10	336.0
165.00	1422.0
166.00	1314.0
167.00	7294.0
168.00	4264.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

169.00	619.0
171.90	609.0
173.00	799.0
174.00	1274.0
175.00	3017.0
176.00	795.0
177.00	1392.0
177.90	504.0
179.00	5187.0
180.00	4512.0
181.00	1807.0
184.00	395.0
185.00	2528.0
186.00	20848.0
187.10	6025.0
188.00	565.0
189.00	1482.0
191.00	1061.0
192.00	2088.0
193.00	2221.0
194.00	400.0
196.00	3618.0
196.80	562.0
197.10	783.0
198.00	140672.0
199.00	8246.0
200.00	1235.0
201.60	694.0
203.10	1298.0
204.00	5834.0
205.00	9061.0
206.10	39232.0
207.10	6975.0
208.10	1576.0
209.00	712.0
210.20	880.0
211.00	1817.0
211.90	315.0
215.00	565.0
216.00	1067.0
217.00	10658.0
218.00	1393.0
221.00	5134.0
221.80	1887.0
223.10	2403.0
224.10	21864.0
225.00	5823.0
226.00	719.0
227.00	10575.0
228.00	1571.0
229.00	2247.0
231.00	876.0
234.00	731.0
235.00	787.0
236.00	420.0
237.00	1116.0
239.00	397.0
241.10	600.0
242.10	1184.0
243.00	1536.0
244.10	17192.0
245.10	2352.0
246.00	4125.0
247.00	754.0
249.00	848.0
252.90	414.0
254.00	883.0
255.00	98392.0
256.00	14132.0
257.10	1116.0
258.00	6199.0
258.90	863.0
264.90	2782.0
266.00	420.0
271.90	319.0
273.00	2972.0
274.00	8117.0
275.00	43776.0
276.00	5575.0
277.00	4375.0
277.90	546.0
281.00	442.0
282.00	446.0
283.00	580.0
285.00	932.0
293.00	922.0
294.00	402.0
295.10	390.0
296.00	15493.0
297.00	1806.0
303.00	1866.0
304.00	612.0
314.10	995.0
315.00	1602.0
316.10	766.0
321.00	611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

323.10	3865.0
324.10	788.0
327.00	842.0
328.00	548.0
332.00	301.0
333.10	433.0
334.00	3001.0
334.90	778.0
341.00	533.0
346.00	825.0
352.10	1538.0
353.00	938.0
354.10	1571.0
355.10	505.0
365.00	6470.0
366.00	953.0
371.00	357.0
372.00	2164.0
373.00	712.0
383.00	624.0
390.10	312.0
402.00	786.0
403.00	1406.0
404.10	361.0
421.10	1311.0
422.00	1180.0
423.10	7288.0
424.10	2271.0
425.00	460.0
441.10	22760.0
442.10	141760.0
443.10	27104.0
444.10	2413.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55