


Data Path : Z:\HPCHEM1\BNA M\DATA\BM112816\
 Data File : BM008098.D
 Acq On : 29 Nov 2016 08:40
 Operator : UM/SJ
 Sample : DFTPP56
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP56

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
 Title : SVOA CALIBRATION
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2534

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.6	54821	PASS
68	69	0.00	2	1.9	1203	PASS
69	198	0.00	100	36.2	62653	PASS
70	69	0.00	2	0.5	343	PASS
127	198	10	80	42.1	73026	PASS
197	198	0.00	2	0.5	901	PASS
198	198	100	100	100.0	173282	PASS
199	198	5	9	6.9	11881	PASS
275	198	10	60	28.9	50104	PASS
365	198	1	100	3.8	6557	PASS
441	443	0.01	100	78.3	19749	PASS
442	198	50	100	77.3	133960	PASS
443	442	15	24	18.8	25218	PASS

Scan 2534 (17.492 min): BM008098.D
DFTPP56

m/z	Abundance
39.90	430.0
44.00	907.0
73.10	464.0
207.00	2063.0
208.00	531.0
281.10	482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
38.00	633.0
39.10	2776.0
40.00	680.0
43.90	762.0
48.90	414.0
50.10	11909.0
51.10	45312.0
52.00	2709.0
55.20	385.0
56.10	1259.0
57.00	2924.0
61.10	579.0
62.00	753.0
63.00	1990.0
65.10	839.0
68.20	894.0
69.00	49752.0
70.10	534.0
73.10	942.0
74.10	5385.0
75.10	8460.0
76.10	3087.0
77.10	55088.0
78.10	4655.0
79.00	4064.0
80.00	3060.0
81.10	3978.0
82.10	1141.0
83.00	1075.0
83.90	457.0
85.10	698.0
86.00	1253.0
87.00	631.0
91.00	905.0
92.00	981.0
93.00	6528.0
94.00	514.0
95.10	357.0
96.10	700.0
98.00	5448.0
99.00	4667.0
100.10	323.0
101.00	2306.0
103.10	646.0
104.10	1316.0
104.90	1184.0
106.10	427.0
107.00	16968.0
108.00	2626.0
109.10	706.0
110.00	30784.0
111.00	4426.0
111.90	403.0
115.90	831.0
117.00	15513.0
118.00	1803.0
122.10	1054.0
123.00	1691.0
123.90	877.0
124.90	637.0
127.00	53528.0
128.00	4181.0
129.00	22472.0
130.00	2127.0
131.00	400.0
133.00	341.0
134.00	747.0
135.00	2199.0
136.00	898.0
137.00	1209.0
140.00	377.0
141.00	3667.0
141.90	821.0
143.10	705.0
145.90	596.0
147.10	1717.0
148.00	3770.0
149.00	875.0
151.20	368.0
153.10	1101.0
154.00	832.0
155.00	1894.0
156.00	2500.0
157.00	316.0
158.00	601.0
158.90	567.0
160.00	1094.0
161.00	1295.0
162.00	368.0
165.00	1286.0
166.00	819.0
167.10	6149.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

168.00	3990.0
169.00	672.0
172.00	539.0
173.10	676.0
174.00	1645.0
175.00	2680.0
176.20	529.0
177.00	1262.0
177.90	346.0
179.00	4723.0
180.00	3439.0
181.00	1192.0
184.00	461.0
185.10	1724.0
186.00	16010.0
187.00	4411.0
187.90	580.0
189.10	1346.0
191.00	938.0
192.00	1426.0
193.00	1747.0
193.90	332.0
196.00	3270.0
196.90	711.0
198.00	116008.0
199.00	8419.0
200.10	614.0
201.40	438.0
203.00	913.0
204.00	4770.0
205.00	7386.0
206.10	32816.0
207.10	6575.0
208.10	1285.0
209.00	757.0
210.10	763.0
211.00	1266.0
212.00	349.0
214.90	357.0
216.00	889.0
217.00	8698.0
218.00	1058.0
220.30	309.0
221.00	3824.0
221.90	1137.0
223.10	1862.0
224.00	18296.0
225.10	4480.0
226.00	455.0
227.00	8196.0
228.00	1426.0
229.00	1762.0
231.00	664.0
234.00	655.0
235.00	993.0
235.90	497.0
237.10	914.0
241.00	338.0
241.90	1101.0
242.90	695.0
244.00	13366.0
245.00	1707.0
246.00	3060.0
247.10	637.0
248.90	704.0
254.10	537.0
255.00	72480.0
256.00	11210.0
257.10	785.0
258.00	3988.0
259.10	645.0
265.00	2114.0
266.10	356.0
273.00	1929.0
274.00	5291.0
275.00	30336.0
276.00	4774.0
277.00	3271.0
278.10	515.0
281.10	670.0
283.00	508.0
285.00	516.0
293.00	657.0
296.10	9763.0
297.00	1242.0
303.10	854.0
313.90	361.0
314.90	851.0
316.10	400.0
323.10	2595.0
324.00	623.0
327.00	679.0
327.90	330.0
334.00	2238.0
335.00	762.0
341.00	368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

346.00	408.0
352.00	1092.0
353.10	555.0
354.00	921.0
355.10	366.0
365.00	4179.0
366.00	449.0
372.00	1438.0
373.00	375.0
383.00	319.0
402.00	506.0
403.10	712.0
404.20	312.0
420.90	367.0
422.00	559.0
423.00	4013.0
424.00	1239.0
441.10	10096.0
442.10	68248.0
443.00	12607.0
444.00	1863.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
36.90	350.0
38.10	865.0
39.00	4528.0
40.10	847.0
41.00	430.0
44.00	719.0
49.10	649.0
50.10	16164.0
51.10	65712.0
52.10	3982.0
55.00	507.0
56.00	2164.0
57.00	4715.0
57.90	325.0
61.00	913.0
62.10	1263.0
63.10	3347.0
64.00	651.0
65.10	1738.0
67.10	339.0
68.00	1454.0
69.00	76104.0
73.10	1661.0
74.10	8334.0
75.10	12282.0
76.10	4251.0
77.10	85120.0
78.10	5666.0
79.00	6488.0
80.00	4201.0
81.00	6635.0
82.10	1664.0
83.00	1684.0
85.10	949.0
86.00	1920.0
87.00	1023.0
91.00	1628.0
92.00	1675.0
93.00	11019.0
93.90	623.0
95.10	322.0
96.00	601.0
97.30	409.0
98.00	9014.0
99.00	7044.0
100.00	653.0
101.00	2966.0
102.00	418.0
103.10	1044.0
104.00	2587.0
105.00	2207.0
106.00	708.0
107.00	26608.0
108.00	4084.0
108.90	631.0
110.00	48592.0
111.00	7447.0
112.00	679.0
113.10	330.0
116.10	1678.0
117.00	23296.0
118.00	2099.0
120.10	517.0
122.00	2070.0
123.00	3070.0
124.00	978.0
125.00	1138.0
126.20	401.0
127.00	87416.0
128.00	6831.0
129.00	36136.0
130.00	3035.0
131.00	815.0
131.90	321.0
132.90	310.0
134.00	987.0
135.00	3089.0
136.00	1488.0
136.90	1710.0
138.00	319.0
140.00	482.0
141.00	5495.0
142.10	1361.0
143.00	1240.0
144.90	308.0
146.10	669.0
147.00	2676.0
148.00	5841.0
149.10	1261.0
150.00	649.0
150.90	561.0
151.60	361.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

153.00	1322.0
154.10	1385.0
155.00	2907.0
156.10	3769.0
157.10	849.0
157.90	957.0
159.10	652.0
160.10	1593.0
161.00	2031.0
162.00	598.0
163.00	324.0
165.00	2154.0
166.10	1806.0
167.00	9217.0
168.00	5163.0
169.10	1129.0
171.00	427.0
172.00	958.0
173.00	943.0
174.00	2248.0
175.00	3815.0
176.00	950.0
176.90	2030.0
177.90	707.0
179.00	7725.0
180.00	5759.0
181.10	2692.0
184.00	739.0
185.00	3951.0
186.00	29728.0
187.00	8553.0
188.00	1264.0
189.00	2194.0
190.00	309.0
191.00	1113.0
192.00	2886.0
193.00	3045.0
194.00	570.0
194.90	415.0
196.00	5685.0
196.90	1141.0
198.00	204800.0
199.00	13981.0
200.00	1012.0
201.60	849.0
202.90	1413.0
204.00	8196.0
205.00	13491.0
206.10	57744.0
207.00	9058.0
208.00	2503.0
209.00	1025.0
209.90	1048.0
210.30	810.0
211.00	2277.0
211.90	726.0
215.00	816.0
216.00	1350.0
217.00	14931.0
218.00	1660.0
221.00	7536.0
222.00	1898.0
223.00	3098.0
224.10	31560.0
225.10	7800.0
226.00	804.0
227.00	14898.0
228.00	1960.0
229.00	2836.0
230.10	526.0
231.00	1419.0
232.00	301.0
232.90	373.0
234.00	973.0
234.90	975.0
235.90	801.0
237.10	1011.0
239.00	566.0
240.20	400.0
241.00	901.0
242.00	1818.0
243.00	1592.0
244.00	24936.0
245.10	2728.0
246.00	5790.0
247.00	852.0
249.10	918.0
253.00	825.0
254.10	902.0
255.00	131000.0
256.00	18480.0
257.00	2090.0
258.00	8707.0
258.90	1492.0
264.00	340.0
265.00	3610.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

266.00	420.0
271.00	321.0
271.80	362.0
273.00	3768.0
274.00	9880.0
275.00	58240.0
276.00	7144.0
277.00	4885.0
278.00	1120.0
281.10	673.0
283.00	1031.0
284.10	377.0
285.00	979.0
292.90	1508.0
294.90	435.0
296.00	16880.0
297.10	2559.0
303.00	2283.0
304.00	806.0
314.00	759.0
315.00	1656.0
316.00	1036.0
321.00	491.0
323.10	5459.0
324.00	705.0
327.00	1344.0
328.00	514.0
331.90	537.0
333.00	409.0
334.00	3241.0
335.10	1121.0
341.10	659.0
342.10	327.0
346.00	1016.0
352.00	1802.0
353.00	1414.0
354.10	1831.0
365.00	7079.0
366.00	1068.0
372.10	2517.0
373.00	608.0
383.00	770.0
390.00	311.0
402.00	1133.0
403.00	1112.0
404.00	630.0
421.00	1149.0
422.00	1061.0
423.10	8246.0
424.10	2610.0
441.10	21080.0
442.00	150144.0
443.10	27296.0
444.10	3156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

m/z	Abundance
38.10	590.0
39.10	4035.0
40.00	731.0
41.00	386.0
44.00	707.0
49.10	506.0
50.10	13380.0
51.10	53440.0
52.10	2614.0
55.10	462.0
56.10	1878.0
57.00	3671.0
61.00	918.0
62.00	844.0
63.00	2179.0
64.00	403.0
65.00	1186.0
65.90	309.0
67.00	303.0
68.10	1263.0
69.00	62104.0
69.90	496.0
73.10	1111.0
74.10	7461.0
75.10	10096.0
76.10	3469.0
77.10	74728.0
78.00	4539.0
79.00	4766.0
80.00	3922.0
81.00	5305.0
81.90	1667.0
83.00	1227.0
85.10	1054.0
86.00	1860.0
87.10	819.0
88.00	375.0
91.10	1154.0
92.00	1624.0
93.00	9581.0
94.00	660.0
96.10	673.0
98.00	7708.0
99.00	5870.0
99.90	472.0
101.00	3474.0
103.00	1164.0
104.00	1990.0
105.00	2025.0
105.90	694.0
107.00	24992.0
108.00	3354.0
109.10	712.0
110.00	42928.0
111.00	6298.0
112.00	836.0
113.00	457.0
116.00	1791.0
117.00	22840.0
118.00	1518.0
120.00	418.0
122.00	1560.0
123.00	2804.0
124.00	1119.0
125.10	1113.0
127.00	78136.0
128.00	5604.0
129.00	32760.0
130.00	3114.0
130.90	346.0
132.10	307.0
133.10	472.0
134.00	828.0
135.00	2744.0
136.00	1116.0
137.10	1369.0
139.90	395.0
141.00	4855.0
142.10	1425.0
143.10	1146.0
144.00	331.0
146.10	1170.0
147.00	3225.0
148.00	5505.0
148.90	1660.0
150.00	495.0
151.00	619.0
151.70	382.0
153.10	1951.0
153.90	1020.0
155.10	2752.0
156.00	3606.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

157.00	818.0
158.10	739.0
159.10	631.0
160.10	1247.0
161.00	2055.0
162.10	517.0
162.90	328.0
165.00	2037.0
166.00	1350.0
167.00	9812.0
168.10	5142.0
169.10	880.0
170.70	347.0
172.00	921.0
173.00	1599.0
174.10	1940.0
175.00	4201.0
176.00	1162.0
177.10	1978.0
178.00	811.0
179.00	6982.0
180.10	6084.0
181.00	2358.0
181.90	385.0
184.00	596.0
185.00	4280.0
186.00	29424.0
187.10	9175.0
188.00	881.0
189.00	1704.0
190.00	447.0
191.00	1100.0
192.00	2649.0
193.00	2902.0
194.00	604.0
196.00	5019.0
196.80	853.0
197.10	891.0
198.00	199040.0
199.00	13245.0
200.00	802.0
201.00	528.0
201.50	1045.0
203.00	1617.0
204.00	7790.0
205.00	13461.0
206.10	55368.0
207.10	9364.0
208.10	2306.0
209.00	1013.0
210.10	858.0
211.10	2444.0
211.70	444.0
215.00	982.0
216.00	1328.0
217.00	13997.0
218.00	1659.0
219.30	343.0
221.10	7705.0
222.10	1485.0
223.00	3352.0
224.00	31760.0
225.00	7938.0
226.00	1013.0
227.00	13594.0
228.10	2060.0
229.00	2727.0
230.00	474.0
231.00	1210.0
233.00	370.0
234.10	1024.0
234.90	1256.0
235.90	671.0
237.00	898.0
238.90	776.0
240.10	432.0
241.00	935.0
242.00	1665.0
243.10	1648.0
244.00	26008.0
245.10	2866.0
246.00	5394.0
247.00	939.0
249.00	1035.0
253.00	857.0
254.10	1101.0
255.00	137152.0
256.00	18728.0
257.00	1327.0
258.00	7509.0
259.00	1793.0
265.00	3631.0
266.00	459.0
272.00	351.0
273.00	4516.0
274.00	12430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56

275.00	61736.0
276.00	8034.0
277.00	6082.0
277.90	1325.0
281.10	797.0
281.90	393.0
283.00	915.0
284.10	550.0
285.00	856.0
289.00	335.0
293.00	1466.0
294.00	324.0
294.90	419.0
296.00	17824.0
297.00	2412.0
301.90	465.0
303.10	2637.0
304.10	593.0
314.20	890.0
315.00	2240.0
316.00	1560.0
321.00	493.0
323.00	5968.0
324.10	1018.0
327.00	1242.0
328.00	551.0
332.00	376.0
333.00	510.0
334.10	3682.0
335.00	975.0
341.10	761.0
345.90	867.0
352.10	1367.0
353.10	1297.0
354.00	1868.0
355.10	369.0
365.00	8413.0
366.00	1475.0
371.00	801.0
372.00	3065.0
372.90	936.0
383.00	885.0
390.00	456.0
402.10	1342.0
403.00	2333.0
404.10	1037.0
421.10	1433.0
422.00	1568.0
423.10	10335.0
424.00	2691.0
425.00	401.0
440.10	341.0
441.00	28072.0
442.10	183488.0
443.10	35752.0
444.00	2837.0

Instrument :
BNA_M
ClientSampleId :
DFTPP56