

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112818\
 Data File : BM017787.D
 Acq On : 29 Nov 2018 00:45
 Operator : JU/SJ
 Sample : J5979-03MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ETO65MS

Quant Time: Nov 29 05:53:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 29 00:55:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	171547	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	765291	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.23	164	474124	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.99	188	1090533	20.00	ng/ul	0.00
77) Chrysene-d12	21.19	240	1188890	20.00	ng/ul	0.00
85) Perylene-d12	23.38	264	1142339	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.18	96	13650	3.64	ng/uL	0.00
5) Phenol-d5	6.77	99	259712	16.58	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	188334	19.93	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	223497	18.23	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	221692	17.32	ng/ul	0.00
19) Nitrobenzene-d5	8.74	128	118289	20.12	ng/ul	0.00
22) 2-Nitrophenol-d4	9.46	143	122622	18.21	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.99	165	217858	16.99	ng/ul	0.00
29) 4-Chloroaniline-d4	10.51	131	308405	28.03	ng/ul	0.00
43) Dimethylphthalate-d6	13.66	166	866676	21.07	ng/ul	0.00
46) Acenaphthylene-d8	13.92	160	925104	18.58	ng/ul	0.00
51) 4-Nitrophenol-d4	14.46	143	97608	13.05	ng/ul	0.00
57) Fluorene-d10	15.23	176	716874	20.18	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.37	200	100354	12.87	ng/ul	0.00
70) Anthracene-d10	17.09	188	1046372	19.27	ng/ul	0.00
78) Pyrene-d10	19.39	212	1149085	19.97	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.24	264	1125377	18.16	ng/ul	0.00

Target Compounds

					Ovalue
6) Phenol	6.80	94	268006	17.137	ng/ul 98
10) 2-Chlorophenol	7.16	128	192686	15.732	ng/ul 97
15) N-Nitroso-di-n-propylamine	8.39	70	187910	18.056	ng/ul 98
33) 4-Chloro-3-methylphenol	11.67	107	210672	15.322	ng/ul 96
44) Dimethylphthalate	13.70	163	130119	3.265	ng/ul 99
49) Acenaphthene	14.30	153	565016	16.861	ng/ul 99
52) 4-Nitrophenol	14.47	109	80105	13.021	ng/ul 95
54) 2,4-Dinitrotoluene	14.63	165	206983	17.455	ng/ul 87
68) Pentachlorophenol	16.64	266	116276	13.493	ng/ul 98
79) Pyrene	19.42	202	1278362	17.713	ng/ul 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112818\
Data File : BM017787.D
Acq On : 29 Nov 2018 00:45
Operator : JU/SJ
Sample : J5979-03MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ETO65MS

Quant Time: Nov 29 05:53:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11918MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 29 00:55:29 2018
Response via : Initial Calibration

