


Data Path : Z:\HPCHEM1\BNA M\DATA\BM112916\
Data File : BM008102.D
Acq On : 29 Nov 2016 15:17
Operator : UM/SJ
Sample : DFTPP57
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM112916.M
Title : SVOA CALIBRATION
Last Update : Wed Nov 30 10:28:38 2016



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2533

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.3	47472	PASS
68	69	0.00	2	1.3	704	PASS
69	198	0.00	100	33.7	54616	PASS
70	69	0.00	2	0.4	212	PASS
127	198	10	80	41.6	67349	PASS
197	198	0.00	2	0.4	606	PASS
198	198	100	100	100.0	162090	PASS
199	198	5	9	6.8	11081	PASS
275	198	10	60	30.3	49090	PASS
365	198	1	100	4.2	6793	PASS
441	443	0.01	100	77.7	20929	PASS
442	198	50	100	87.4	141605	PASS
443	442	15	24	19.0	26928	PASS

m/z	Abundance
39.10	349.0
40.00	711.0
41.00	515.0
42.00	552.0
43.10	817.0
44.00	1867.0
55.10	823.0
56.10	359.0
57.10	937.0
60.00	1296.0
66.90	300.0
69.10	594.0
70.10	392.0
70.90	505.0
73.10	1727.0
74.00	370.0
77.10	695.0
83.90	501.0
91.00	363.0
96.90	394.0
191.00	421.0
193.00	470.0
207.00	2257.0
208.00	415.0
281.10	891.0
282.10	341.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	816.0
39.10	3846.0
40.00	679.0
41.00	798.0
42.00	436.0
43.10	760.0
44.00	2114.0
45.00	693.0
49.00	554.0
50.10	14096.0
51.10	52560.0
52.10	2362.0
53.10	612.0
55.10	1004.0
56.00	1575.0
57.00	4655.0
60.10	1374.0
61.10	921.0
61.90	917.0
63.10	2302.0
65.00	1166.0
68.10	1094.0
69.00	60136.0
69.90	637.0
70.90	483.0
73.10	2889.0
74.00	6844.0
75.00	10557.0
76.10	3340.0
77.10	68336.0
78.10	4738.0
79.00	4684.0
80.00	3897.0
81.00	5129.0
82.00	1461.0
83.10	1228.0
83.90	538.0
85.00	1281.0
86.00	1742.0
87.00	753.0
88.10	368.0
88.90	373.0
91.00	1411.0
92.10	1543.0
93.00	8992.0
94.00	589.0
95.10	346.0
96.10	504.0
97.00	443.0
98.00	7702.0
99.00	5282.0
100.00	314.0
101.00	2557.0
103.00	1076.0
104.00	2161.0
105.00	1972.0
105.90	603.0
107.00	22216.0
108.00	3038.0
109.00	819.0
110.00	37904.0
111.10	5750.0
112.00	841.0
115.90	981.0
117.00	19552.0
118.10	1000.0
120.00	380.0
121.00	321.0
122.10	1502.0
123.00	2245.0
124.00	1018.0
125.00	857.0
127.00	70656.0
128.10	6035.0
129.00	28168.0
130.00	2851.0
130.90	540.0
132.90	385.0
134.00	1106.0
135.00	2673.0
136.00	1113.0
137.10	1312.0
139.90	430.0
141.00	4549.0
142.00	1139.0
143.00	648.0
143.90	382.0
145.90	592.0
147.00	1956.0
148.00	5144.0
149.00	1403.0
149.90	597.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

151.20	746.0
153.00	1450.0
154.00	822.0
155.00	2325.0
156.00	3172.0
157.00	753.0
157.90	673.0
158.90	516.0
160.00	1347.0
161.10	1865.0
162.00	419.0
164.00	314.0
165.00	1967.0
166.00	1198.0
167.00	8333.0
168.00	4951.0
169.00	740.0
169.90	521.0
170.90	326.0
171.90	771.0
173.00	1209.0
174.00	1290.0
175.00	2877.0
176.00	1191.0
177.00	1783.0
178.00	528.0
179.00	6170.0
180.00	4597.0
181.00	1791.0
182.00	353.0
182.80	367.0
184.00	730.0
185.00	2692.0
186.00	24160.0
187.00	7143.0
187.90	666.0
188.90	1727.0
189.90	315.0
191.00	1037.0
192.00	2495.0
193.00	2449.0
194.00	674.0
195.20	375.0
196.00	4678.0
196.80	952.0
198.00	155776.0
199.00	11002.0
200.00	917.0
201.40	715.0
201.80	352.0
203.00	1495.0
204.00	6167.0
205.00	10110.0
206.10	44536.0
207.00	7588.0
208.00	2149.0
209.10	829.0
210.00	1033.0
211.00	1640.0
211.80	300.0
212.10	319.0
215.00	849.0
216.00	1027.0
217.00	10955.0
218.00	1489.0
221.00	5339.0
221.70	1979.0
223.00	2712.0
224.00	23912.0
225.10	6700.0
226.10	802.0
227.00	11351.0
228.00	1486.0
229.00	2123.0
230.00	424.0
231.00	1107.0
234.00	501.0
235.00	948.0
235.90	718.0
237.00	987.0
239.00	466.0
241.00	561.0
242.00	1287.0
243.10	1155.0
244.00	18336.0
245.10	3061.0
246.00	4048.0
246.90	994.0
248.90	714.0
249.90	344.0
252.90	505.0
254.10	1112.0
255.00	102528.0
256.00	14950.0
257.10	1185.0
258.00	5934.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

259.00	1149.0
264.90	2337.0
266.00	411.0
273.00	3318.0
274.00	8922.0
275.00	45096.0
276.00	5937.0
277.00	4298.0
278.00	752.0
281.00	802.0
283.00	638.0
284.90	993.0
293.00	892.0
295.10	540.0
296.00	13824.0
297.00	2074.0
303.00	1528.0
304.10	509.0
313.90	610.0
315.00	1335.0
316.00	585.0
321.10	397.0
323.00	4299.0
324.10	731.0
327.00	917.0
327.90	447.0
332.90	384.0
334.00	2501.0
335.10	801.0
341.20	585.0
346.00	1129.0
351.90	953.0
353.00	695.0
354.00	1315.0
355.00	425.0
365.00	6297.0
366.00	928.0
372.00	1944.0
373.10	538.0
383.00	505.0
402.00	713.0
403.00	1150.0
404.00	481.0
421.00	669.0
421.90	787.0
423.10	5755.0
424.10	1611.0
441.10	16029.0
442.10	108784.0
443.00	21496.0
444.10	1908.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
37.00	404.0
38.00	845.0
39.00	4335.0
40.10	525.0
41.00	668.0
41.90	499.0
43.10	901.0
44.00	1834.0
44.90	505.0
49.10	920.0
50.00	15294.0
51.10	56280.0
52.00	2655.0
55.00	865.0
55.90	1913.0
57.10	4800.0
58.10	322.0
60.00	918.0
61.00	1000.0
62.00	918.0
63.00	2542.0
65.10	1161.0
67.00	383.0
68.20	1019.0
69.00	66368.0
70.10	631.0
71.00	360.0
73.10	1790.0
74.10	6966.0
75.00	9955.0
76.10	4321.0
77.10	74184.0
78.10	4936.0
79.10	6462.0
80.00	4266.0
81.00	5882.0
82.00	1556.0
82.90	1203.0
83.90	669.0
85.10	1328.0
86.00	2012.0
87.00	1038.0
88.00	383.0
91.00	1910.0
92.00	1865.0
93.00	9320.0
94.00	842.0
94.90	435.0
96.00	583.0
97.00	448.0
98.00	8207.0
99.00	5795.0
99.90	652.0
101.00	3133.0
103.00	1279.0
104.00	1949.0
105.10	1979.0
106.00	748.0
107.00	26328.0
108.00	4002.0
108.90	823.0
110.00	43904.0
111.00	7758.0
112.00	965.0
116.00	1610.0
117.00	21112.0
118.10	1948.0
119.00	373.0
120.00	354.0
122.00	1695.0
123.00	2976.0
124.00	1348.0
125.10	1013.0
125.80	336.0
127.00	80192.0
128.00	5285.0
129.00	34368.0
129.90	2573.0
131.00	629.0
133.10	468.0
134.00	1119.0
135.00	3206.0
136.10	1368.0
137.00	1216.0
137.90	492.0
139.00	338.0
140.00	606.0
141.00	5519.0
142.00	1636.0
143.00	1228.0
144.00	356.0
145.00	300.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

146.00	1095.0
147.00	3160.0
148.00	6305.0
149.00	1327.0
149.90	440.0
150.90	630.0
151.80	384.0
153.00	1450.0
153.90	1252.0
155.00	2316.0
156.00	4002.0
157.00	854.0
157.90	991.0
158.90	695.0
160.00	1314.0
161.10	2106.0
162.10	681.0
163.00	412.0
165.00	1854.0
166.00	1707.0
167.00	9716.0
168.00	4632.0
169.10	982.0
170.00	448.0
171.00	367.0
172.00	1066.0
173.00	1571.0
173.90	2016.0
175.00	4084.0
176.10	843.0
177.00	1837.0
177.90	632.0
179.00	7016.0
180.00	5170.0
181.00	2250.0
182.20	414.0
183.90	740.0
185.00	3468.0
186.00	25952.0
187.00	8442.0
188.10	754.0
189.00	1769.0
191.00	1153.0
192.00	2308.0
193.10	3045.0
194.00	503.0
196.00	5429.0
197.10	1038.0
198.00	194752.0
199.00	13277.0
200.00	1096.0
201.20	377.0
201.70	525.0
203.00	1350.0
204.00	7696.0
205.00	12519.0
206.10	55392.0
207.00	9256.0
208.10	2692.0
209.00	1166.0
210.20	743.0
211.10	1945.0
215.00	1004.0
216.00	1288.0
217.00	14188.0
218.00	2220.0
221.10	7224.0
221.80	1912.0
223.00	3636.0
224.00	32816.0
225.00	7904.0
226.00	889.0
227.00	14164.0
228.00	1705.0
229.00	2866.0
230.00	590.0
231.00	981.0
232.90	322.0
234.00	905.0
235.00	1138.0
235.80	573.0
237.00	1046.0
238.90	568.0
239.90	353.0
241.00	933.0
242.10	1789.0
243.00	1745.0
244.00	23264.0
245.00	3052.0
246.00	4366.0
247.00	973.0
249.10	1116.0
249.90	308.0
250.90	309.0
252.90	709.0
254.00	1111.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

255.00	130952.0
256.00	18416.0
257.00	1894.0
258.00	8290.0
259.00	1337.0
260.00	325.0
265.00	3963.0
265.90	548.0
272.00	390.0
273.00	3269.0
274.00	10468.0
275.00	58664.0
276.00	8418.0
277.00	4767.0
278.00	872.0
281.00	844.0
282.00	424.0
282.90	694.0
284.20	401.0
285.10	1022.0
293.00	1257.0
294.10	333.0
295.00	376.0
296.00	18360.0
297.00	2702.0
303.00	1914.0
304.10	647.0
314.10	813.0
315.00	2167.0
315.90	865.0
321.00	396.0
322.00	341.0
323.10	5917.0
324.10	781.0
327.00	789.0
328.00	689.0
331.90	395.0
333.00	707.0
334.10	4159.0
335.10	863.0
340.90	780.0
346.00	1185.0
352.00	1561.0
353.00	1036.0
354.10	1524.0
355.10	382.0
365.00	7853.0
365.90	896.0
371.00	362.0
372.10	3450.0
373.10	955.0
383.00	840.0
383.90	321.0
391.00	408.0
401.90	920.0
403.10	1726.0
404.00	734.0
421.00	1433.0
422.00	1230.0
423.00	8962.0
424.00	2097.0
441.00	23632.0
442.10	160704.0
443.00	30288.0
444.10	3114.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
38.10	487.0
39.00	2932.0
40.00	593.0
41.00	761.0
42.00	492.0
43.10	1125.0
44.00	1802.0
45.10	502.0
49.00	716.0
50.00	8013.0
51.10	33576.0
52.00	1814.0
53.00	357.0
55.00	853.0
56.00	1398.0
57.00	3318.0
58.10	382.0
60.00	1062.0
61.00	542.0
62.10	486.0
63.00	1435.0
65.10	991.0
67.90	881.0
69.00	39128.0
70.00	560.0
70.90	423.0
73.10	1903.0
74.10	4343.0
75.00	7101.0
76.10	2353.0
77.10	46640.0
78.10	2660.0
79.00	3716.0
80.00	2163.0
81.00	3529.0
81.90	792.0
83.00	1047.0
83.90	377.0
85.00	867.0
86.00	1316.0
86.90	637.0
88.10	330.0
91.00	1188.0
92.00	1023.0
93.00	6127.0
94.00	652.0
95.00	351.0
96.00	619.0
97.10	468.0
98.10	4992.0
99.00	4021.0
100.00	512.0
101.10	2266.0
102.90	478.0
104.00	1214.0
105.00	1281.0
107.00	15780.0
108.00	3100.0
109.00	551.0
110.00	27584.0
111.00	4810.0
112.00	792.0
116.00	839.0
117.00	14996.0
118.00	1128.0
122.00	1183.0
123.00	1801.0
124.00	961.0
125.00	826.0
126.20	357.0
127.00	51200.0
128.10	3596.0
129.00	21728.0
129.90	1771.0
130.90	553.0
133.00	494.0
134.10	575.0
135.00	2034.0
136.00	793.0
137.00	979.0
141.00	3353.0
142.00	911.0
143.00	620.0
146.00	607.0
147.10	2334.0
148.00	3646.0
149.00	1078.0
150.90	423.0
151.60	383.0
152.90	1061.0
154.10	558.0
155.00	2050.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

156.00	2328.0
156.90	356.0
158.00	768.0
158.90	425.0
160.00	894.0
161.00	1558.0
162.10	384.0
162.90	302.0
165.00	1321.0
166.00	1113.0
167.00	6049.0
168.10	3243.0
169.00	446.0
172.10	516.0
173.10	856.0
173.90	1141.0
175.00	2297.0
176.10	921.0
177.00	1336.0
178.10	461.0
179.00	4765.0
180.00	3847.0
181.00	1503.0
183.10	300.0
184.00	458.0
185.10	2306.0
186.00	19152.0
187.10	5484.0
188.00	591.0
188.90	1378.0
191.00	847.0
192.00	1531.0
193.10	2208.0
194.00	506.0
196.10	3869.0
196.90	866.0
198.00	135744.0
199.00	8965.0
200.00	377.0
201.40	727.0
203.00	934.0
204.00	5255.0
205.00	8402.0
206.10	36592.0
207.00	7186.0
208.00	2339.0
209.10	694.0
210.10	590.0
211.10	1508.0
211.90	335.0
214.90	580.0
216.10	824.0
217.00	11281.0
218.00	1560.0
221.10	4714.0
221.80	1437.0
223.00	2560.0
224.00	21720.0
225.10	5851.0
226.00	975.0
227.00	10137.0
228.00	1205.0
229.00	2192.0
231.00	861.0
233.90	655.0
235.00	744.0
236.10	684.0
237.00	1072.0
239.00	402.0
241.10	455.0
242.00	1251.0
243.00	1421.0
244.10	17184.0
245.00	1984.0
246.00	3667.0
247.00	825.0
249.10	663.0
252.90	485.0
253.90	607.0
255.00	94456.0
256.00	14074.0
257.00	1177.0
258.00	5910.0
258.90	816.0
261.00	304.0
265.00	2858.0
266.00	436.0
266.90	340.0
273.00	2959.0
274.00	7800.0
275.00	43512.0
276.00	5503.0
277.00	3679.0
278.00	815.0
281.10	810.0
283.10	615.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

284.00	565.0
284.90	827.0
293.00	744.0
293.90	347.0
296.00	14186.0
297.00	2123.0
303.00	1796.0
304.00	539.0
310.00	304.0
313.90	607.0
315.00	1640.0
316.00	974.0
323.10	4046.0
324.10	730.0
326.90	926.0
327.90	468.0
332.00	323.0
332.90	326.0
334.00	3098.0
335.10	807.0
341.00	736.0
345.90	1035.0
352.00	1237.0
353.10	867.0
354.00	1204.0
355.00	558.0
365.00	6231.0
366.00	939.0
371.00	438.0
372.00	2343.0
373.00	543.0
382.80	552.0
383.90	309.0
402.00	937.0
403.10	1127.0
404.10	318.0
421.00	909.0
422.00	944.0
423.00	8737.0
424.00	1989.0
441.10	23128.0
442.10	155328.0
443.10	29000.0
444.00	2648.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57