

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112922\
 Data File : BM037795.D
 Acq On : 29 Nov 2022 17:31
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4182

Quant Time: Nov 29 23:36:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 23:34:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	5595	0.400	ng/ul	0.00
4) Naphthalene-d8	10.927	136	17321	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.730	164	8875	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.467	188	16441	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.629	240	9597	0.400	ng/ul	# 0.00
23) Perylene-d12	24.116	264	7872	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.439	96	3104	0.477	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.499	152	10577	0.368	ng/ul	0.00
18) Fluoranthene-d10	19.482	212	16022	0.505	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.472	88	3129	0.462	ng/ul#	56
5) Naphthalene	10.976	128	22587	0.429	ng/ul	98
7) 2-Methylnaphthalene	12.571	142	13838	0.387	ng/ul	98
8) 1-Methylnaphthalene	12.791	142	14226	0.388	ng/ul	99
10) Acenaphthylene	14.452	152	18620	0.378	ng/ul	99
11) Acenaphthene	14.790	153	15345	0.446	ng/ul	99
12) Fluorene	15.771	166	16403	0.396	ng/ul	98
14) Pentachlorophenol	17.113	266	1853	0.502	ng/ul	99
15) Phenanthrene	17.510	178	24793	0.441	ng/ul	99
16) Anthracene	17.598	178	21977	0.405	ng/ul	98
19) Fluoranthene	19.510	202	25000	0.540	ng/ul#	85
20) Pyrene	19.872	202	25816	0.536	ng/ul#	86
21) Benzo(a)anthracene	21.614	228	18608	0.434	ng/ul	99
22) Chrysene	21.670	228	18569	0.436	ng/ul	99
24) Benzo(b)fluoranthene	23.350	252	17840	0.497	ng/ul#	80
25) Benzo(k)fluoranthene	23.403	252	17101	0.472	ng/ul#	80
26) Benzo(a)pyrene	24.002	252	14518	0.465	ng/ul#	70
27) Indeno(1,2,3-cd)pyrene	26.712	276	17857	0.432	ng/ul#	78
28) Dibenzo(a,h)anthracene	26.732	278	13860	0.427	ng/ul#	86
29) Benzo(g,h,i)perylene	27.514	276	14913	0.425	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112922\
Data File : BM037795.D
Acq On : 29 Nov 2022 17:31
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4182

Quant Time: Nov 29 23:36:59 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM11622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 23:34:25 2022
Response via : Initial Calibration

