


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112922\
 Data File : BM037783.D
 Acq On : 29 Nov 2022 08:32
 Operator : CG/JU
 Sample : DFTPP086
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP086

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 29 23:34:25 2022



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	66.7	220377	PASS
68	69	0.00	2	1.6	3087	PASS
69	198	0.00	100	58.7	193968	PASS
70	69	0.00	2	0.6	1174	PASS
127	198	10	80	61.7	203973	PASS
197	198	0.00	2	0.6	1926	PASS
198	198	100	100	100.0	330517	PASS
199	198	5	9	6.7	22251	PASS
275	198	10	60	23.8	78696	PASS
365	198	1	100	4.0	13228	PASS
441	198	0.01	100	9.9	32560	PASS
442	198	50	100	60.5	199915	PASS
443	442	15	24	19.3	38659	PASS

m/z	Abundance
35.20	204.0
36.00	280.0
36.90	155.0
39.00	387.0
40.00	3114.0
41.10	526.0
43.00	684.0
44.00	10837.0
45.00	411.0
47.00	295.0
48.00	159.0
49.00	783.0
50.00	237.0
51.10	354.0
55.10	316.0
57.10	386.0
67.10	208.0
69.20	235.0
70.90	229.0
73.10	2196.0
74.10	446.0
75.00	212.0
77.10	438.0
78.10	314.0
81.10	228.0
82.90	210.0
84.00	525.0
86.10	416.0
91.00	178.0
96.00	552.0
102.90	213.0
114.90	184.0
119.00	252.0
133.00	681.0
135.00	201.0
147.00	947.0
149.10	197.0
177.10	317.0
190.90	822.0
191.90	173.0
193.00	563.0
207.00	5567.0
208.00	1142.0
209.00	593.0
209.90	168.0
221.10	611.0
249.10	187.0
266.90	184.0
281.00	1121.0
282.10	357.0
283.00	161.0
340.90	253.0
355.10	366.0
356.00	269.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

m/z	Abundance
36.00	378.0
37.10	1346.0
38.10	4148.0
39.10	23152.0
40.10	3562.0
41.10	1213.0
41.90	201.0
43.10	759.0
44.00	8875.0
45.10	910.0
47.00	245.0
48.10	273.0
49.10	2648.0
50.10	78232.0
51.10	289152.0
52.10	14719.0
53.00	607.0
55.00	2131.0
56.00	7665.0
57.10	17856.0
58.00	789.0
59.10	357.0
59.90	188.0
61.00	3515.0
62.10	3581.0
63.10	11232.0
64.10	1149.0
65.00	4696.0
66.00	329.0
67.00	501.0
68.10	3771.0
69.00	247872.0
70.00	1542.0
70.90	290.0
73.10	3804.0
74.10	28048.0
75.10	39352.0
76.10	14077.0
77.10	275392.0
78.10	17880.0
79.00	18960.0
80.00	15531.0
81.00	20760.0
82.00	4992.0
83.10	4604.0
84.00	604.0
85.00	2976.0
86.10	6017.0
87.10	2521.0
88.00	1214.0
89.00	524.0
91.00	4767.0
92.10	4520.0
93.00	35776.0
94.00	1959.0
95.00	599.0
96.00	1547.0
97.10	665.0
98.00	28728.0
99.00	21272.0
100.00	1847.0
101.00	10902.0
102.00	549.0
103.00	3677.0
104.00	6785.0
105.00	6336.0
106.10	1734.0
107.00	84320.0
108.00	12562.0
109.10	2174.0
110.00	129920.0
111.00	21232.0
112.00	2383.0
112.90	888.0
113.90	209.0
114.20	226.0
115.20	442.0
116.10	4399.0
117.00	88840.0
118.10	5392.0
119.00	989.0
120.00	1278.0
121.10	392.0
122.00	5216.0
123.00	8924.0
124.00	4030.0
125.00	3259.0
126.20	1912.0
127.00	251520.0
128.00	18528.0
129.00	105824.0
130.00	9082.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

131.00	1967.0
132.10	1028.0
133.10	1168.0
134.00	2866.0
135.10	8738.0
136.00	3565.0
137.00	4461.0
138.10	989.0
138.90	568.0
140.10	1301.0
141.00	14090.0
142.00	4594.0
143.00	2975.0
144.00	770.0
145.00	501.0
146.10	2114.0
147.00	7855.0
148.00	19504.0
149.00	3927.0
150.00	869.0
151.00	1883.0
151.80	847.0
153.00	4080.0
154.00	3319.0
155.00	7850.0
156.10	11144.0
157.10	1684.0
158.00	2709.0
159.00	2145.0
160.00	4105.0
161.00	5597.0
162.00	1804.0
163.10	596.0
163.90	493.0
165.00	4646.0
166.10	3984.0
167.00	31112.0
168.00	16800.0
169.00	2517.0
170.10	813.0
171.00	1334.0
172.00	2239.0
173.00	3151.0
174.00	5742.0
175.10	11111.0
176.00	1978.0
177.00	4408.0
178.10	1597.0
179.00	19920.0
180.00	13816.0
181.00	5862.0
182.10	817.0
182.80	320.0
184.10	1627.0
185.00	9743.0
186.00	66664.0
187.10	19672.0
188.10	2218.0
189.00	4510.0
190.00	716.0
191.00	2264.0
192.00	7052.0
193.00	8217.0
194.00	1240.0
195.10	901.0
196.00	11288.0
198.00	393664.0
199.00	27912.0
200.00	2140.0
201.40	1462.0
202.20	292.0
203.00	3630.0
204.10	18736.0
205.00	32864.0
206.10	113896.0
207.00	21000.0
208.00	5664.0
209.10	2251.0
210.00	2326.0
211.10	4763.0
212.00	800.0
213.00	255.0
214.90	1528.0
216.10	3028.0
217.00	36520.0
218.00	4255.0
219.10	446.0
221.10	10924.0
222.00	4229.0
223.00	7587.0
224.10	63456.0
225.10	16324.0
226.10	2243.0
227.00	31912.0
228.10	4306.0
229.00	6409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

230.00	682.0
231.10	1898.0
232.10	395.0
233.00	461.0
234.00	1682.0
235.00	2575.0
236.00	1516.0
237.10	2438.0
238.00	325.0
239.00	1036.0
240.10	1065.0
241.00	1707.0
242.00	3634.0
243.10	3431.0
244.10	46088.0
245.10	6702.0
246.00	11314.0
247.00	2311.0
248.00	368.0
249.00	1526.0
249.90	277.0
251.10	548.0
252.00	579.0
253.10	1643.0
254.10	3285.0
255.00	232896.0
256.00	35512.0
257.00	3034.0
258.00	16912.0
259.00	3060.0
260.10	454.0
261.00	271.0
263.00	185.0
264.00	341.0
265.00	7664.0
266.00	1060.0
267.00	420.0
269.70	317.0
271.00	704.0
272.00	908.0
273.00	7457.0
274.00	17744.0
275.00	88280.0
276.00	11427.0
277.00	9209.0
278.00	2381.0
278.90	400.0
281.00	1117.0
282.00	362.0
283.10	1269.0
284.00	833.0
285.10	1440.0
286.00	341.0
289.00	459.0
290.00	486.0
291.10	193.0
292.00	310.0
293.00	1974.0
294.10	582.0
295.10	610.0
296.00	36104.0
297.00	5130.0
298.00	473.0
301.00	484.0
302.10	601.0
303.10	3919.0
304.10	797.0
308.10	458.0
308.90	305.0
310.00	365.0
313.00	271.0
314.00	1317.0
315.00	3617.0
316.00	1931.0
316.90	357.0
321.00	1066.0
321.90	369.0
323.10	8320.0
324.10	1457.0
325.00	224.0
326.00	161.0
327.00	1526.0
328.00	800.0
329.10	214.0
331.90	422.0
333.00	934.0
334.10	5464.0
335.10	1633.0
336.00	172.0
339.00	250.0
341.20	643.0
342.00	215.0
346.00	1853.0
347.10	381.0
352.10	2715.0
353.00	1460.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

354.10	2232.0
355.10	738.0
356.10	172.0
365.10	13092.0
366.00	1724.0
366.90	185.0
369.90	319.0
371.00	591.0
372.10	3530.0
373.00	914.0
374.10	192.0
383.00	1342.0
384.00	365.0
390.10	509.0
391.10	462.0
392.00	222.0
401.20	218.0
402.10	1439.0
403.00	2336.0
404.00	777.0
405.00	396.0
415.00	169.0
421.10	1598.0
422.00	1641.0
423.10	9891.0
424.10	3155.0
425.00	451.0
435.90	175.0
438.00	174.0
439.20	306.0
439.70	464.0
441.10	30112.0
442.10	184320.0
443.10	37848.0
444.10	3625.0
445.20	202.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

m/z	Abundance
36.00	356.0
37.10	1271.0
38.10	3457.0
39.10	21376.0
40.10	3426.0
41.00	1332.0
42.00	379.0
43.10	750.0
44.10	9915.0
45.00	745.0
47.00	286.0
48.00	249.0
49.00	2421.0
50.10	64384.0
51.10	253888.0
52.10	11838.0
53.00	543.0
55.00	1554.0
56.00	7011.0
57.10	14912.0
58.00	754.0
59.10	362.0
60.00	258.0
61.10	3003.0
62.10	3662.0
63.10	8597.0
64.10	1216.0
65.00	4239.0
66.00	440.0
66.90	473.0
68.00	4029.0
69.00	223744.0
70.00	1303.0
71.10	282.0
73.10	4006.0
74.10	24360.0
75.10	36152.0
76.10	12622.0
77.10	249280.0
78.10	16440.0
79.00	17144.0
80.00	12986.0
81.00	17560.0
82.00	4601.0
83.00	3749.0
84.00	812.0
85.00	2779.0
86.00	5738.0
87.00	2390.0
88.00	876.0
89.00	653.0
91.00	4059.0
92.10	5151.0
93.00	33008.0
94.00	2036.0
95.00	705.0
96.00	1876.0
97.00	674.0
98.00	24856.0
99.00	19408.0
100.00	1601.0
101.00	9604.0
102.00	542.0
103.00	3169.0
104.00	5497.0
105.00	5979.0
106.00	2019.0
107.00	79624.0
108.00	11231.0
109.10	1980.0
110.00	121336.0
111.00	19456.0
112.10	2727.0
113.00	790.0
114.00	234.0
115.10	547.0
116.00	3791.0
117.00	82424.0
118.10	5246.0
119.00	1039.0
120.10	1180.0
121.00	274.0
122.00	4877.0
123.00	8401.0
124.00	3611.0
125.00	3089.0
126.20	1022.0
127.00	237632.0
128.00	17432.0
129.00	96472.0
130.10	7458.0
130.90	1448.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

132.00	789.0
133.00	1219.0
134.00	2950.0
135.00	8804.0
136.00	3438.0
137.00	4221.0
138.00	890.0
139.10	446.0
140.00	1617.0
141.00	13617.0
142.00	4128.0
143.00	2728.0
144.00	663.0
145.00	665.0
146.00	2491.0
147.00	7625.0
148.00	18096.0
149.00	3754.0
150.10	864.0
151.00	1851.0
151.80	746.0
153.00	4133.0
154.00	2590.0
155.00	6669.0
156.00	11002.0
157.10	1477.0
158.00	2489.0
159.00	1763.0
160.10	4444.0
161.10	6160.0
162.00	1953.0
163.00	766.0
164.00	649.0
165.00	5225.0
166.00	4129.0
167.00	31120.0
168.00	16608.0
169.10	2494.0
170.00	834.0
171.00	1268.0
172.00	2256.0
173.00	3261.0
174.00	5482.0
175.10	10021.0
176.00	2492.0
177.00	4822.0
178.00	1442.0
179.00	20952.0
180.00	13359.0
181.00	6269.0
182.10	975.0
182.80	272.0
183.20	358.0
184.10	1564.0
185.00	10496.0
186.10	64432.0
187.00	18704.0
188.00	2013.0
189.00	4395.0
190.00	977.0
191.00	3112.0
192.00	6965.0
193.00	7790.0
194.10	1491.0
195.10	704.0
196.00	11021.0
197.20	5777.0
198.00	393792.0
199.00	26088.0
200.00	2232.0
201.50	1396.0
203.00	4563.0
204.00	19576.0
205.10	31168.0
206.10	117336.0
207.10	20024.0
208.00	5710.0
209.00	2729.0
210.00	2452.0
211.00	4848.0
212.20	722.0
212.90	359.0
214.00	227.0
215.00	1857.0
216.00	3340.0
217.00	37072.0
218.00	5098.0
218.90	358.0
221.10	11998.0
222.00	4837.0
223.00	7273.0
224.10	67424.0
225.10	16552.0
226.00	2210.0
227.00	33832.0
228.00	4906.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

229.00	6321.0
230.10	804.0
231.00	2121.0
232.10	376.0
232.90	406.0
234.00	2063.0
235.00	2408.0
236.00	1867.0
237.10	2691.0
238.00	419.0
239.10	1302.0
240.00	1260.0
241.00	1329.0
242.10	4317.0
243.10	4203.0
244.10	46432.0
245.10	6105.0
246.00	12342.0
247.00	2441.0
248.00	380.0
249.00	1653.0
250.10	353.0
251.10	468.0
251.80	340.0
252.30	451.0
253.10	1381.0
254.10	2906.0
255.00	253120.0
256.00	35784.0
257.00	2699.0
258.00	18448.0
259.00	3214.0
260.00	536.0
261.00	280.0
263.00	419.0
264.00	584.0
265.00	7759.0
266.00	1301.0
267.10	305.0
268.00	161.0
270.10	349.0
271.00	809.0
272.10	924.0
273.00	7453.0
274.00	19264.0
275.00	93992.0
276.00	12555.0
277.00	11029.0
278.00	2234.0
279.10	535.0
281.10	1176.0
282.10	506.0
283.00	1530.0
284.00	819.0
285.00	1876.0
286.00	449.0
289.00	512.0
290.10	296.0
291.00	309.0
292.00	360.0
293.00	2081.0
294.00	505.0
295.10	810.0
296.10	39592.0
297.00	5302.0
298.10	415.0
301.00	372.0
302.00	773.0
303.00	3971.0
304.00	879.0
304.90	159.0
308.20	468.0
309.00	333.0
310.00	437.0
313.00	382.0
314.10	1584.0
315.00	4240.0
316.10	2484.0
317.10	244.0
320.10	158.0
321.00	903.0
322.00	431.0
323.10	9856.0
324.00	1498.0
324.90	198.0
325.80	172.0
327.00	2234.0
328.00	1132.0
329.00	238.0
332.10	828.0
333.10	1294.0
334.10	6570.0
335.10	1391.0
336.10	189.0
339.00	273.0
341.10	1105.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

342.10	229.0
346.00	2263.0
346.90	251.0
350.00	179.0
351.00	227.0
352.10	3150.0
353.00	1889.0
354.10	2915.0
355.10	787.0
357.10	191.0
358.90	250.0
365.00	15789.0
366.00	2487.0
367.00	159.0
369.90	163.0
370.80	479.0
372.10	4459.0
373.10	1186.0
383.00	1396.0
383.90	492.0
390.10	559.0
391.00	270.0
392.10	239.0
401.00	243.0
402.10	1723.0
403.00	2589.0
404.00	689.0
421.10	2841.0
422.00	1850.0
423.10	13342.0
424.10	3016.0
425.10	338.0
437.50	170.0
438.80	488.0
439.60	377.0
441.10	39304.0
442.10	242240.0
443.10	45616.0
444.10	4040.0
445.00	344.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

m/z	Abundance
35.10	182.0
36.00	498.0
37.00	585.0
38.10	1778.0
39.10	9345.0
40.00	2668.0
41.10	899.0
42.00	253.0
43.10	881.0
44.10	9762.0
45.00	469.0
46.90	229.0
47.90	175.0
49.00	1656.0
50.10	31576.0
51.10	119152.0
52.10	5631.0
53.10	449.0
55.10	1104.0
56.00	3450.0
57.00	7812.0
58.00	325.0
59.00	281.0
60.00	171.0
61.00	1216.0
62.00	1496.0
63.10	4856.0
64.00	625.0
65.00	2237.0
66.00	248.0
67.00	248.0
68.10	1460.0
69.00	110992.0
70.00	676.0
71.10	276.0
73.10	2803.0
74.10	11756.0
75.00	18504.0
76.10	6207.0
77.10	122064.0
78.10	8671.0
79.00	8806.0
80.00	5879.0
81.00	8419.0
82.00	2069.0
83.10	1941.0
83.90	615.0
85.00	1630.0
86.00	2711.0
87.10	1616.0
87.90	458.0
88.90	381.0
91.00	2399.0
92.00	2737.0
93.00	16488.0
94.00	1025.0
94.90	397.0
96.10	1189.0
97.10	524.0
98.00	12952.0
99.00	9046.0
100.00	611.0
101.00	5559.0
101.90	198.0
102.90	1656.0
104.00	3510.0
105.00	3201.0
106.10	926.0
107.00	38608.0
108.00	5443.0
109.00	1261.0
110.00	62488.0
111.00	9640.0
112.00	1089.0
113.00	308.0
115.10	437.0
116.00	2146.0
117.00	42152.0
118.00	3271.0
119.10	479.0
120.00	635.0
120.90	231.0
122.00	2357.0
123.00	4177.0
124.00	1748.0
125.00	1491.0
127.00	122768.0
128.00	9027.0
129.00	52608.0
130.00	4049.0
131.00	1035.0
132.00	368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

133.00	727.0
134.00	1417.0
135.00	4256.0
136.00	1603.0
137.00	2249.0
138.00	653.0
139.00	403.0
140.10	636.0
141.00	6556.0
142.00	2066.0
143.00	1163.0
144.00	454.0
145.00	434.0
145.90	1237.0
147.00	4050.0
148.00	10066.0
149.10	2022.0
150.00	370.0
151.10	1068.0
152.00	349.0
153.00	2305.0
154.00	1614.0
155.10	3875.0
156.00	5210.0
157.00	778.0
158.00	1254.0
159.00	1085.0
160.10	2223.0
161.00	3054.0
161.90	945.0
163.00	431.0
164.00	432.0
165.00	2413.0
166.00	2155.0
167.00	14905.0
168.00	8217.0
169.00	1489.0
170.00	481.0
171.00	594.0
172.00	1282.0
173.00	1981.0
174.00	2707.0
175.00	5016.0
176.00	1216.0
177.00	2552.0
178.10	899.0
179.00	10132.0
180.00	6405.0
181.00	2787.0
182.00	471.0
183.10	155.0
183.90	1046.0
185.10	5391.0
186.00	32256.0
187.00	10458.0
188.00	1097.0
189.00	2166.0
190.00	369.0
191.00	1658.0
192.00	2522.0
193.00	4907.0
194.10	887.0
195.00	500.0
196.00	6070.0
198.00	204096.0
199.00	12754.0
200.00	1129.0
201.40	752.0
202.10	317.0
202.90	1820.0
204.10	10019.0
205.10	15972.0
206.10	62016.0
207.10	13703.0
208.00	3842.0
209.00	1729.0
210.00	1045.0
211.00	2657.0
212.00	267.0
212.90	177.0
215.00	878.0
216.00	1929.0
217.00	18904.0
218.00	2484.0
219.10	313.0
220.20	444.0
221.00	7050.0
222.00	2272.0
223.00	4362.0
224.10	37488.0
225.00	8352.0
226.10	1232.0
227.00	18968.0
228.00	2395.0
229.00	3667.0
229.90	414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

231.10	840.0
232.00	314.0
233.00	418.0
234.10	1280.0
235.00	1039.0
236.00	959.0
237.00	1364.0
238.00	173.0
239.00	716.0
239.90	571.0
241.00	1093.0
242.00	2128.0
243.10	2293.0
244.10	27168.0
245.00	3333.0
246.00	6980.0
247.00	1475.0
248.10	246.0
249.00	1063.0
249.80	259.0
251.00	481.0
252.00	567.0
253.10	1311.0
254.10	2135.0
255.00	140096.0
256.00	19904.0
257.00	1619.0
258.00	10635.0
259.00	1453.0
260.00	327.0
264.00	342.0
265.00	4706.0
266.00	553.0
267.00	277.0
270.00	191.0
271.00	366.0
272.10	543.0
273.00	4770.0
274.00	11511.0
275.00	53816.0
276.00	7441.0
277.00	5718.0
278.00	993.0
278.80	190.0
281.00	1177.0
282.00	484.0
283.10	911.0
284.10	669.0
285.10	1130.0
288.90	322.0
289.80	159.0
290.80	249.0
292.00	216.0
293.00	1268.0
293.90	310.0
295.10	615.0
296.00	23376.0
297.00	2771.0
298.00	249.0
301.10	385.0
302.10	388.0
303.10	2441.0
304.10	732.0
307.90	307.0
310.10	298.0
312.80	168.0
314.00	745.0
315.00	2682.0
316.00	1370.0
317.00	259.0
321.00	674.0
322.10	422.0
323.00	5070.0
324.00	1121.0
327.00	1294.0
328.00	600.0
332.10	510.0
333.10	709.0
334.10	4601.0
335.00	942.0
336.00	184.0
340.00	172.0
341.10	685.0
342.10	251.0
346.00	1739.0
347.00	328.0
352.10	1798.0
353.10	1206.0
354.10	1684.0
355.10	763.0
365.00	10804.0
366.10	1081.0
370.10	193.0
370.90	478.0
372.10	2548.0
373.10	717.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086

383.00	905.0
384.00	262.0
390.00	378.0
390.90	200.0
401.00	323.0
402.00	1337.0
403.10	2094.0
404.10	789.0
405.00	171.0
414.90	162.0
421.00	1238.0
422.10	1308.0
423.10	8544.0
424.10	2200.0
425.00	321.0
436.50	175.0
438.60	161.0
439.10	176.0
439.50	210.0
440.10	195.0
441.10	28264.0
442.10	173184.0
443.10	32512.0
444.10	3239.0
445.10	401.0

Instrument :
BNA_M
ClientSampleId :
DFTPP086