

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
 Data File : BM003321.D
 Acq On : 30 Nov 2015 12:56
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 01 13:31:12 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 01 13:09:54 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.488	0.476	2.5	103	-0.02
3	n-Nitrosodimethylamine	0.462	0.465	-0.6	103	0.00
4 S	2-Fluorophenol	0.986	0.986	0.0	105	0.00
5 S	Phenol-d6	1.131	1.131	0.0	106	0.00
6	bis(2-Chloroethyl)ether	1.156	1.160	-0.3	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.222	0.222	0.0	108	0.00
9	Nitrobenzene	0.243	0.239	1.6	108	0.00
10	Naphthalene	1.081	1.090	-0.8	106	0.00
11	Hexachlorobutadiene	0.164	0.169	-3.0	110	-0.02
12	2-Methylnaphthalene	0.678	0.691	-1.9	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.090	0.097	-7.8	111	0.02
15 S	2-Fluorobiphenyl	1.614	1.582	2.0	107	0.00
16	Acenaphthylene	1.832	1.863	-1.7	107	0.00
17	Acenaphthene	1.414	1.430	-1.1	118	0.00
18	Fluorene	1.427	1.506	-5.5	111	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.170	0.170	0.0	129	0.00
21	Hexachlorobenzene	0.176	0.166	5.7	99	0.02
22	Pentachlorophenol	0.066	0.073	-10.6	154#	0.00
23	Phenanthrene	1.145	1.151	-0.5	120	0.02
24	Anthracene	0.933	0.936	-0.3	96	0.00
25	Fluoranthene	1.090	1.145	-5.0	112	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	90	0.01
27	Benzydine	0.229	0.254	-10.9	141	0.01
28	Pvrene	1.528	1.565	-2.4	110	0.00
29 S	Terphenyl-d14	0.697	0.709	-1.7	101	0.00
30	Benzo(a)anthracene	1.266	1.274	-0.6	102	0.01
31	3,3'-Dichlorobenzidine	0.270	0.269	0.4	132	0.01
32	Chrysene	1.515	1.601	-5.7	115	0.00
33	Bis(2-ethylhexyl)phthalate	0.518	0.533	-2.9	154#	0.00
34	Indeno(1,2,3-cd)pyrene	1.160	1.220	-5.2	109	0.00
35 I	Perylene-d12	1.000	1.000	0.0	104	0.00
36	Benzo(b)fluoranthene	1.439	1.464	-1.7	103	0.00
37	Benzo(k)fluoranthene	1.294	1.321	-2.1	112	0.00
38 C	Benzo(a)pyrene	1.146	1.138	0.7	108	0.00
39	Dibenzo(a,h)anthracene	0.992	1.038	-4.6	110	0.00
40	Benzo(a,h,i)perylene	1.190	1.200	-0.8	108	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
Data File : BM003321.D
Acq On : 30 Nov 2015 12:56
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 01 13:31:12 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 01 13:09:54 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				