

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
 Data File : BM003321.D
 Acq On : 30 Nov 2015 12:56
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 01 13:31:12 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 01 13:09:54 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	1.051	-5.1	103	-0.02
3	n-Nitrosodimethylamine	1.000	1.007	-0.7	103	0.00
4 S	2-Fluorophenol	1.000	1.000	0.0	105	0.00
5 S	Phenol-d6	1.000	0.999	0.1	106	0.00
6	bis(2-Chloroethyl)ether	1.000	1.004	-0.4	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	103	0.00
8 S	Nitrobenzene-d5	1.000	0.999	0.1	108	0.00
9	Nitrobenzene	1.000	0.982	1.8	108	0.00
10	Naphthalene	1.000	1.009	-0.9	106	0.00
11	Hexachlorobutadiene	1.000	1.031	-3.1	110	-0.02
12	2-Methylnaphthalene	1.000	1.020	-2.0	107	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	1.000	1.035	-3.5	111	0.02
15 S	2-Fluorobiphenyl	1.000	0.980	2.0	107	0.00
16	Acenaphthylene	1.000	1.017	-1.7	107	0.00
17	Acenaphthene	1.000	1.011	-1.1	118	0.00
18	Fluorene	1.000	1.055	-5.5	111	0.02
19 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	1.000	1.072	-7.2	129	0.00
21	Hexachlorobenzene	1.000	0.943	5.7	99	0.02
22	Pentachlorophenol	1.000	1.302	-30.2#	154	0.00
23	Phenanthrene	1.000	1.084	-8.4	120	0.02
24	Anthracene	1.000	1.003	-0.3	96	0.00
25	Fluoranthene	1.000	1.050	-5.0	112	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	90	0.01
27	Benzydine	1.000	1.260	-26.0#	141	0.01
28	Pvrene	1.000	1.024	-2.4	110	0.00
29 S	Terphenyl-d14	1.000	1.017	-1.7	101	0.00
30	Benzo(a)anthracene	1.000	1.006	-0.6	102	0.01
31	3,3'-Dichlorobenzidine	1.000	1.109	-10.9	132	0.01
32	Chrysene	1.000	1.136	-13.6	115	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.335	-33.5#	154	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	1.051	-5.1	109	0.00
35 I	Perylene-d12	5.000	5.000	0.0	104	0.00
36	Benzo(b)fluoranthene	1.000	1.018	-1.8	103	0.00
37	Benzo(k)fluoranthene	1.000	1.021	-2.1	112	0.00
38 C	Benzo(a)pyrene	1.000	0.993	0.7	108	0.00
39	Dibenzo(a,h)anthracene	1.000	1.047	-4.7	110	0.00
40	Benzo(a,h,i)perylene	1.000	1.008	-0.8	108	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
Data File : BM003321.D
Acq On : 30 Nov 2015 12:56
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 01 13:31:12 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 01 13:09:54 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				