

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
 Data File : BM003326.D
 Acq On : 30 Nov 2015 19:35
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 01 02:48:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	-0.02
2	1,4-Dioxane	0.488	0.452	7.4	126	-0.02
3	n-Nitrosodimethylamine	0.462	0.500	-8.2	143	-0.02
4 S	2-Fluorophenol	0.986	1.040	-5.5	143	0.00
5 S	Phenol-d6	1.131	1.307	-15.6	157#	-0.02
6	bis(2-Chloroethyl)ether	1.156	1.228	-6.2	144	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	142	-0.02
8 S	Nitrobenzene-d5	0.222	0.259	-16.7	173#	-0.01
9	Nitrobenzene	0.243	0.278	-14.4	172#	-0.01
10	Naphthalene	1.081	1.086	-0.5	145	0.00
11	Hexachlorobutadiene	0.164	0.176	-7.3	157#	-0.02
12	2-Methylnaphthalene	0.678	0.746	-10.0	158#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	160#	0.00
14 S	2,4,6-Tribromophenol	0.090	0.120	-33.3#	218#	0.00
15 S	2-Fluorobiphenyl	1.614	1.517	6.0	163#	0.00
16	Acenaphthylene	1.832	2.100	-14.6	191#	0.00
17	Acenaphthene	1.414	1.281	9.4	168#	0.00
18	Fluorene	1.427	1.467	-2.8	172#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	154#	0.00
20	4-Bromophenyl-phenylether	0.170	0.141	17.1	173#	0.00
21	Hexachlorobenzene	0.176	0.171	2.8	164#	0.00
22	Pentachlorophenol	0.066	0.062	6.1	213#	0.00
23	Phenanthrene	1.145	0.895	21.8	151#	0.00
24	Anthracene	0.933	1.089	-16.7	180#	0.00
25	Fluoranthene	1.090	0.992	9.0	157#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	110	-0.01
27	Benzydine	0.229	0.457	-99.6#	309#	0.00
28	Pvrene	1.528	1.788	-17.0	154#	0.00
29 S	Terphenyl-d14	0.697	0.792	-13.6	138	0.00
30	Benzo(a)anthracene	1.266	1.374	-8.5	135	-0.01
31	3,3'-Dichlorobenzidine	0.270	0.392	-45.2#	236#	0.01
32	Chrysene	1.515	1.586	-4.7	139	0.00
33	Bis(2-ethylhexyl)phthalate	0.518	0.806	-55.6#	284#	0.00
34	Indeno(1,2,3-cd)pyrene	1.160	1.304	-12.4	143	0.00
35 I	Perylene-d12	1.000	1.000	0.0	123	0.00
36	Benzo(b)fluoranthene	1.439	1.440	-0.1	120	0.00
37	Benzo(k)fluoranthene	1.294	1.442	-11.4	144	-0.01
38 C	Benzo(a)pyrene	1.146	1.274	-11.2	143	0.00
39	Dibenzo(a,h)anthracene	0.992	1.199	-20.9	150#	-0.01
40	Benzo(a,h,i)perylene	1.190	1.254	-5.4	134	-0.01

Data Path : Z:\HPCHEM1\BNA M\Data\BM113015\
Data File : BM003326.D
Acq On : 30 Nov 2015 19:35
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 01 02:48:43 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				