

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113015\
 Data File : BM003328.D
 Acq On : 30 Nov 2015 20:47
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 01 03:45:05 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	-0.02
2	1,4-Dioxane	0.488	0.454	7.0	126	-0.02
3	n-Nitrosodimethylamine	0.462	0.495	-7.1	141	-0.02
4 S	2-Fluorophenol	0.986	1.022	-3.7	140	0.00
5 S	Phenol-d6	1.131	1.276	-12.8	153#	-0.02
6	bis(2-Chloroethyl)ether	1.156	1.241	-7.4	145	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	142	-0.02
8 S	Nitrobenzene-d5	0.222	0.253	-14.0	170#	-0.01
9	Nitrobenzene	0.243	0.273	-12.3	169#	-0.01
10	Naphthalene	1.081	1.091	-0.9	145	0.00
11	Hexachlorobutadiene	0.164	0.166	-1.2	148	-0.02
12	2-Methylnaphthalene	0.678	0.752	-10.9	159#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.090	0.139	-54.4#	235#	0.00
15 S	2-Fluorobiphenyl	1.614	1.650	-2.2	165#	-0.01
16	Acenaphthylene	1.832	2.072	-13.1	175#	0.00
17	Acenaphthene	1.414	1.440	-1.8	176#	-0.02
18	Fluorene	1.427	1.808	-26.7#	198#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
20	4-Bromophenyl-phenylether	0.170	0.198	-16.5	221#	-0.03
21	Hexachlorobenzene	0.176	0.248	-40.9#	216#	0.00
22	Pentachlorophenol	0.066	0.067	-1.5	209#	-0.03
23	Phenanthrene	1.145	1.253	-9.4	192#	0.00
24	Anthracene	0.933	1.099	-17.8	165#	0.00
25	Fluoranthene	1.090	1.094	-0.4	158#	-0.02
26 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
27	Benzydine	0.229	0.414	-80.8#	316#	0.00
28	Pvrene	1.528	1.539	-0.7	149	-0.02
29 S	Terphenyl-d14	0.697	0.688	1.3	134	0.00
30	Benzo(a)anthracene	1.266	1.307	-3.2	144	0.00
31	3,3'-Dichlorobenzidine	0.270	0.364	-34.8#	247#	0.00
32	Chrysene	1.515	1.218	19.6	120	-0.02
33	Bis(2-ethylhexyl)phthalate	0.518	0.411	20.7	163#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.160	1.095	5.6	135	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	116	0.00
36	Benzo(b)fluoranthene	1.439	1.444	-0.3	114	0.00
37	Benzo(k)fluoranthene	1.294	1.424	-10.0	134	-0.01
38 C	Benzo(a)pyrene	1.146	1.242	-8.4	132	0.00
39	Dibenzo(a,h)anthracene	0.992	1.202	-21.2	142	-0.01
40	Benzo(a,h,i)perylene	1.190	1.253	-5.3	126	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113015\
Data File : BM003328.D
Acq On : 30 Nov 2015 20:47
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 01 03:45:05 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0