

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113015\
 Data File : BM003335.D
 Acq On : 01 Dec 2015 00:57
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 01 02:52:46 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 27 16:36:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	-0.02
2	1,4-Dioxane	0.488	0.457	6.4	130	-0.02
3	n-Nitrosodimethylamine	0.462	0.492	-6.5	143	-0.02
4 S	2-Fluorophenol	0.986	0.991	-0.5	139	0.00
5 S	Phenol-d6	1.131	1.229	-8.7	151#	-0.02
6	bis(2-Chloroethyl)ether	1.156	1.246	-7.8	148	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	145	-0.02
8 S	Nitrobenzene-d5	0.222	0.238	-7.2	164#	-0.01
9	Nitrobenzene	0.243	0.260	-7.0	165#	-0.01
10	Naphthalene	1.081	1.092	-1.0	149	0.00
11	Hexachlorobutadiene	0.164	0.174	-6.1	158#	-0.02
12	2-Methylnaphthalene	0.678	0.744	-9.7	162#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	166#	0.00
14 S	2,4,6-Tribromophenol	0.090	0.111	-23.3	210#	0.00
15 S	2-Fluorobiphenyl	1.614	1.500	7.1	167#	0.00
16	Acenaphthylene	1.832	2.010	-9.7	190#	0.00
17	Acenaphthene	1.414	1.292	8.6	176#	0.00
18	Fluorene	1.427	1.455	-2.0	177#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	162#	0.00
20	4-Bromophenyl-phenylether	0.170	0.139	18.2	179#	-0.03
21	Hexachlorobenzene	0.176	0.178	-1.1	180#	0.00
22	Pentachlorophenol	0.066	0.053	19.7	189#	0.00
23	Phenanthrene	1.145	0.909	20.6	161#	0.00
24	Anthracene	0.933	1.034	-10.8	179#	0.00
25	Fluoranthene	1.090	0.910	16.5	151#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	114	-0.01
27	Benzydine	0.229	0.394	-72.1#	275#	0.00
28	Pvrene	1.528	1.706	-11.6	151#	0.00
29 S	Terphenyl-d14	0.697	0.802	-15.1	144	0.00
30	Benzo(a)anthracene	1.266	1.470	-16.1	149	-0.01
31	3,3'-Dichlorobenzidine	0.270	0.295	-9.3	183#	0.00
32	Chrysene	1.515	1.281	15.4	116	0.00
33	Bis(2-ethylhexyl)phthalate	0.518	0.532	-2.7	193#	0.00
34	Indeno(1,2,3-cd)pyrene	1.160	1.175	-1.3	133	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	116	-0.01
36	Benzo(b)fluoranthene	1.439	1.468	-2.0	116	0.00
37	Benzo(k)fluoranthene	1.294	1.369	-5.8	129	-0.01
38 C	Benzo(a)pyrene	1.146	1.197	-4.5	127	-0.01
39	Dibenzo(a,h)anthracene	0.992	1.192	-20.2	141	-0.01
40	Benzo(a,h,i)perylene	1.190	1.246	-4.7	125	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113015\
Data File : BM003335.D
Acq On : 01 Dec 2015 00:57
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 01 02:52:46 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM112715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 27 16:36:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				