

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113016\
 Data File : BM008127.D
 Acq On : 30 Nov 2016 23:56
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02043

Quant Time: Dec 01 10:36:37 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 02:52:19 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.465	0.472	-1.5	99	0.00
3 S	1,4-Dioxane-d8	0.405	0.416	-2.7	100	0.00
4	Benzaldehyde	1.079	1.103	-2.2	103	0.00
5 S	Phenol-d5	1.491	1.458	2.2	102	0.00
6	Phenol	1.531	1.547	-1.0	104	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.857	0.868	-1.3	102	0.00
8	Bis(2-Chloroethyl)ether	1.162	1.172	-0.9	104	0.00
9 S	2-Chlorophenol-d4	1.154	1.151	0.3	102	0.00
10	2-Chlorophenol	1.182	1.197	-1.3	104	0.00
11	2-Methylphenol	1.135	1.108	2.4	103	0.00
12	2,2'-oxybis(1-Chloropropane	1.291	1.318	-2.1	104	0.00
13 S	4-Methylphenol-d8	1.159	1.118	3.5	101	0.00
14	Acetophenone	1.854	1.847	0.4	104	0.00
15 P	N-Nitroso-di-n-propylamine	0.973	0.965	0.8	105	0.00
16	4-Methylphenol	1.223	1.217	0.5	104	0.00
17	Hexachloroethane	0.527	0.520	1.3	98	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
19 S	Nitrobenzene-d5	0.144	0.143	0.7	103	0.00
20	Nitrobenzene	0.357	0.358	-0.3	104	0.00
21	Isophorone	0.658	0.642	2.4	102	0.00
22 S	2-Nitrophenol-d4	0.163	0.152	6.7	97	0.00
23 C	2-Nitrophenol	0.167	0.161	3.6	100	0.00
24	2,4-Dimethylphenol	0.358	0.355	0.8	103	0.00
25	Bis(2-Chloroethoxy)methane	0.382	0.379	0.8	104	0.00
26 S	2,4-Dichlorophenol-d3	0.298	0.291	2.3	102	0.00
27 C	2,4-Dichlorophenol	0.291	0.289	0.7	103	0.00
28	Naphthalene	0.929	0.931	-0.2	104	0.00
29 S	4-Chloroaniline-d4	0.290	0.283	2.4	105	0.00
30	4-Chloroaniline	0.298	0.302	-1.3	108	0.00
31 C	Hexachlorobutadiene	0.231	0.233	-0.9	103	0.00
32	Caprolactam	0.088	0.073	17.0	88	0.00
33 C	4-Chloro-3-methylphenol	0.313	0.293	6.4	98	0.00
34	2-Methylnaphthalene	0.665	0.654	1.7	103	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
36	1,2,4,5-Tetrachlorobenzene	0.621	0.643	-3.5	103	0.00
37	Hexachlorocyclopentadiene	0.308	0.234	24.0#	90	0.00
38 C	2,4,6-Trichlorophenol	0.372	0.363	2.4	99	0.00
39	2,4,5-Trichlorophenol	0.395	0.375	5.1	95	0.00
40	1,1'-Biphenyl	1.348	1.371	-1.7	103	0.00
41	2-Chloronaphthalene	1.030	1.046	-1.6	101	0.00
42	2-Nitroaniline	0.294	0.286	2.7	95	0.00
43 S	Dimethylphthalate-d6	1.289	1.262	2.1	98	0.00
44	Dimethylphthalate	1.264	1.242	1.7	97	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113016\
 Data File : BM008127.D
 Acq On : 30 Nov 2016 23:56
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02043

Quant Time: Dec 01 10:36:37 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 02:52:19 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.251	0.245	2.4	96	0.00
46 S	Acenaphthylene-d8	1.584	1.572	0.8	99	0.00
47	Acenaphthylene	1.598	1.609	-0.7	100	0.00
48	3-Nitroaniline	0.235	0.213	9.4	90	0.00
49 C	Acenaphthene	1.092	1.079	1.2	99	0.00
50	2,4-Dinitrophenol	0.158	0.104	34.2#	74	0.00
51 S	4-Nitrophenol-d4	0.199	0.162	18.6	86	0.00
52	4-Nitrophenol	0.187	0.149	20.3#	81	0.00
53	Dibenzofuran	1.553	1.542	0.7	100	0.00
54	2,4-Dinitrotoluene	0.357	0.337	5.6	93	0.00
55	2,3,4,6-Tetrachlorophenol	0.354	0.333	5.9	94	0.00
56	Diethylphthalate	1.353	1.190	12.0	90	0.00
57 S	Fluorene-d10	1.112	1.090	2.0	97	0.00
58	Fluorene	1.261	1.239	1.7	98	0.00
59	4-Chlorophenyl-phenylether	0.667	0.663	0.6	99	0.00
60	4-Nitroaniline	0.236	0.206	12.7	91	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.106	10.9	91	0.00
63	4,6-Dinitro-2-methylphenol	0.122	0.106	13.1	86	0.00
64	N-Nitrosodiphenylamine	0.558	0.563	-0.9	97	0.00
65	4-Bromophenyl-phenylether	0.223	0.222	0.4	96	0.00
66	Hexachlorobenzene	0.245	0.245	0.0	95	0.00
67	Atrazine	0.219	0.206	5.9	91	0.00
68 C	Pentachlorophenol	0.142	0.122	14.1	86	0.00
69	Phenanthrene	1.000	0.999	0.1	94	0.00
70 S	Anthracene-d10	0.860	0.858	0.2	94	0.00
71	Anthracene	1.021	1.024	-0.3	94	0.00
72	Carbazole	0.886	0.855	3.5	91	0.00
73	Di-n-butylphthalate	1.051	0.986	6.2	87	0.00
74 C	Fluoranthene	1.161	1.136	2.2	90	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
76 S	Pyrene-d10	0.809	0.798	1.4	91	0.00
77	Pyrene	1.029	1.005	2.3	91	0.00
78	Butylbenzylphthalate	0.396	0.371	6.3	84	0.00
79	3,3'-Dichlorobenzidine	0.353	0.353	0.0	87	0.00
80	Benzo(a)anthracene	1.032	1.020	1.2	88	0.00
81	Bis(2-ethylhexyl)phthalate	0.597	0.551	7.7	84	0.00
82	Chrysene	0.975	0.979	-0.4	89	0.00
83 I	Perylene-d12	1.000	1.000	0.0	87	0.00
84	Di-n-octyl phthalate	1.030	0.970	5.8	84	0.00
85	Benzo(b)fluoranthene	1.071	1.052	1.8	86	0.00
86	Benzo(k)fluoranthene	1.010	1.021	-1.1	89	0.00
87 S	Benzo(a)pyrene-d12	0.833	0.822	1.3	86	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM113016\
Data File : BM008127.D
Acq On : 30 Nov 2016 23:56
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02043

Quant Time: Dec 01 10:36:37 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 01 02:52:19 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.029	1.035	-0.6	88	0.00
89	Indeno(1,2,3-cd)pyrene	1.255	1.240	1.2	86	0.00
90	Dibenzo(a,h)anthracene	1.058	1.051	0.7	86	0.00
91	Benzo(a,h,i)perylene	1.044	1.040	0.4	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0