

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113023\
 Data File : BM043110.D
 Acq On : 01 Dec 2023 06:57
 Operator : MA/JU
 Sample : 05459-11
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AJ6

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/01/2023
 Supervised By :mohammad ahmed 12/04/2023

Quant Time: Dec 01 07:30:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 00:42:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.800	152	1235m	0.400	ng/ul	-0.06
4) Naphthalene-d8	10.600	136	3532	0.400	ng/ul	#-0.03
9) Acenaphthene-d10	14.450	164	2117m	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.203	188	4236m	0.400	ng/ul	-0.03
17) Chrysene-d12	21.406	240	2811m	0.400	ng/ul	0.00
23) Perylene-d12	23.744	264	2630m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.155	96	6272	3.310	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.205	152	1934	0.409	ng/ul	-0.03
18) Fluoranthene-d10	19.233	212	4600	0.449	ng/ul	-0.02
Target Compounds						
					Qvalue	
5) Naphthalene	10.649	128	1134	0.118	ng/ul	98
7) 2-Methylnaphthalene	12.288	142	936	0.165	ng/ul	100
8) 1-Methylnaphthalene	12.497	142	826	0.131	ng/ul	95
15) Phenanthrene	17.245	178	1495	0.124	ng/ul	95
19) Fluoranthene	19.266	202	2570	0.161	ng/ul	97
20) Pyrene	19.628	202	2413	0.142	ng/ul#	87
21) Benzo(a)anthracene	21.391	228	1063	0.117	ng/ul#	85
22) Chrysene	21.444	228	930	0.055	ng/ul#	89
24) Benzo(b)fluoranthene	23.025	252	1270	0.143	ng/ul	67
25) Benzo(k)fluoranthene	23.072	252	658m	0.054	ng/ul	
27) Indeno(1,2,3-cd)pyrene	26.229	276	760	0.058	ng/ul#	99
29) Benzo(g,h,i)perylene	26.990	276	834	0.065	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113023\
Data File : BM043110.D
Acq On : 01 Dec 2023 06:57
Operator : MA/JU
Sample : 05459-11
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AJ6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/01/2023
Supervised By :mohammad ahmed 12/04/2023

Quant Time: Dec 01 07:30:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 00:42:42 2023
Response via : Initial Calibration

