

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120116\
 Data File : BM008129.D
 Acq On : 01 Dec 2016 09:27
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02044

Quant Time: Dec 02 02:30:48 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 02:52:19 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.465	0.459	1.3	103	0.00
3 S	1,4-Dioxane-d8	0.405	0.392	3.2	100	0.00
4	Benzaldehyde	1.079	1.103	-2.2	109	0.00
5 S	Phenol-d5	1.491	1.471	1.3	109	0.00
6	Phenol	1.531	1.538	-0.5	110	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.857	0.873	-1.9	109	0.00
8	Bis(2-Chloroethyl)ether	1.162	1.183	-1.8	111	0.00
9 S	2-Chlorophenol-d4	1.154	1.141	1.1	107	0.00
10	2-Chlorophenol	1.182	1.167	1.3	107	0.00
11	2-Methylphenol	1.135	1.114	1.9	110	0.00
12	2,2'-oxybis(1-Chloropropane	1.291	1.371	-6.2	114	0.00
13 S	4-Methylphenol-d8	1.159	1.171	-1.0	112	0.00
14	Acetophenone	1.854	1.889	-1.9	112	0.00
15 P	N-Nitroso-di-n-propylamine	0.973	1.012	-4.0	116	0.00
16	4-Methylphenol	1.223	1.223	0.0	111	0.00
17	Hexachloroethane	0.527	0.519	1.5	104	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
19 S	Nitrobenzene-d5	0.144	0.139	3.5	110	0.00
20	Nitrobenzene	0.357	0.369	-3.4	118	0.00
21	Isophorone	0.658	0.664	-0.9	116	0.00
22 S	2-Nitrophenol-d4	0.163	0.151	7.4	106	0.00
23 C	2-Nitrophenol	0.167	0.164	1.8	112	0.00
24	2,4-Dimethylphenol	0.358	0.360	-0.6	115	0.00
25	Bis(2-Chloroethoxy)methane	0.382	0.385	-0.8	116	0.00
26 S	2,4-Dichlorophenol-d3	0.298	0.293	1.7	113	0.00
27 C	2,4-Dichlorophenol	0.291	0.288	1.0	113	0.00
28	Naphthalene	0.929	0.912	1.8	112	0.00
29 S	4-Chloroaniline-d4	0.290	0.302	-4.1	123	0.00
30	4-Chloroaniline	0.298	0.306	-2.7	121	0.00
31 C	Hexachlorobutadiene	0.231	0.231	0.0	111	0.00
32	Caprolactam	0.088	0.079	10.2	105	0.00
33 C	4-Chloro-3-methylphenol	0.313	0.316	-1.0	116	0.00
34	2-Methylnaphthalene	0.665	0.666	-0.2	115	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
36	1,2,4,5-Tetrachlorobenzene	0.621	0.626	-0.8	117	0.00
37	Hexachlorocyclopentadiene	0.308	0.231	25.0#	104	0.00
38 C	2,4,6-Trichlorophenol	0.372	0.359	3.5	114	0.00
39	2,4,5-Trichlorophenol	0.395	0.366	7.3	108	0.00
40	1,1'-Biphenyl	1.348	1.348	0.0	118	0.00
41	2-Chloronaphthalene	1.030	1.032	-0.2	117	0.00
42	2-Nitroaniline	0.294	0.291	1.0	113	0.00
43 S	Dimethylphthalate-d6	1.289	1.285	0.3	117	0.00
44	Dimethylphthalate	1.264	1.260	0.3	115	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120116\
 Data File : BM008129.D
 Acq On : 01 Dec 2016 09:27
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02044

Quant Time: Dec 02 02:30:48 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 01 02:52:19 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.251	0.251	0.0	115	0.00
46 S	Acenaphthylene-d8	1.584	1.574	0.6	115	0.00
47	Acenaphthylene	1.598	1.600	-0.1	116	0.00
48	3-Nitroaniline	0.235	0.224	4.7	111	0.00
49 C	Acenaphthene	1.092	1.091	0.1	117	0.00
50	2,4-Dinitrophenol	0.158	0.113	28.5#	94	0.00
51 S	4-Nitrophenol-d4	0.199	0.164	17.6	101	0.00
52	4-Nitrophenol	0.187	0.158	15.5	100	0.00
53	Dibenzofuran	1.553	1.562	-0.6	118	0.00
54	2,4-Dinitrotoluene	0.357	0.354	0.8	114	0.00
55	2,3,4,6-Tetrachlorophenol	0.354	0.348	1.7	114	0.00
56	Diethylphthalate	1.353	1.224	9.5	109	0.00
57 S	Fluorene-d10	1.112	1.106	0.5	115	0.00
58	Fluorene	1.261	1.258	0.2	117	0.00
59	4-Chlorophenyl-phenylether	0.667	0.682	-2.2	119	0.00
60	4-Nitroaniline	0.236	0.193	18.2	100	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.105	11.8	106	0.00
63	4,6-Dinitro-2-methylphenol	0.122	0.109	10.7	103	0.00
64	N-Nitrosodiphenylamine	0.558	0.572	-2.5	116	0.00
65	4-Bromophenyl-phenylether	0.223	0.225	-0.9	114	0.00
66	Hexachlorobenzene	0.245	0.246	-0.4	112	0.00
67	Atrazine	0.219	0.209	4.6	109	0.00
68 C	Pentachlorophenol	0.142	0.123	13.4	102	0.00
69	Phenanthrene	1.000	1.007	-0.7	111	0.00
70 S	Anthracene-d10	0.860	0.854	0.7	110	0.00
71	Anthracene	1.021	1.026	-0.5	111	0.00
72	Carbazole	0.886	0.843	4.9	105	0.00
73	Di-n-butylphthalate	1.051	0.964	8.3	100	0.00
74 C	Fluoranthene	1.161	1.098	5.4	102	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
76 S	Pyrene-d10	0.809	0.845	-4.4	101	0.00
77	Pyrene	1.029	1.065	-3.5	101	0.00
78	Butylbenzylphthalate	0.396	0.377	4.8	90	0.00
79	3,3'-Dichlorobenzidine	0.353	0.347	1.7	90	0.00
80	Benzo(a)anthracene	1.032	1.029	0.3	93	0.00
81	Bis(2-ethylhexyl)phthalate	0.597	0.542	9.2	87	0.00
82	Chrysene	0.975	0.979	-0.4	93	0.00
83 I	Perylene-d12	1.000	1.000	0.0	88	0.00
84	Di-n-octyl phthalate	1.030	0.973	5.5	84	0.00
85	Benzo(b)fluoranthene	1.071	1.059	1.1	86	0.00
86	Benzo(k)fluoranthene	1.010	1.039	-2.9	91	0.00
87 S	Benzo(a)pyrene-d12	0.833	0.823	1.2	87	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM120116\
Data File : BM008129.D
Acq On : 01 Dec 2016 09:27
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02044

Quant Time: Dec 02 02:30:48 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM112916.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 01 02:52:19 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.029	1.028	0.1	88	0.00
89	Indeno(1,2,3-cd)pyrene	1.255	1.257	-0.2	87	0.00
90	Dibenzo(a,h)anthracene	1.058	1.057	0.1	87	0.00
91	Benzo(a,h,i)perylene	1.044	1.034	1.0	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0