


Data Path : Z:\HPCHEM1\BNA M\DATA\BM120116\
 Data File : BM008128.D
 Acq On : 01 Dec 2016 08:51
 Operator : UM/SJ
 Sample : DFTPP59
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP59

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM112916.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 02 03:12:28 2016



AutoFind: Scans 2538, 2539, 2540; Background Corrected with Scan 2533

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.7	52744	PASS
68	69	0.00	2	2.0	1186	PASS
69	198	0.00	100	35.8	59634	PASS
70	69	0.00	2	0.4	217	PASS
127	198	10	80	41.9	69765	PASS
197	198	0.00	2	0.6	1029	PASS
198	198	100	100	100.0	166421	PASS
199	198	5	9	6.9	11521	PASS
275	198	10	60	30.0	49853	PASS
365	198	1	100	4.3	7153	PASS
441	443	0.01	100	80.3	21989	PASS
442	198	50	100	83.3	138674	PASS
443	442	15	24	19.8	27400	PASS

Scan 2533 (17.486 min): BM008128.D
DFTPP59

m/z	Abundance
40.00	436.0
44.00	567.0
73.10	613.0
207.00	1998.0
208.10	591.0
280.90	531.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
38.10	618.0
39.10	3581.0
40.10	404.0
44.00	859.0
49.30	302.0
50.10	13929.0
51.10	52016.0
52.00	3094.0
56.00	1897.0
57.00	3363.0
61.00	651.0
62.00	782.0
63.00	2275.0
64.10	316.0
65.10	1134.0
68.10	1145.0
69.00	56888.0
70.00	303.0
73.00	945.0
74.10	6088.0
75.00	9336.0
76.00	3146.0
77.10	65472.0
78.10	4610.0
79.00	4958.0
80.00	3387.0
81.00	5354.0
82.00	1318.0
83.00	1125.0
85.10	937.0
86.00	1421.0
87.10	635.0
91.00	1190.0
92.10	1215.0
93.00	8994.0
94.00	621.0
96.00	365.0
98.10	6459.0
99.00	6211.0
100.00	378.0
101.00	2698.0
103.00	1014.0
104.00	1617.0
105.00	1626.0
105.90	654.0
107.00	21408.0
108.00	3035.0
109.20	502.0
110.00	35432.0
111.00	5492.0
112.00	835.0
112.90	301.0
116.00	939.0
117.00	19560.0
118.00	1411.0
122.00	1168.0
122.90	2118.0
124.00	801.0
125.00	755.0
126.20	478.0
127.00	62960.0
128.10	4857.0
129.00	25952.0
130.00	2714.0
131.10	713.0
133.00	503.0
134.10	873.0
135.10	2234.0
136.00	1140.0
136.90	1453.0
140.20	364.0
141.00	3446.0
142.00	1098.0
143.00	903.0
146.00	530.0
147.00	2081.0
148.00	5311.0
149.00	1025.0
151.10	363.0
153.00	1141.0
153.90	769.0
155.00	1981.0
156.00	3191.0
157.20	614.0
158.00	531.0
159.00	598.0
160.00	1210.0
161.10	1740.0
162.00	609.0
165.00	1606.0
166.00	855.0
167.00	7986.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

168.00	4602.0
168.90	622.0
170.00	324.0
170.90	550.0
172.00	669.0
172.90	1395.0
174.00	1528.0
175.10	3151.0
176.00	725.0
177.00	1905.0
178.00	568.0
178.90	5280.0
180.00	4316.0
181.00	2017.0
181.90	410.0
184.00	430.0
185.00	2474.0
186.00	21592.0
187.10	6411.0
188.00	568.0
189.00	1553.0
190.90	883.0
192.00	1562.0
193.00	2261.0
193.90	443.0
194.90	301.0
196.00	3784.0
196.80	754.0
197.10	721.0
198.00	143616.0
199.00	9858.0
199.90	544.0
201.40	586.0
202.90	1309.0
204.10	5358.0
205.00	8451.0
206.00	40448.0
207.10	7446.0
208.10	1837.0
209.00	721.0
210.20	552.0
211.00	1734.0
215.00	534.0
216.10	1056.0
217.00	11209.0
218.00	1327.0
221.10	5121.0
222.00	1573.0
223.00	2164.0
224.00	22752.0
225.10	5848.0
226.10	985.0
227.00	9287.0
228.00	1243.0
229.10	2445.0
230.90	895.0
233.10	310.0
234.00	590.0
235.00	1129.0
235.90	496.0
237.00	1521.0
239.00	570.0
240.10	317.0
241.00	446.0
242.00	1581.0
243.00	1573.0
244.10	17280.0
245.00	2452.0
246.00	4220.0
247.00	998.0
249.00	744.0
253.00	524.0
254.10	724.0
255.00	91960.0
256.00	13861.0
257.00	1443.0
258.00	6073.0
259.00	876.0
260.00	312.0
265.00	1982.0
265.90	452.0
273.00	2682.0
274.00	7043.0
275.00	39200.0
276.00	5437.0
277.00	3553.0
278.00	1068.0
281.10	379.0
282.90	633.0
283.90	355.0
285.10	813.0
292.90	852.0
293.90	385.0
296.00	11884.0
297.00	2304.0
303.00	1873.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

304.00	433.0
314.00	588.0
315.00	1151.0
316.10	709.0
321.10	330.0
323.00	3346.0
324.10	741.0
326.90	842.0
327.90	329.0
333.10	387.0
334.00	1945.0
335.00	609.0
341.00	568.0
345.90	568.0
352.00	1101.0
353.00	703.0
354.00	1155.0
355.00	376.0
365.00	5388.0
366.00	931.0
372.00	1827.0
372.90	582.0
383.10	418.0
401.90	893.0
402.90	1203.0
403.90	311.0
421.10	782.0
422.00	710.0
423.00	5085.0
424.00	1987.0
441.00	14936.0
442.10	89176.0
443.10	19672.0
444.00	1934.0
445.10	317.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
37.00	384.0
38.10	647.0
39.10	4234.0
40.00	593.0
44.10	717.0
49.00	476.0
50.10	16704.0
51.10	64912.0
52.10	3254.0
55.00	426.0
56.00	1959.0
57.10	4076.0
61.10	785.0
62.00	1315.0
63.10	2450.0
64.00	529.0
65.10	1705.0
68.00	1363.0
69.00	72656.0
70.00	349.0
73.10	1073.0
74.10	8062.0
75.00	11185.0
76.10	4076.0
77.10	81056.0
78.10	5620.0
79.00	6323.0
80.00	4843.0
81.00	6425.0
82.00	1453.0
83.00	1520.0
83.90	319.0
85.00	1118.0
86.00	2038.0
87.00	1027.0
90.90	1591.0
92.00	1737.0
93.00	10367.0
94.00	795.0
96.00	520.0
98.00	8704.0
99.00	6923.0
100.00	564.0
101.00	3375.0
103.00	1302.0
104.00	2718.0
105.00	2128.0
106.00	577.0
107.00	28472.0
108.10	4139.0
109.10	910.0
110.00	45040.0
111.00	6732.0
112.10	1032.0
113.10	403.0
114.90	355.0
116.00	1634.0
117.00	25384.0
118.00	1801.0
119.80	331.0
122.00	1739.0
123.00	2802.0
123.90	1395.0
125.00	1200.0
126.20	746.0
127.00	83664.0
128.00	6442.0
129.00	34272.0
130.00	2717.0
131.10	611.0
132.10	322.0
133.10	329.0
134.00	944.0
135.00	3049.0
136.00	1113.0
137.00	2142.0
137.90	469.0
140.00	543.0
141.00	6027.0
142.00	1857.0
143.00	1250.0
144.00	622.0
146.10	727.0
147.00	2763.0
148.00	6242.0
149.00	1358.0
150.10	457.0
151.00	543.0
151.20	539.0
151.90	336.0
153.00	1773.0
154.00	1426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

155.00	2788.0
156.00	4072.0
157.10	880.0
157.90	885.0
158.90	649.0
160.00	1544.0
161.00	2610.0
162.00	791.0
164.00	396.0
165.00	1868.0
166.00	1450.0
167.00	10243.0
168.00	5825.0
169.00	1141.0
169.90	411.0
170.90	527.0
172.00	854.0
173.10	1267.0
174.00	2326.0
175.00	5040.0
176.00	942.0
176.90	1917.0
178.10	736.0
179.00	7880.0
180.00	5551.0
181.00	2669.0
182.00	519.0
184.00	615.0
185.00	3972.0
186.00	29000.0
187.00	8349.0
188.00	852.0
189.00	2178.0
189.90	390.0
191.00	884.0
192.00	3053.0
193.00	3115.0
194.10	683.0
194.90	340.0
196.00	5963.0
197.10	1333.0
198.00	200960.0
199.00	14435.0
200.00	1184.0
201.40	603.0
202.10	306.0
203.00	1776.0
204.00	8152.0
205.00	14094.0
206.10	55944.0
207.00	9099.0
208.00	2422.0
209.00	956.0
210.00	804.0
211.00	2241.0
211.90	465.0
212.90	337.0
215.00	750.0
216.10	1081.0
217.00	15775.0
218.00	2967.0
221.00	7761.0
222.00	1535.0
223.00	3993.0
224.00	32704.0
225.10	8299.0
226.00	1033.0
227.00	15638.0
228.00	2006.0
229.00	2815.0
230.10	370.0
231.00	1516.0
233.00	465.0
233.90	931.0
234.90	932.0
236.00	922.0
237.00	1665.0
238.90	582.0
239.90	602.0
241.10	758.0
241.90	1712.0
243.10	1646.0
244.00	23448.0
245.00	3188.0
246.00	6250.0
247.00	919.0
249.00	1029.0
253.00	618.0
253.90	704.0
254.20	960.0
255.00	139520.0
256.00	19736.0
257.00	1749.0
258.00	9038.0
258.90	1642.0
265.00	3837.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

270.90	553.0
273.00	4644.0
274.00	10481.0
275.00	59352.0
276.00	8325.0
277.00	4961.0
278.00	897.0
281.00	557.0
282.00	436.0
283.00	742.0
284.10	448.0
285.10	1162.0
291.90	364.0
293.00	1240.0
295.00	567.0
296.00	19072.0
297.00	2758.0
303.00	2062.0
304.10	594.0
308.90	331.0
314.00	1302.0
315.00	2370.0
315.90	1059.0
321.00	651.0
323.10	5469.0
324.00	1400.0
327.00	829.0
328.00	658.0
332.00	314.0
333.00	503.0
334.00	3849.0
335.00	1035.0
341.00	692.0
346.00	1652.0
346.90	307.0
352.00	1583.0
353.00	1005.0
354.00	1744.0
355.10	509.0
365.00	8682.0
365.90	1152.0
371.00	320.0
372.00	2676.0
373.00	763.0
383.00	880.0
390.00	570.0
401.90	1031.0
403.00	1519.0
404.00	485.0
420.90	1615.0
422.00	1342.0
423.10	9627.0
424.00	3124.0
425.00	589.0
441.00	25280.0
442.10	161664.0
443.10	32216.0
444.00	3072.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

m/z	Abundance
38.10	480.0
39.00	2670.0
40.10	567.0
44.00	640.0
49.00	366.0
50.10	11335.0
51.10	41304.0
52.10	1989.0
55.00	367.0
56.00	1104.0
57.00	2817.0
61.00	697.0
62.00	838.0
63.10	1704.0
63.90	358.0
65.00	789.0
68.10	1050.0
69.00	49360.0
73.00	745.0
74.10	5887.0
75.00	8076.0
76.10	2599.0
77.10	56200.0
78.10	4246.0
79.00	4355.0
80.00	3298.0
81.00	4696.0
82.10	951.0
83.00	1189.0
85.00	893.0
86.00	1927.0
87.00	541.0
90.90	990.0
92.10	1334.0
93.00	6961.0
94.00	340.0
96.00	492.0
98.00	6606.0
99.00	5203.0
100.00	391.0
100.90	2384.0
103.00	890.0
104.00	1760.0
105.00	1713.0
106.00	682.0
107.00	18496.0
108.10	2474.0
109.00	633.0
110.00	32992.0
111.00	5673.0
112.00	498.0
116.00	979.0
117.00	18552.0
117.90	1244.0
118.90	314.0
122.00	1432.0
123.00	2413.0
124.00	824.0
124.80	755.0
127.00	62672.0
128.00	5005.0
129.00	25056.0
130.00	2502.0
130.90	683.0
132.00	306.0
133.00	338.0
134.00	916.0
135.00	2343.0
136.00	937.0
137.10	1150.0
140.00	329.0
141.00	3574.0
142.00	1076.0
143.00	880.0
145.00	385.0
146.00	990.0
147.10	1891.0
148.00	4643.0
148.90	949.0
149.90	300.0
150.90	492.0
151.80	516.0
153.00	1434.0
153.90	823.0
155.00	1953.0
156.00	3237.0
157.00	471.0
157.90	602.0
159.00	469.0
160.00	1097.0
161.00	1757.0
161.90	493.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

165.00	1738.0
166.00	1376.0
167.00	8340.0
168.00	4763.0
168.90	570.0
170.00	437.0
171.10	456.0
171.90	698.0
173.00	966.0
174.00	1613.0
175.00	2605.0
176.00	1149.0
177.00	1799.0
178.00	666.0
179.00	5709.0
180.00	4709.0
181.00	1987.0
182.00	309.0
184.00	430.0
185.00	3252.0
186.00	22200.0
187.10	6529.0
187.90	664.0
189.00	1389.0
190.10	392.0
191.10	600.0
192.00	2249.0
193.10	2175.0
193.90	361.0
195.20	309.0
196.00	4112.0
197.20	1033.0
198.00	154688.0
199.00	10271.0
200.00	1013.0
201.50	792.0
203.10	1569.0
204.00	6149.0
205.10	10675.0
206.10	46000.0
207.10	7080.0
208.00	2019.0
209.00	1031.0
210.00	719.0
211.10	1753.0
212.00	500.0
214.90	527.0
216.10	941.0
217.00	12768.0
218.00	1682.0
221.00	6489.0
221.90	1609.0
222.90	2535.0
224.00	26288.0
225.00	6751.0
226.00	810.0
227.00	12020.0
228.00	1656.0
229.00	2789.0
231.00	714.0
234.00	658.0
235.00	1144.0
235.90	376.0
237.00	1374.0
239.00	532.0
240.00	345.0
241.00	676.0
242.10	1347.0
243.00	1445.0
244.00	19024.0
245.00	2933.0
246.00	5013.0
246.90	1163.0
247.90	339.0
248.90	726.0
253.00	366.0
254.10	869.0
255.00	109720.0
256.00	15632.0
257.00	1244.0
258.00	7859.0
259.00	1157.0
265.00	2827.0
265.90	311.0
271.00	342.0
272.00	502.0
273.00	3519.0
274.00	8467.0
275.00	51008.0
276.00	6343.0
277.10	4298.0
277.90	701.0
281.10	498.0
283.00	603.0
284.10	488.0
285.10	777.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59

289.90	341.0
293.00	919.0
294.00	345.0
296.00	16496.0
297.00	3027.0
302.00	343.0
303.00	1616.0
314.00	1028.0
315.00	1504.0
316.00	1081.0
321.00	730.0
323.00	4924.0
324.00	953.0
327.00	858.0
328.00	615.0
333.00	587.0
334.00	3552.0
334.90	969.0
341.10	605.0
346.00	1441.0
346.80	326.0
352.00	1646.0
352.90	1132.0
353.90	1758.0
365.00	7391.0
366.00	1234.0
371.00	430.0
372.10	2478.0
373.00	729.0
383.00	609.0
390.00	315.0
402.00	924.0
402.90	1196.0
404.00	515.0
421.00	863.0
422.00	1147.0
423.00	8488.0
424.00	2163.0
425.00	431.0
441.10	25752.0
442.00	165184.0
443.10	30312.0
444.10	2423.0

Instrument :
BNA_M
ClientSampleId :
DFTPP59