

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037814.D
 Acq On : 01 Dec 2022 17:38
 Operator : CG/JU
 Sample : SST02098
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020049

Quant Time: Dec 02 00:18:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 00:13:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	234345	20.000	ng/u1	0.00
20) Naphthalene-d8	10.922	136	1119768	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.728	164	752110	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.463	188	1624230	20.000	ng/u1	0.00
79) Chrysene-d12	21.633	240	1724202	20.000	ng/u1	0.00
88) Perylene-d12	24.121	264	1740332	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	41110	7.275	ng/uL	0.00
4) Pyridine-d5	3.870	84	322504	17.534	ng/u1	0.00
7) Phenol-d5	7.246	99	430758	18.425	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.422	67	259024	16.608	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	337504	20.775	ng/u1	0.00
15) 4-Methylphenol-d8	8.805	113	341984	18.382	ng/u1	0.00
21) Nitrobenzene-d5	9.269	128	183586	20.913	ng/u1	0.00
24) 2-Nitrophenol-d4	9.999	143	189540	21.913	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	354801	22.414	ng/u1	0.00
31) 4-Chloroaniline-d4	11.051	131	530994	20.215	ng/u1	0.00
46) Dimethylphthalate-d6	14.134	166	1063811	19.969	ng/u1	0.00
49) Acenaphthylene-d8	14.422	160	1295890	20.609	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	214638	18.291	ng/u1	0.00
60) Fluorene-d10	15.710	176	975287	20.903	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	165890	18.986	ng/u1	0.00
73) Anthracene-d10	17.563	188	1559652	20.534	ng/u1	0.00
81) Pyrene-d10	19.845	212	1808198	20.637	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	1827032	20.396	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.464	88	46736	7.523	ng/uL	100
5) Pyridine	3.887	79	322594	17.248	ng/u1	100
6) Benzaldehyde	7.228	77	232398	18.939	ng/u1	100
8) Phenol	7.275	94	438115	17.877	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.511	93	363135	18.278	ng/u1	100
12) 2-Chlorophenol	7.658	128	362205	20.410	ng/u1	100
13) 2-Methylphenol	8.540	108	352897	18.876	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.634	45	683756	19.053	ng/u1	100
16) Acetophenone	8.928	105	550517	17.975	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.916	70	287725	16.029	ng/u1	100
18) 4-Methylphenol	8.869	108	383419	18.563	ng/u1	100
19) Hexachloroethane	9.187	117	140745	18.337	ng/u1	100
22) Nitrobenzene	9.310	77	384212	16.052	ng/u1	100
23) Isophorone	9.840	82	834315	17.294	ng/u1	100
25) 2-Nitrophenol	10.028	139	207567	21.216	ng/u1	100
26) 2,4-Dimethylphenol	10.081	107	401482	18.686	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.322	93	504694	18.237	ng/u1	100
29) 2,4-Dichlorophenol	10.563	162	357320	22.065	ng/u1	100
30) Naphthalene	10.969	128	1251335	20.438	ng/u1	100
32) 4-Chloroaniline	11.081	127	539214	19.699	ng/u1	100
33) Hexachlorobutadiene	11.251	225	222011	26.116	ng/u1	100
34) Caprolactam	11.857	113	128315	19.025	ng/u1	100
35) 4-Chloro-3-methylphenol	12.187	107	375973	18.162	ng/u1	100
36) 2-Methylnaphthalene	12.569	142	872111	20.505	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037814.D
 Acq On : 01 Dec 2022 17:38
 Operator : CG/JU
 Sample : SST02098
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020049

Quant Time: Dec 02 00:18:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 00:13:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.787	142	871858	20.376	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.928	216	444317	24.059	ng/ul	100
40) Hexachlorocyclopentadiene	12.910	237	243045	21.134	ng/ul	100
41) 2,4,6-Trichlorophenol	13.169	196	290451	21.841	ng/ul	100
42) 2,4,5-Trichlorophenol	13.234	196	313085	22.004	ng/ul	100
43) 1,1'-Biphenyl	13.563	154	1128152	19.753	ng/ul	100
44) 2-Chloronaphthalene	13.610	162	887234	19.986	ng/ul	100
45) 2-Nitroaniline	13.810	65	253475	15.061	ng/ul	100
47) Dimethylphthalate	14.181	163	1111298	19.625	ng/ul	100
48) 2,6-Dinitrotoluene	14.298	165	229395	20.979	ng/ul	100
50) Acenaphthylene	14.451	152	1427149	20.079	ng/ul	100
51) 3-Nitroaniline	14.628	138	261740	19.726	ng/ul	100
52) Acenaphthene	14.792	153	991144	19.498	ng/ul	100
53) 2,4-Dinitrophenol	14.828	184	117171	18.264	ng/ul	100
55) 4-Nitrophenol	14.928	109	129656	12.330	ng/ul	100
56) Dibenzofuran	15.122	168	1349328	20.338	ng/ul	100
57) 2,4-Dinitrotoluene	15.081	165	324345	20.215	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.345	232	279238	23.919	ng/ul	100
59) Diethylphthalate	15.534	149	1136209	18.590	ng/ul	100
61) Fluorene	15.769	166	1146239	19.944	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.757	204	544858	21.732	ng/ul	100
63) 4-Nitroaniline	15.786	138	264487	20.973	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.839	198	171249	18.467	ng/ul	100
67) N-Nitrosodiphenylamine	15.969	169	977500	19.369	ng/ul	100
68) 4-Bromophenyl-phenylether	16.651	248	345769	26.467	ng/ul	100
69) Hexachlorobenzene	16.769	284	406781	24.244	ng/ul	100
70) Atrazine	16.916	200	339791	19.836	ng/ul	100
71) Pentachlorophenol	17.110	266	240411	21.640	ng/ul	100
72) Phenanthrene	17.504	178	1836538	19.303	ng/ul	100
74) Anthracene	17.598	178	1895031	19.618	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.534	216	445641	22.935	ng/ul	100
76) Pentachlorobenzene	15.039	250	493331	23.872	ng/ul	100
77) Carbazole	17.863	167	1771888	19.294	ng/ul	100
78) Di-n-butylphthalate	18.416	149	2196917	16.526	ng/ul	100
80) Fluoranthene	19.510	202	2221260	19.845	ng/ul	100
82) Pyrene	19.874	202	2317990	19.607	ng/ul	100
83) Butylbenzylphthalate	20.757	149	1007265	18.290	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.545	252	780628	22.543	ng/ul	100
85) Benzo(a)anthracene	21.616	228	2275291	19.725	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.527	149	1536574	18.467	ng/ul	100
87) Chrysene	21.674	228	2143902	19.751	ng/ul	100
89) Di-n-octyl phthalate	22.480	149	2720294	18.298	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	2278126	19.256	ng/ul	100
91) Benzo(k)fluoranthene	23.410	252	2287828	19.829	ng/ul	100
93) Benzo(a)pyrene	24.009	252	1991586	19.368	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.721	276	2238098	18.972	ng/ul	100
95) Dibenzo(a,h)anthracene	26.739	278	2039797	19.259	ng/ul	100
96) Benzo(g,h,i)perylene	27.527	276	1932733	18.814	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.523 ng/uL
RT: 3.464 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 88 Resp: 46736
Ion Ratio Lower Upper
88 100
43 64.1 51.3 76.9
58 98.8 79.0 118.6



#5
Pyridine
Concen: 17.248 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 79 Resp: 322594
Ion Ratio Lower Upper
79 100
52 94.0 75.2 112.8
51 42.4 33.9 50.9









#6
Benzaldehyde
Concen: 18.939 ng/ul
RT: 7.228 min Scan# 611
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 77 Resp: 232398
Ion Ratio Lower Upper
77 100
105 101.5 81.2 121.8
106 98.8 79.0 118.6



#8
Phenol
Concen: 17.877 ng/ul
RT: 7.275 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 94 Resp: 438115
Ion Ratio Lower Upper
94 100
65 26.4 21.1 31.7
66 31.5 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.278 ng/ul
RT: 7.511 min Scan# 701
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 93 Resp: 363135
Ion Ratio Lower Upper
93 100
63 82.1 65.7 98.5
95 33.5 26.8 40.2



#12
2-Chlorophenol
Concen: 20.410 ng/ul
RT: 7.658 min Scan# 726
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:128 Resp: 362205
Ion Ratio Lower Upper
128 100
64 46.6 37.3 55.9
130 33.9 27.1 40.7









#13
2-Methylphenol
Concen: 18.876 ng/ul
RT: 8.540 min Scan# 876
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:108 Resp: 352897
Ion Ratio Lower Upper
108 100
107 87.6 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.053 ng/ul
RT: 8.634 min Scan# 892
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 45 Resp: 683756
Ion Ratio Lower Upper
45 100
77 10.3 8.2 12.4
79 8.4 6.7 10.1









#16
Acetophenone
Concen: 17.975 ng/ul
RT: 8.928 min Scan# 942
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:105 Resp: 550517
Ion Ratio Lower Upper
105 100
77 73.7 59.0 88.4
51 30.5 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 16.029 ng/ul
RT: 8.916 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 70 Resp: 287725
Ion Ratio Lower Upper
70 100
42 69.7 55.8 83.6
101 10.0 8.0 12.0
130 23.2 18.6 27.8









#18
4-Methylphenol
Concen: 18.563 ng/ul
RT: 8.869 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:108 Resp: 383419
Ion Ratio Lower Upper
108 100
107 111.6 89.3 133.9



#19
Hexachloroethane
Concen: 18.337 ng/ul
RT: 9.187 min Scan# 986
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:117 Resp: 140745
Ion Ratio Lower Upper
117 100
201 102.9 82.3 123.5
199 62.2 49.8 74.6









#22
Nitrobenzene
Concen: 16.052 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion: 77 Resp: 384212
Ion Ratio Lower Upper
77 100
123 51.7 41.4 62.0
65 13.6 10.9 16.3



#23
Isophorone
Concen: 17.294 ng/ul
RT: 9.840 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 82 Resp: 834315
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 19.1 15.3 22.9









#25
2-Nitrophenol
Concen: 21.216 ng/ul
RT: 10.028 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:139 Resp: 207567
Ion Ratio Lower Upper
139 100
65 45.7 36.6 54.8
109 26.5 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 18.686 ng/ul
RT: 10.081 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:107 Resp: 401482
Ion Ratio Lower Upper
107 100
121 52.3 41.8 62.8
122 91.4 73.1 109.7





#27

Bis(2-Chloroethoxy)methane

Concen: 18.237 ng/ul

RT: 10.322 min Scan# 1179

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion: 93 Resp: 504694

Ion Ratio Lower Upper

93 100

95 34.9 27.9 41.9

123 12.1 9.7 14.5



#29

2,4-Dichlorophenol

Concen: 22.065 ng/ul

RT: 10.563 min Scan# 1220

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:162 Resp: 357320


Ion Ratio Lower Upper

162 100

164 63.4 50.7 76.1

98 34.1 27.3 40.9









#30
Naphthalene
Concen: 20.438 ng/ul
RT: 10.969 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:128 Resp: 1251335
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 19.699 ng/ul
RT: 11.081 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:127 Resp: 539214
Ion Ratio Lower Upper
127 100
129 33.6 26.9 40.3





#33

Hexachlorobutadiene

Concen: 26.116 ng/ul

RT: 11.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:225 Resp: 222011

Ion Ratio Lower Upper

225 100

223 61.6 49.3 73.9

227 61.7 49.4 74.0



#34

Caprolactam

Concen: 19.025 ng/ul

RT: 11.857 min Scan# 1440

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:113 Resp: 128315


Ion Ratio Lower Upper

113 100

55 232.7 186.2 279.2

56 149.8 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 18.162 ng/ul
RT: 12.187 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:107 Resp: 375973
Ion Ratio Lower Upper
107 100
144 28.8 23.0 34.6
142 86.4 69.1 103.7



#36
2-Methylnaphthalene
Concen: 20.505 ng/ul
RT: 12.569 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:142 Resp: 872111
Ion Ratio Lower Upper
142 100
141 87.6 70.1 105.1





#37

1-Methylnaphthalene

Concen: 20.376 ng/ul

RT: 12.787 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:142 Resp: 871858

Ion Ratio Lower Upper

142 100

141 91.8 73.4 110.2

116 3.3 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.059 ng/ul

RT: 12.928 min Scan# 1622

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Tgt Ion:216 Resp: 444317




Ion Ratio Lower Upper


216 100

214 78.4 62.7 94.1

179 19.9 15.9 23.9

108 16.2 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 21.134 ng/ul
RT: 12.910 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:237 Resp: 243045
Ion Ratio Lower Upper
237 100
235 61.5 49.2 73.8
272 13.1 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 21.841 ng/ul
RT: 13.169 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:196 Resp: 290451
Ion Ratio Lower Upper
196 100
198 94.5 75.6 113.4
200 30.5 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 22.004 ng/ul
RT: 13.234 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:196 Resp: 313085
Ion Ratio Lower Upper
196 100
198 95.1 76.1 114.1
200 31.0 24.8 37.2



#43
1,1'-Biphenyl
Concen: 19.753 ng/ul
RT: 13.563 min Scan# 1730
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:154 Resp: 1128152
Ion Ratio Lower Upper
154 100
153 39.2 31.4 47.0
76 12.7 10.2 15.2









#44
2-Chloronaphthalene
Concen: 19.986 ng/ul
RT: 13.610 min Scan# 1738
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:162 Resp: 887234
Ion Ratio Lower Upper
162 100
164 33.1 26.5 39.7
127 38.1 30.5 45.7



#45
2-Nitroaniline
Concen: 15.061 ng/ul
RT: 13.810 min Scan# 1772
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion: 65 Resp: 253475
Ion Ratio Lower Upper
65 100
92 68.4 54.7 82.1
138 121.0 96.8 145.2









#47
Dimethylphthalate
Concen: 19.625 ng/ul
RT: 14.181 min Scan# 1835
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:163 Resp: 1111298
Ion Ratio Lower Upper
163 100
194 4.5 3.6 5.4
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.979 ng/ul
RT: 14.298 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:165 Resp: 229395
Ion Ratio Lower Upper
165 100
89 50.7 40.6 60.8
121 20.5 16.4 24.6





#50

Acenaphthylene

Concen: 20.079 ng/ul

RT: 14.451 min Scan# 1881

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:152 Resp: 1427149

Ion Ratio Lower Upper

152 100

151 19.7 15.8 23.6

153 13.2 10.6 15.8



#51

3-Nitroaniline

Concen: 19.726 ng/ul

RT: 14.628 min Scan# 1911

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:138 Resp: 261740


Ion Ratio Lower Upper

138 100

108 9.8 7.8 11.8

92 104.7 83.8 125.6









#52
Acenaphthene
Concen: 19.498 ng/ul
RT: 14.792 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:153 Resp: 991144
Ion Ratio Lower Upper
153 100
152 46.2 37.0 55.4
154 87.2 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 18.264 ng/ul
RT: 14.828 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:184 Resp: 117171
Ion Ratio Lower Upper
184 100
63 54.2 43.4 65.0
154 49.6 39.7 59.5









#55
4-Nitrophenol
Concen: 12.330 ng/ul
RT: 14.928 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:109 Resp: 129656
Ion Ratio Lower Upper
109 100
139 164.5 131.6 197.4
65 136.7 109.4 164.0



#56
Dibenzofuran
Concen: 20.338 ng/ul
RT: 15.122 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:168 Resp: 1349328
Ion Ratio Lower Upper
168 100
139 34.5 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 20.215 ng/ul

RT: 15.081 min Scan# 1988

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:165 Resp: 324345

Ion Ratio Lower Upper

165 100

63 37.7 30.2 45.2

182 2.3 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.919 ng/ul

RT: 15.345 min Scan# 2033

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Tgt Ion:232 Resp: 279238




Ion Ratio Lower Upper


232 100

131 40.9 32.7 49.1

130 2.2 1.8 2.6

166 27.8 22.2 33.4









#59
Diethylphthalate
Concen: 18.590 ng/ul
RT: 15.534 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:149 Resp: 1136209
Ion Ratio Lower Upper
149 100
177 22.5 18.0 27.0
150 12.3 9.8 14.8



#61
Fluorene
Concen: 19.944 ng/ul
RT: 15.769 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:166 Resp: 1146239
Ion Ratio Lower Upper
166 100
165 96.3 77.0 115.6
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 21.732 ng/ul
RT: 15.757 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:204 Resp: 544858
Ion Ratio Lower Upper
204 100
206 32.8 26.2 39.4
141 55.9 44.7 67.1



#63
4-Nitroaniline
Concen: 20.973 ng/ul
RT: 15.786 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:138 Resp: 264487
Ion Ratio Lower Upper
138 100
92 45.7 36.6 54.8
108 60.2 48.2 72.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.467 ng/ul
 RT: 15.839 min Scan# 2117
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049


Tgt Ion:198 Resp: 171249
 Ion Ratio Lower Upper
 198 100
 182 3.6 2.9 4.3
 77 21.9 17.5 26.3



#67
 N-Nitrosodiphenylamine
 Concen: 19.369 ng/ul
 RT: 15.969 min Scan# 2139
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38

Tgt Ion:169 Resp: 977500
 Ion Ratio Lower Upper
 169 100
 168 67.2 53.8 80.6
 167 36.9 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 26.467 ng/ul
RT: 16.651 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:248 Resp: 345769
Ion Ratio Lower Upper
248 100
250 97.6 78.1 117.1
141 64.5 51.6 77.4



#69
Hexachlorobenzene
Concen: 24.244 ng/ul
RT: 16.769 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:284 Resp: 406781
Ion Ratio Lower Upper
284 100
142 30.6 24.5 36.7
249 28.6 22.9 34.3





#70

Atrazine

Concen: 19.836 ng/ul

RT: 16.916 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:200 Resp: 339791

Ion Ratio Lower Upper

200 100

173 27.5 22.0 33.0

215 50.0 40.0 60.0



#71

Pentachlorophenol

Concen: 21.640 ng/ul

RT: 17.110 min Scan# 2333

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:266 Resp: 240411


Ion Ratio Lower Upper

266 100

264 62.9 50.3 75.5

268 65.2 52.2 78.2





#72

Phenanthrene

Concen: 19.303 ng/ul

RT: 17.504 min Scan# 2400

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:178 Resp: 1836538

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.1 15.3 22.9



#74

Anthracene

Concen: 19.618 ng/ul

RT: 17.598 min Scan# 2416

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:178 Resp: 1895031


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.935 ng/uL
RT: 13.534 min Scan# 1725
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:216 Resp: 445641
Ion Ratio Lower Upper
216 100
214 78.8 63.0 94.6
218 50.0 40.0 60.0



#76
Pentachlorobenzene
Concen: 23.872 ng/uL
RT: 15.039 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:250 Resp: 493331
Ion Ratio Lower Upper
250 100
248 63.1 50.5 75.7
252 63.9 51.1 76.7









#77
 Carbazole
 Concen: 19.294 ng/ul
 RT: 17.863 min Scan# 2461
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049


Tgt Ion:167 Resp: 1771888
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.2 25.8
 139 12.5 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 16.526 ng/ul
 RT: 18.416 min Scan# 2555
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38

Tgt Ion:149 Resp: 2196917
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.2 3.4 5.0









#80
Fluoranthene
Concen: 19.845 ng/ul
RT: 19.510 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:202 Resp: 2221260
Ion Ratio Lower Upper
202 100
101 12.1 9.7 14.5
100 9.6 7.7 11.5



#82
Pyrene
Concen: 19.607 ng/ul
RT: 19.874 min Scan# 2803
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:202 Resp: 2317990
Ion Ratio Lower Upper
202 100
101 13.7 11.0 16.4
100 11.3 9.0 13.6





#83

Butylbenzylphthalate

Concen: 18.290 ng/ul

RT: 20.757 min Scan# 2953

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:149 Resp: 1007265

Ion Ratio Lower Upper

149 100

91 64.6 51.7 77.5

206 21.4 17.1 25.7



#84

3,3'-Dichlorobenzidine

Concen: 22.543 ng/ul

RT: 21.545 min Scan# 3087

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:252 Resp: 780628


Ion Ratio Lower Upper

252 100

254 65.3 52.2 78.4

126 11.0 8.8 13.2









#85
Benzo(a)anthracene
Concen: 19.725 ng/ul
RT: 21.616 min Scan# 3099
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38




Instrument :
BNA_M
ClientSampleId :
SSTD020049


Tgt Ion:228 Resp: 2275291
Ion Ratio Lower Upper
228 100
229 20.2 16.2 24.2
226 27.9 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.467 ng/ul
RT: 21.527 min Scan# 3084
Delta R.T. 0.000 min
Lab File: BM037814.D
Acq: 01 Dec 2022 17:38

Tgt Ion:149 Resp: 1536574
Ion Ratio Lower Upper
149 100
167 29.1 23.3 34.9
279 4.5 3.6 5.4





#87

Chrysene

Concen: 19.751 ng/ul

RT: 21.674 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:228 Resp: 2143902

Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.2

229 19.7 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 18.298 ng/ul




RT: 22.480 min Scan# 3246


Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Tgt Ion:149 Resp: 2720294





#90

Benzo(b)fluoranthene

Concen: 19.256 ng/ul

RT: 23.357 min Scan# 3395

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:252 Resp: 2278126

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 11.6 9.3 13.9



#91

Benzo(k)fluoranthene

Concen: 19.829 ng/ul

RT: 23.410 min Scan# 3404

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:252 Resp: 2287828


Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 11.6 9.3 13.9





#93

Benzo(a)pyrene

Concen: 19.368 ng/ul

RT: 24.009 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38

Instrument :

BNA_M

ClientSampleId :

SSTD020049





Tgt Ion:252 Resp: 1991586

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 12.8 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.972 ng/ul

RT: 26.721 min Scan# 3967

Delta R.T. 0.000 min

Lab File: BM037814.D

Acq: 01 Dec 2022 17:38




Tgt Ion:276 Resp: 2238098


Ion Ratio Lower Upper

276 100

138 28.4 22.7 34.1

277 25.1 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 19.259 ng/ul
 RT: 26.739 min Scan# 3970
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020049

Tgt Ion:278 Resp: 2039797
 Ion Ratio Lower Upper
 278 100
 139 17.4 13.9 20.9
 279 23.8 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 18.814 ng/ul
 RT: 27.527 min Scan# 4104
 Delta R.T. 0.000 min
 Lab File: BM037814.D
 Acq: 01 Dec 2022 17:38

Tgt Ion:276 Resp: 1932733
 Ion Ratio Lower Upper
 276 100
 138 24.0 19.2 28.8
 277 23.9 19.1 28.7

