

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037817.D
 Acq On : 01 Dec 2022 19:29
 Operator : CG/JU
 Sample : SSTD16002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160003

Quant Time: Dec 02 00:19:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 00:13:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.098	152	266161	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	1260690	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.733	164	824668	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.468	188	1809107	20.000	ng/u1	0.00
79) Chrysene-d12	21.645	240	1574971	20.000	ng/u1	0.01
88) Perylene-d12	24.133	264	1727458	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	333871	52.022	ng/uL	0.00
4) Pyridine-d5	3.863	84	3132834	149.966	ng/u1	0.00
7) Phenol-d5	7.269	99	4021492	151.448	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.434	67	2194122	123.869	ng/u1	0.01
11) 2-Chlorophenol-d4	7.634	132	3138921	170.120	ng/u1	0.01
15) 4-Methylphenol-d8	8.828	113	3157824	149.445	ng/u1	0.02
21) Nitrobenzene-d5	9.287	128	1682209	170.209	ng/u1	0.02
24) 2-Nitrophenol-d4	10.010	143	1839306	188.871	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.551	165	3171793	177.977	ng/u1	0.02
31) 4-Chloroaniline-d4	11.069	131	4639521	156.886	ng/u1	0.02
46) Dimethylphthalate-d6	14.151	166	8707268	149.064	ng/u1	0.02
49) Acenaphthylene-d8	14.433	160	10151572	147.237	ng/u1	0.01
54) 4-Nitrophenol-d4	14.957	143	1967975	152.952	ng/u1	0.05
60) Fluorene-d10	15.727	176	7164577	140.047	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.857	200	1657124	170.277	ng/u1	0.04
73) Anthracene-d10	17.580	188	10793590	127.584	ng/u1	0.02
81) Pyrene-d10	19.857	212	11874610	148.366	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.986	264	12305175	138.389	ng/u1	0.03
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	367693	52.113	ng/uL#	78
5) Pyridine	3.881	79	3023847	142.349	ng/u1	99
6) Benzaldehyde	7.234	77	1401272	100.544	ng/u1	97
8) Phenol	7.299	94	4119140	147.991	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.528	93	3087294	136.818	ng/u1	96
12) 2-Chlorophenol	7.669	128	3318644	164.646	ng/u1	97
13) 2-Methylphenol	8.551	108	3211943	151.266	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.640	45	5548803	136.138	ng/u1	99
16) Acetophenone	8.945	105	4699007	135.089	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.963	70	2316815	113.640	ng/u1	96
18) 4-Methylphenol	8.910	108	3495220	148.992	ng/u1	96
19) Hexachloroethane	9.193	117	1226091	140.644	ng/u1	99
22) Nitrobenzene	9.334	77	3482520	129.230	ng/u1	100
23) Isophorone	9.881	82	7172867	132.066	ng/u1	99
25) 2-Nitrophenol	10.045	139	1969176	178.780	ng/u1	94
26) 2,4-Dimethylphenol	10.098	107	3521883	145.597	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.339	93	4247923	136.340	ng/u1	97
29) 2,4-Dichlorophenol	10.581	162	3192462	175.103	ng/u1	98
30) Naphthalene	10.981	128	10484115	152.094	ng/u1	99
32) 4-Chloroaniline	11.098	127	4619855	149.907	ng/u1	98
33) Hexachlorobutadiene	11.257	225	1813470	189.477	ng/u1	97
34) Caprolactam	11.951	113	573883	75.578	ng/u1	83
35) 4-Chloro-3-methylphenol	12.222	107	3377860	144.935	ng/u1	97
36) 2-Methylnaphthalene	12.575	142	7070254	147.652	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120122\
 Data File : BM037817.D
 Acq On : 01 Dec 2022 19:29
 Operator : CG/JU
 Sample : SSTD16002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160003

Quant Time: Dec 02 00:19:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 00:13:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.798	142	7057555	146.503	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.939	216	3368573	166.355	ng/ul	99
40) Hexachlorocyclopentadiene	12.916	237	2014473	159.755	ng/ul	99
41) 2,4,6-Trichlorophenol	13.180	196	2491376	170.862	ng/ul	97
42) 2,4,5-Trichlorophenol	13.257	196	2719789	174.331	ng/ul	99
43) 1,1'-Biphenyl	13.580	154	8505838	135.824	ng/ul	97
44) 2-Chloronaphthalene	13.622	162	6950797	142.802	ng/ul	98
45) 2-Nitroaniline	13.833	65	2299179	124.592	ng/ul	98
47) Dimethylphthalate	14.198	163	8796169	141.666	ng/ul	99
48) 2,6-Dinitrotoluene	14.327	165	2054343	171.343	ng/ul	100
50) Acenaphthylene	14.463	152	10643714	136.572	ng/ul	97
51) 3-Nitroaniline	14.657	138	2390030	164.279	ng/ul	100
52) Acenaphthene	14.804	153	7399811	132.763	ng/ul	97
53) 2,4-Dinitrophenol	14.851	184	1383656	196.697	ng/ul	94
55) 4-Nitrophenol	14.974	109	1218251	105.657	ng/ul	95
56) Dibenzofuran	15.139	168	9792096	134.605	ng/ul	93
57) 2,4-Dinitrotoluene	15.110	165	2867168	162.972	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.357	232	2240775	175.055	ng/ul	89
59) Diethylphthalate	15.557	149	8881949	132.537	ng/ul	99
61) Fluorene	15.786	166	7119035	112.970	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.769	204	3398917	123.641	ng/ul	93
63) 4-Nitroaniline	15.839	138	2152251	155.653	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.874	198	1672701	161.947	ng/ul	99
67) N-Nitrosodiphenylamine	15.986	169	6944086	123.535	ng/ul	100
68) 4-Bromophenyl-phenylether	16.663	248	2455716	168.762	ng/ul	95
69) Hexachlorobenzene	16.780	284	2882606	154.245	ng/ul	94
70) Atrazine	16.939	200	2395346	125.544	ng/ul	99
71) Pentachlorophenol	17.121	266	2015711	162.898	ng/ul	100
72) Phenanthrene	17.521	178	12380264	116.823	ng/ul	97
74) Anthracene	17.615	178	12034508	111.855	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.539	216	3508039	162.094	ng/ul	99
76) Pentachlorobenzene	15.057	250	3542231	153.889	ng/ul	99
77) Carbazole	17.880	167	12306635	120.314	ng/ul	96
78) Di-n-butylphthalate	18.421	149	13817150	93.317	ng/ul#	95
80) Fluoranthene	19.521	202	13678686	133.789	ng/ul	98
82) Pyrene	19.892	202	13034791	120.705	ng/ul	98
83) Butylbenzylphthalate	20.762	149	6907730	137.315	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.556	252	4266809	134.892	ng/ul	100
85) Benzo(a)anthracene	21.627	228	13497615	128.104	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.527	149	8868869	116.689	ng/ul	98
87) Chrysene	21.692	228	12126815	122.304	ng/ul	98
89) Di-n-octyl phthalate	22.480	149	15500475	105.042	ng/ul	100
90) Benzo(b)fluoranthene	23.380	252	16782747	142.911	ng/ul	98
91) Benzo(k)fluoranthene	23.439	252	12797454	111.745	ng/ul	98
93) Benzo(a)pyrene	24.044	252	13143727	128.774	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.762	276	14520150	124.004	ng/ul	97
95) Dibenzo(a,h)anthracene	26.797	278	12596753	119.817	ng/ul	98
96) Benzo(g,h,i)perylene	27.591	276	13509648	132.491	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 52.113 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion: 88 Resp: 367693
Ion Ratio Lower Upper
88 100
43 36.7 51.3 76.9#
58 85.0 79.0 118.6



#5
Pyridine
Concen: 142.349 ng/uL
RT: 3.881 min Scan# 84
Delta R.T. -0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 79 Resp: 3023847
Ion Ratio Lower Upper
79 100
52 94.6 75.2 112.8
51 40.4 33.9 50.9









#6
Benzaldehyde
Concen: 100.544 ng/ul
RT: 7.234 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion: 77 Resp: 1401272
Ion Ratio Lower Upper
77 100
105 103.5 81.2 121.8
106 102.5 79.0 118.6



#8
Phenol
Concen: 147.991 ng/ul
RT: 7.299 min Scan# 665
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 94 Resp: 4119140
Ion Ratio Lower Upper
94 100
65 25.9 21.1 31.7
66 31.5 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 136.818 ng/ul
RT: 7.528 min Scan# 701
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion: 93 Resp: 3087294
Ion Ratio Lower Upper
93 100
63 77.7 65.7 98.5
95 32.5 26.8 40.2



#12
2-Chlorophenol
Concen: 164.646 ng/ul
RT: 7.669 min Scan# 728
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:128 Resp: 3318644
Ion Ratio Lower Upper
128 100
64 43.6 37.3 55.9
130 33.5 27.1 40.7









#13
2-Methylphenol
Concen: 151.266 ng/ul
RT: 8.551 min Scan# 81
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:108 Resp: 3211943
Ion Ratio Lower Upper
108 100
107 87.5 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 136.138 ng/ul
RT: 8.640 min Scan# 893
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 45 Resp: 5548803
Ion Ratio Lower Upper
45 100
77 10.6 8.2 12.4
79 8.7 6.7 10.1









#16
Acetophenone
Concen: 135.089 ng/ul
RT: 8.945 min Scan# 945
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:105 Resp: 4699007
Ion Ratio Lower Upper
105 100
77 72.2 59.0 88.4
51 28.7 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 113.640 ng/ul
RT: 8.963 min Scan# 948
Delta R.T. 0.047 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 70 Resp: 2316815
Ion Ratio Lower Upper
70 100
42 66.0 55.8 83.6
101 10.1 8.0 12.0
130 24.3 18.6 27.8









#18
4-Methylphenol
Concen: 148.992 ng/ul
RT: 8.910 min Scan# 910
Delta R.T. 0.041 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:108 Resp: 3495220
Ion Ratio Lower Upper
108 100
107 106.9 89.3 133.9



#19
Hexachloroethane
Concen: 140.644 ng/ul
RT: 9.193 min Scan# 987
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:117 Resp: 1226091
Ion Ratio Lower Upper
117 100
201 102.6 82.3 123.5
199 64.5 49.8 74.6









#22
Nitrobenzene
Concen: 129.230 ng/ul
RT: 9.334 min Scan# 1104
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion: 77 Resp: 3482520
Ion Ratio Lower Upper
77 100
123 51.9 41.4 62.0
65 13.3 10.9 16.3



#23
Isophorone
Concen: 132.066 ng/ul
RT: 9.881 min Scan# 1104
Delta R.T. 0.041 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 82 Resp: 7172867
Ion Ratio Lower Upper
82 100
95 6.4 5.5 8.3
138 19.3 15.3 22.9









#25
2-Nitrophenol
Concen: 178.780 ng/ul
RT: 10.045 min Scan# 1132
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:139 Resp: 1969176
Ion Ratio Lower Upper
139 100
65 42.5 36.6 54.8
109 23.0 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 145.597 ng/ul
RT: 10.098 min Scan# 1141
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:107 Resp: 3521883
Ion Ratio Lower Upper
107 100
121 52.7 41.8 62.8
122 92.4 73.1 109.7





#27

Bis(2-Chloroethoxy)methane

Concen: 136.340 ng/ul

RT: 10.339 min Scan# 1182

Delta R.T. 0.017 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :





BNA_M

ClientSampleId :

SSTD160003

Tgt Ion: 93 Resp: 4247923

Ion	Ratio	Lower	Upper
93	100		
95	32.8	27.9	41.9
123	11.8	9.7	14.5



#29

2,4-Dichlorophenol

Concen: 175.103 ng/ul

RT: 10.581 min Scan# 1223




Delta R.T. 0.017 min


Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Tgt Ion:162 Resp: 3192462

Ion	Ratio	Lower	Upper
162	100		
164	65.7	50.7	76.1
98	34.7	27.3	40.9









#30
Naphthalene
Concen: 152.094 ng/ul
RT: 10.981 min Scan# 1311
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:128 Resp:10484115
Ion Ratio Lower Upper
128 100
129 11.4 8.9 13.3
127 13.5 10.4 15.6



#32
4-Chloroaniline
Concen: 149.907 ng/ul
RT: 11.098 min Scan# 1311
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:127 Resp: 4619855
Ion Ratio Lower Upper
127 100
129 32.5 26.9 40.3









#33
Hexachlorobutadiene
Concen: 189.477 ng/ul
RT: 11.257 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:225 Resp: 1813470
Ion Ratio Lower Upper
225 100
223 63.2 49.3 73.9
227 64.8 49.4 74.0



#34
Caprolactam
Concen: 75.578 ng/ul
RT: 11.951 min Scan# 1456
Delta R.T. 0.094 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:113 Resp: 573883
Ion Ratio Lower Upper
113 100
55 195.7 186.2 279.2
56 137.1 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 144.935 ng/ul
RT: 12.222 min Scan# 1
Delta R.T. 0.035 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:107 Resp: 3377860
Ion Ratio Lower Upper
107 100
144 27.6 23.0 34.6
142 83.9 69.1 103.7



#36
2-Methylnaphthalene
Concen: 147.652 ng/ul
RT: 12.575 min Scan# 1562
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:142 Resp: 7070254
Ion Ratio Lower Upper
142 100
141 87.7 70.1 105.1





#37

1-Methylnaphthalene

Concen: 146.503 ng/ul

RT: 12.798 min Scan# 1

Delta R.T. 0.011 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:142 Resp: 7057555

Ion Ratio Lower Upper

142 100

141 92.0 73.4 110.2

116 3.6 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 166.355 ng/ul

RT: 12.939 min Scan# 1624

Delta R.T. 0.011 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Tgt Ion:216 Resp: 3368573




Ion Ratio Lower Upper


216 100

214 78.8 62.7 94.1

179 20.8 15.9 23.9

108 17.6 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 159.755 ng/ul
RT: 12.916 min Scan# 1619
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:237 Resp: 2014473
Ion Ratio Lower Upper
237 100
235 62.6 49.2 73.8
272 13.2 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 170.862 ng/ul
RT: 13.180 min Scan# 1665
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:196 Resp: 2491376
Ion Ratio Lower Upper
196 100
198 97.4 75.6 113.4
200 31.4 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 174.331 ng/ul
RT: 13.257 min Scan# 1
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:196 Resp: 2719789
Ion Ratio Lower Upper
196 100
198 96.6 76.1 114.1
200 31.5 24.8 37.2



#43
1,1'-Biphenyl
Concen: 135.824 ng/ul
RT: 13.580 min Scan# 1733
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:154 Resp: 8505838
Ion Ratio Lower Upper
154 100
153 40.9 31.4 47.0
76 11.8 10.2 15.2









#44
2-Chloronaphthalene
Concen: 142.802 ng/ul
RT: 13.622 min Scan# 1738
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:162 Resp: 6950797
Ion Ratio Lower Upper
162 100
164 33.7 26.5 39.7
127 39.7 30.5 45.7



#45
2-Nitroaniline
Concen: 124.592 ng/ul
RT: 13.833 min Scan# 1776
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion: 65 Resp: 2299179
Ion Ratio Lower Upper
65 100
92 71.8 54.7 82.1
138 122.2 96.8 145.2









#47
Dimethylphthalate
Concen: 141.666 ng/ul
RT: 14.198 min Scan# 1838
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:163 Resp: 8796169
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 10.6 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 171.343 ng/ul
RT: 14.327 min Scan# 1860
Delta R.T. 0.029 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:165 Resp: 2054343
Ion Ratio Lower Upper
165 100
89 50.8 40.6 60.8
121 20.0 16.4 24.6









#50
Acenaphthylene
Concen: 136.572 ng/ul
RT: 14.463 min Scan# 1883
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:152 Resp:10643714
Ion Ratio Lower Upper
152 100
151 21.0 15.8 23.6
153 14.1 10.6 15.8



#51
3-Nitroaniline
Concen: 164.279 ng/ul
RT: 14.657 min Scan# 1916
Delta R.T. 0.029 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:138 Resp: 2390030
Ion Ratio Lower Upper
138 100
108 9.3 7.8 11.8
92 104.4 83.8 125.6









#52
Acenaphthene
Concen: 132.763 ng/ul
RT: 14.804 min Scan# 1941
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:153 Resp: 7399811
Ion Ratio Lower Upper
153 100
152 48.7 37.0 55.4
154 89.7 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 196.697 ng/ul
RT: 14.851 min Scan# 1949
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:184 Resp: 1383656
Ion Ratio Lower Upper
184 100
63 58.9 43.4 65.0
154 53.8 39.7 59.5









#55
4-Nitrophenol
Concen: 105.657 ng/ul
RT: 14.974 min Scan# 1970
Delta R.T. 0.047 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:109 Resp: 1218251
Ion Ratio Lower Upper
109 100
139 169.7 131.6 197.4
65 143.5 109.4 164.0



#56
Dibenzofuran
Concen: 134.605 ng/ul
RT: 15.139 min Scan# 1998
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:168 Resp: 9792096
Ion Ratio Lower Upper
168 100
139 38.3 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 162.972 ng/ul
RT: 15.110 min Scan# 1988
Delta R.T. 0.029 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:165 Resp: 2867168
Ion Ratio Lower Upper
165 100
63 39.4 30.2 45.2
182 2.4 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 175.055 ng/ul
RT: 15.357 min Scan# 2035
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:232 Resp: 2240775
Ion Ratio Lower Upper
232 100
131 49.0 32.7 49.1
130 2.5 1.8 2.6
166 32.5 22.2 33.4









#59
Diethylphthalate
Concen: 132.537 ng/ul
RT: 15.557 min Scan# 2069
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:149 Resp: 8881949
Ion Ratio Lower Upper
149 100
177 22.8 18.0 27.0
150 12.6 9.8 14.8



#61
Fluorene
Concen: 112.970 ng/ul
RT: 15.786 min Scan# 2108
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:166 Resp: 7119035
Ion Ratio Lower Upper
166 100
165 96.9 77.0 115.6
167 13.5 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 123.641 ng/ul
RT: 15.769 min Scan# 2105
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:204 Resp: 3398917
Ion Ratio Lower Upper
204 100
206 33.0 26.2 39.4
141 63.6 44.7 67.1



#63
4-Nitroaniline
Concen: 155.653 ng/ul
RT: 15.839 min Scan# 2117
Delta R.T. 0.053 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:138 Resp: 2152251
Ion Ratio Lower Upper
138 100
92 45.0 36.6 54.8
108 67.5 48.2 72.2









#66
4,6-Dinitro-2-methylphenol
Concen: 161.947 ng/ul
RT: 15.874 min Scan# 2117
Delta R.T. 0.035 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:198 Resp: 1672701
Ion Ratio Lower Upper
198 100
182 3.0 2.9 4.3
77 21.5 17.5 26.3



#67
N-Nitrosodiphenylamine
Concen: 123.535 ng/ul
RT: 15.986 min Scan# 2142
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:169 Resp: 6944086
Ion Ratio Lower Upper
169 100
168 67.3 53.8 80.6
167 37.1 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 168.762 ng/ul
RT: 16.663 min Scan# 21
Delta R.T. 0.012 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:248 Resp: 2455716
Ion Ratio Lower Upper
248 100
250 97.9 78.1 117.1
141 74.3 51.6 77.4



#69
Hexachlorobenzene
Concen: 154.245 ng/ul
RT: 16.780 min Scan# 2277
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:284 Resp: 2882606
Ion Ratio Lower Upper
284 100
142 35.4 24.5 36.7
249 29.9 22.9 34.3





#70

Atrazine

Concen: 125.544 ng/ul

RT: 16.939 min Scan# 21

Delta R.T. 0.023 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:200 Resp: 2395346

Ion Ratio Lower Upper

200 100

173 27.7 22.0 33.0

215 48.8 40.0 60.0



#71

Pentachlorophenol

Concen: 162.898 ng/ul

RT: 17.121 min Scan# 2335

Delta R.T. 0.011 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29




Tgt Ion:266 Resp: 2015711


Ion Ratio Lower Upper

266 100

264 62.4 50.3 75.5

268 65.1 52.2 78.2





#72

Phenanthrene

Concen: 116.823 ng/ul

RT: 17.521 min Scan# 2419

Delta R.T. 0.017 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:178 Resp:12380264

Ion Ratio Lower Upper

178 100

179 16.4 12.5 18.7

176 20.8 15.3 22.9



#74

Anthracene

Concen: 111.855 ng/ul

RT: 17.615 min Scan# 2419

Delta R.T. 0.017 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29




Tgt Ion:178 Resp:12034508


Ion Ratio Lower Upper

178 100

179 16.6 12.2 18.4

176 20.2 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 162.094 ng/uL
RT: 13.539 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:216 Resp: 3508039
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.6
218 49.1 40.0 60.0



#76
Pentachlorobenzene
Concen: 153.889 ng/uL
RT: 15.057 min Scan# 1984
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:250 Resp: 3542231
Ion Ratio Lower Upper
250 100
248 63.2 50.5 75.7
252 64.7 51.1 76.7









#77
Carbazole
Concen: 120.314 ng/ul
RT: 17.880 min Scan# 2464
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:167 Resp:12306635
Ion Ratio Lower Upper
167 100
166 23.0 17.2 25.8
139 14.2 10.0 15.0



#78
Di-n-butylphthalate
Concen: 93.317 ng/ul
RT: 18.421 min Scan# 2556
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:149 Resp:13817150
Ion Ratio Lower Upper
149 100
150 11.1 7.4 11.2
104 5.8 3.4 5.0#









#80
Fluoranthene
Concen: 133.789 ng/uI
RT: 19.521 min Scan# 21
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:202 Resp:13678686
Ion Ratio Lower Upper
202 100
101 11.6 9.7 14.5
100 8.8 7.7 11.5



#82
Pyrene
Concen: 120.705 ng/uI
RT: 19.892 min Scan# 2806
Delta R.T. 0.017 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:202 Resp:13034791
Ion Ratio Lower Upper
202 100
101 13.1 11.0 16.4
100 10.7 9.0 13.6









#83
Butylbenzylphthalate
Concen: 137.315 ng/ul
RT: 20.762 min Scan# 2953
Delta R.T. 0.006 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:149 Resp: 6907730
Ion Ratio Lower Upper
149 100
91 70.2 51.7 77.5
206 20.0 17.1 25.7



#84
3,3'-Dichlorobenzidine
Concen: 134.892 ng/ul
RT: 21.556 min Scan# 3089
Delta R.T. 0.012 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:252 Resp: 4266809
Ion Ratio Lower Upper
252 100
254 65.5 52.2 78.4
126 11.6 8.8 13.2









#85
Benzo(a)anthracene
Concen: 128.104 ng/ul
RT: 21.627 min Scan# 3101
Delta R.T. 0.011 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:228 Resp:13497615
Ion Ratio Lower Upper
228 100
229 21.0 16.2 24.2
226 28.4 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 116.689 ng/ul
RT: 21.527 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:149 Resp: 8868869
Ion Ratio Lower Upper
149 100
167 28.2 23.3 34.9
279 4.0 3.6 5.4





#87

Chrysene

Concen: 122.304 ng/ul

RT: 21.692 min Scan# 3109

Delta R.T. 0.017 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:228 Resp:12126815

Ion Ratio Lower Upper

228 100

226 31.2 24.2 36.2

229 20.6 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 105.042 ng/ul




RT: 22.480 min Scan# 3246


Delta R.T. -0.000 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Tgt Ion:149 Resp:15500475









#90
Benzo(b)fluoranthene
Concen: 142.911 ng/ul
RT: 23.380 min Scan# 31
Delta R.T. 0.023 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29




Instrument :
BNA_M
ClientSampleId :
SSTD160003


Tgt Ion:252 Resp:16782747
Ion Ratio Lower Upper
252 100
253 23.0 17.7 26.5
125 10.7 9.3 13.9



#91
Benzo(k)fluoranthene
Concen: 111.745 ng/ul
RT: 23.439 min Scan# 3409
Delta R.T. 0.029 min
Lab File: BM037817.D
Acq: 01 Dec 2022 19:29

Tgt Ion:252 Resp:12797454
Ion Ratio Lower Upper
252 100
253 23.1 17.8 26.8
125 10.3 9.3 13.9





#93

Benzo(a)pyrene

Concen: 128.774 ng/ul

RT: 24.044 min Scan# 3

Delta R.T. 0.035 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:252 Resp:13143727

Ion Ratio Lower Upper

252 100

253 23.1 17.4 26.2

125 11.5 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 124.004 ng/ul

RT: 26.762 min Scan# 3974

Delta R.T. 0.041 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29




Tgt Ion:276 Resp:14520150


Ion Ratio Lower Upper

276 100

138 25.8 22.7 34.1

277 25.4 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 119.817 ng/ul

RT: 26.797 min Scan# 3970

Delta R.T. 0.059 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Instrument :

BNA_M

ClientSampleId :

SSTD160003





Tgt Ion:278 Resp:12596753

Ion Ratio Lower Upper

278 100

139 16.0 13.9 20.9

279 24.6 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 132.491 ng/ul

RT: 27.591 min Scan# 4115

Delta R.T. 0.064 min

Lab File: BM037817.D

Acq: 01 Dec 2022 19:29

Tgt Ion:276 Resp:13509648

Ion Ratio Lower Upper

276 100

138 23.1 19.2 28.8

277 24.3 19.1 28.7

